B　滅菌・乾燥システム
a　高圧蒸気滅菌器庫内の容量は、1,200Ｌ以上であること。
b　緊急時用に別途クリーン蒸気発生装置を有すること。
c　ＥＯガス滅菌は、排気ガスを無毒化できるＥＯガス排出処理装置を有すること。
d　高圧蒸気滅菌・ＥＯガスともに、環境保護に配慮されていること。
e　全ての滅菌器は、両扉方式で開閉できること。
f　滅菌工程がパソコン上で管理できること。滅菌工程が確認可能であること。
g　作業台は、作業者に合わせて、高さ調節が行えること。
h　流し台は、ジェットガン・エアーガン等を有するシステム流し台であること。
C　滅菌物保管システム
a　病棟払い出し物品用パスボックスは、滅菌物の収納、運搬が速やかに行えること。
b　収納スペースが自由に調整できること。
c　清掃がしやすい素材であること。
D　滅菌物管理システム
a　滅菌物管理システムは、滅菌物の品質管理を総合的に行う物であり、バーコードやＩＣタグ等の使用で運用され、洗浄・滅菌システムと連動し、滅菌物のマスター管理・所在管理・処理工程管理・使用管理が行えるものであること。
b　システム始動後に器材が増えた場合でも、追加対応が可能であること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成22年10月18日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒514―8507 津市江戸橋２―174 三重大学財務部病院再開発調達プロジェクトチーム　丹羽　　章　電話059―231―5716
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成22年９月１日から平成22年10月18日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成22年９月14日10時00分
⑵　開催場所　三重大学医学部会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Material washing sterilization system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Cleaning system
a　It shall be a single tank type with full automatic operation, being able to make a track record.
b　The cleaning process is managed by computer system, and the washing method shall be automatically selected according to the type of materials.
c　The system is equipped with the RO water generator for the final washing.
d　The system shall be environmentally conscious.
B　Sterilization and dry system
a　The capacity of high-pressure steam sterilizer shall be 1,200 litter or more.
b　The system is separately equipped with a pure steam generator for an emergency use.
c　EO gas sterilizer is equiped with EO gas exhaust processor to make the exhaust nontoxic.
d　Both of high-pressure steam and EO gas sterilers shall be environmentally conscious.
e　All sterilizer are equipped with double doors for pass-through operation.
f　The sterilization process shall be managed by computer system and be monitored.
g　The height of working table shall be adjustable for that of workers.
h　Sink cabinet shall be equipped with jet or air gun.
C　Storing system of sterilized materials
a　The system shall pocess pass boxes for secure storage and　rapid delivery of sterilized matrrials to in-patient wards.
b　Storage space shall be adjusted freely.
c　The systems shall be made of the materials cleaned up easily.
D　Management systemt of sterilized materials
a　Management system shall enable the quality control of the sterilized materials in a comprehensive manner, being in cooperation with cleaning and sterilizing systems. All the necessary information, stored in the master database, should be retrieved by RFID and bar-coding system to trace where the instruments are, in what process and whether the instruments are in use or not.
b　When the number of tools and materials increases even after the system starts, it shall be adjustable.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 18, October, 2010
⑹　Contact point for the notice : Akira Niwa, University Hospital Redevelopment Project Team, Finance Department, Mie University, 2―174 Edobashi Tsu-shi 514―8507 Japan, TEL 059―231―5716

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成22年９月１日
独立行政法人国立病院機構
弘前病院長　佐藤　年信
◎調達機関番号　597　◎所在地番号　02
○第 00807 号（No. 00807）
1　調達内容
⑴　品目分類番号　22、31
⑵　購入等物品及び数量　マルチスライスＣＴ１式
2　意見の提出方法
⑴　意見の提出期限　平成22年10月６日17時15分（郵送の場合は必着のこと。）
⑵　提出先　〒036―8545 青森県弘前市大字富野町１　独立行政法人国立病院機構弘前病院企画課長　谷下田喜代志　電話0172―32―4311　内線131
3　仕様書案の交付
⑴　交付期間　平成22年９月１日から平成22年10月５日まで。
⑵　交付場所　上記２⑵に同じ。
4　仕様書案の説明会
⑴　開催日時　平成22年10月７日14時00分
⑵　開催場所　院内会議室
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be procured : Multi-slice Computed Tomography 1 set
⑶　Time-limit for the submission of comments : 5 : 15 P.M October 6, 2010
⑷　Contact point for the notice : Kiyoshi Yageta, Director, Accounting Division, Hirosaki National Hospital, 1 Tomino-cho Hirosaki-shi Aomori-ken 036―8545, Japan. TEL 0172―32―4311 ext. 131

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成22年９月１日
経理責任者　独立行政法人国立病院機構
仙台医療センター院長　和田　裕一
◎調達機関番号　597　◎所在地番号　04
○第 01504 号（No. 01504）
1　調達内容
⑴　品目分類番号　71、27
⑵　借入物品及び数量　病院情報システム（電子カルテシステム）　一式
2　意見の提出方法
⑴　意見の提出期限　平成22年10月１日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒983―8520 宮城県仙台市宮城野区宮城野２－８－８　独立行政法人国立病院機構仙台医療センター企画課長　伊藤　幸淑　電話022―293―1111　内線2201