③ Serial Angiographic System
A The system should be designed safely for interventional procedures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography
B C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary direction
C The flat panel detector should be especially designed to fulfill the interventional angiography
D The digital imaging processing system should have supporting and analysis function for intravascular treatment
E The system should equip the radiation exposure reduction unit for patient, operator, and staff
④ Gamma Camera System
A The system should be multiple digital gamma camera
B The system should have function of SPECT and Whole body acquisition
C The system should equip the fusion software and quantitative application-replete
⑤ X-ray fluoroscope Radiography System
A The system is available in fluoroscopic radiography
B The system should capable of recording dynamic images
C The system should be capable of positioning the bed position to horizontal
D The system should equip the flat-panel detector
⑥ General Radiographic System
A The system should equip X-ray tube
B The system should equip a table forstanding and lying position, in addition, radiography of whole body couldable to perform speedily and safely
C The system should equip the flat-panel detector
⑦ Digital X rays equipment to mammography
A The system should be digital X rays equipment to mammography
B The system must be able to digital stereotactic mammography
C The system should equip the high resolution monitor system for mammography diagnosis
D The system must be able to Tomosynthesis
⑧ Digital Mobile Imaiging System : The system should be digital mobile imaiging system
⑸ Time-limit for the submission of the requested material : 5 : 00 P.M. March 17, 2017
⑹ Contact point for the notice : Hiroto Fujiwara, Section Chief, Facilities Planning Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2―5―21 Higashigaoka Meguro-ku Tokyo 152―8621, Japan. TEL 03―5712―5071

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成29年1月30日
経理責任者　独立行政法人国立病院機構
小倉医療センター院長　澄井　俊彦
◎調達機関番号　597　◎所在地番号　40
○第12601号（№12601）
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　高度放射線治療装置　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成29年度第４・四半期
⑸　調達に必要とされる基本的な要求要件
①　X線のエネルギーは２種類以上、電子線のエネルギーは５種類以上を有するリニアック装置であること。
②　多分割コリメータを有すること。
③　検出器の数が64列以上のマルチディテクターＣＴ装置であること。
④　３次元放射線治療計画システムを有すること。
⑤　放射線治療情報管理システムを有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成29年３月21日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒802―8533 福岡県北九州市小倉南区春ヶ丘10―１　独立行政法人国立病院機構小倉医療センター　企画課長　大石　任美　電話093―921―8881　内線8400
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成29年１月30日から平成29年３月17日まで。
⑵　交付場所　上記２⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : High Radiationtherapy System 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　It must be a linac device with more than 2 types of X-ray energy and 5 or more types of electron beam energy.
②　The system should have a multi-part collimator.
③　The system must be a multi-detector CT device with 64 or more detectors.
④　The system shall have a three dimensional radiation therapy planning system.
⑤　The system should have radiotherapy information management system.
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M. March 21, 2017
⑹　Contact point for the notice : Takumi Ooishi, Director, Accounting Division, National Hospital Organization Kokura Medical Center, 10―1 Harugaoka Kokuraminami-ku, Kitakyushu-shi, Fukuoka-ken 802―8533, Japan. TEL 093―921―8881 ext. 8400

意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成29年１月30日
独立行政法人国際協力機構
契約担当役　理事　神崎　康史
◎調達機関番号　567　◎所在地番号　13
1　調達内容
⑴　品目分類番号　71、27
⑵　調達内容及び数量　国際協力人材部及び青年海外協力隊事務局次期業務主管システム設計・開発業務　一式
2　意見の提出方法
⑴　意見の提出期限　平成29年２月21日正午（必着のこと）
⑵　提出先　〒102―8012 東京都千代田区二番町５―25　二番町センタービル　独立行政法人国際協力機構　調達部契約第三課　電話03―5226―6609　電子メールアドレス：ippankeiyakudesk@jica.go.jp
3　仕様書案の掲載
⑴　掲載開始日　平成29年１月30日
⑵　交付期間　平成29年１月30日から平成29年２月20日までの土曜日、日曜日及び祝日を除く毎日、10時00分から17時00分まで（12時30分から13時30分までの間を除く）期間
⑶　掲載場所　http://www.jica.go.jp/chotatsu/buppin/koji2016.html
⑷　交付場所　〒102―8012 東京都千代田区二番町５―25　二番町センタービル　独立行政法人国際協力機構　国際協力人材部　計画・専門員支援課　電話03―5226―6700