b The incubator could keep the temperature in the range of 37 degrees Celsius±0.1 degrees Celsius.
c The processing capacity of the colorimetric item could be 2,000 test／time or more in total.
d The number of the simultaneous measurement items at the same time could be more than 70 items.
e The sampling probe should have the function detect the disorder.
f It should be direct optical measurement of reactor tube, and photometry being possible more than 33 points.
g It could measure in the photometry range of 0 - 3.0Abs.
h It should be the measurement wave length have 12 wavelengths or more.
i Operation by the existing specimen barcode could be possible, and available of the online data input and output with the existing laboratory tests system.

K Full automatic sugar analyzers
a Being electrode method using a D-D-glucose oxidase immobilized enzyme.
b Serum, plasma, urine being measurable as a specimen.
c Have processing capacity, 150 or more／hour.
d Even a mixture of blood-collecting tube and sample cup being measurable.
e Automatic proofreading of two points with the standard solution or one point being possible.
f Being able to connect it with a glycosylated hemoglobin analyzer.
g Operation in the existing specimen bar code is possible, and be available for the online data input and output with the existing laboratory tests system.

L Glycosylated hemoglobin analyzer
a Principle of measurement assume high performance liquid chromatography.
b The thing that is measurable with whole blood.
c Have processing capacity, 60 tests or more／hour.
d Being able to connect it with all automatic sugar analyzers
e Operation in the existing specimen bar code is possible, and be available for the online data input and output with the existing laboratory tests system.

M Osmometric device
a Osmotic pressure of body fluid such as serum, plasma, the urine being measurable at resolution of 1mOsm.
b The specimen consumption have CV 1% precision it or less in 200 μL or less.
c Have processing capacity, 20 or more test／hour.

N Specimen stocker
a In sample stocker, blood-collecting vessel, a conservation specimen being put on more than 700 together.
b Have cooling function (4±3 degrees Celsius).

(5) Time limit for the submission of the requested material : 17 : 00　19 March, 2018
(6) Contact point for the notice : Hisanori Morinaga, Procurement Section, Accounting Division, National University Corporation University of Miyazaki, 1—1 Gakuen Kibana-dai Nishi Miyazaki-shi 889—2192 Japan, TEL 0985—58—7574

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成30年2月1日
総務省大臣官房会計課企画官　笠木　繁樹
◎調達機関番号　012　◎所在地番号　13
1　調達内容
(1) 品目分類番号　71、27
(2) 購入等特定役務及び数量　総務省共通基盤支援システム更改に係るシステム基盤構築等の請負　一式
2　意見の提出方法
(1) 意見の提出期限　平成30年2月21日午後5時（郵送の場合は必着のこと。）
(2) 提出先　〒100—8926 東京都千代田区霞が関二丁目1番2号　総務省大臣官房会計課契約第二係担当　坪内　　大　電話03—5253—5132
3　仕様書案の交付
(1) 交付期間　平成30年2月1日から平成30年2月21日まで
(2) 交付場所　政府電子調達システムへアクセスし入手すること。
4　Summary
(1) Classification of the services to be procured : 71, 27
(2) Nature and quantity of the services to be required : System construction associated with renewal of the common platform support system of MIC 1 set.
(3) Time limit for the submission of comments : 5 : 00 PM 21 February 2018
(4) Contact point for the notice : Masaru Tsubouchi, Accounts Division, Minister's Secretariat, Ministry of Internal Affairs and Communications, 1—2 Kasumigaseki 2chome, Chiyoda-ku Tokyo 100—8926 Japan. TEL 03—5253—5132

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成30年2月1日
国税庁長官官房会計課長　大内　　聡
◎調達機関番号　015　◎所在地番号　13
1　調達内容
(1) 品目分類番号　17
(2) 購入等件名及び数量　平成30年度乗用自動車の購入等（区分1～区分7）一式
2　意見の提出方法
(1) 意見の提出期限　平成30年2月22日17時00分（送付の場合は必着のこと。）
(2) 提出先　〒100—8978 東京都千代田区霞が関3－1－1　国税庁長官官房会計課契約第二係　伊藤　卓也　電話03—3581—4161　内線3646
3　仕様書案の交付
(1) 交付期間　平成30年2月1日から平成30年2月22日まで
(2) 交付場所　〒100—8978 東京都千代田区霞が関3－1－1　東受付室内入札説明室
4　Summary
(1) Classification of the products to be procured : 17
(2) Nature and quantity of the products to be purchased : Automobiles etc ①～⑦ 1 set in the 2018 fiscal year
(3) Time-limit for the submission of comments : 17 : 00　22 February 2018.
(4) Contact point for the notice : Takuya Ito Procurement Section, Accounts Division, Commissioner's Secretariat, National Tax Agency, 1—1 Kasumigaseki 3chome, Chiyoda-ku, Tokyo 100—8978 Japan. TEL 03—3581—4161 ex. 3646

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成30年2月1日
国土交通省国土地理院長　村上　広史
◎調達機関番号　020　◎所属地番号　08
1　調達内容
(1) 品目分類番号　26
(2) 購入等特定役務及び数量
航空重力計及び関連装置の購入　一式
2　意見の提出方法
(1) 意見の提出期限　平成30年2月23日17時00分（郵送の場合は必着のこと。）
(2) 提出先　〒305—0811 茨城県つくば市北郷1番　国土交通省国土地理院測地部物理測地課技術専門員　小板橋　勝　電話029—864—1111（内線4341）
3　仕様書案の交付
(1) 交付期間　平成30年2月1日から平成30年2月23日まで。
(2) 交付場所　上記2(2)に同じ。
4　Summary
(1) Classification of the Products to be procured : 26
(2) Nature and quantity of the Products to be purchased : Purchase of Airborne Gravimeter and related apparatus, 1 set