5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : X-ray Angiography system 2 sets (should be installed catheter room in the new center hospital)
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The system should a bi-plane system which is able to safely fluoroscopy and radiography from the front and the side
②　The system should be able to safely diagnose and treat arrhythmia
③　The system should have a function to examine the whole body, from head to peripheral, without moving the patient
④　The system should carry a flat-panel detector which has high X-ray conversion efficiency and density resolution
⑤　The flat-panel detector should be able to diagnose and treat by catheter and ablation
⑥　The digital image processing system should have the function to assist the ablation with mapping system
⑦　The system should have a large LCD which is able to project multiple images
⑧　The system should be able to reconstruct by using 3D image data
⑨　The system should be able to connect with existing HIS, RIS, PACS, and diagnosis system
⑩　The bidder should provide peripheral equipment that is effective for managing machines and devices
⑪　The better should have an organization which provides quick user support and adequate maintenance of the system
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M 6, July 2018
⑹　Contact point for the notice : Masataka Tsujimoto, Finance and Accounting Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext. 2982

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年5月22日
国立研究開発法人
国立循環器病研究センター
理事長　小川　久雄
◎調達機関番号　820　◎所在地番号　27
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　血管撮影装置　1式（新センター　カテーテル室⑥に該当）
⑶　調達方法　購入等
⑷　導入予定時期　平成31年1月末日
⑸　調達に必要とされる基本的な要求要件
①　正面・側面の2方向から透視、撮影が可能なバイプレーンシステムであること。
②　脳血管障害に対する血管造影検査に対応可能なシステムであること。
③　患者を移動することなく、頭部から下肢まで全身を､検査可能な機能を有すること。
④　X線変換効率、濃度分解能が高い平面検出器（フラットパネルディテクタ）搭載装置であること。
⑤　脳血管内治療特有のデバイス等に適した高精細な画像を得られる専用モードを有すること。
⑥　パルスレート可変デジタル透視、自動付加フィルタ挿入など、複合的に被ばく低減可能なプログラムを有すること。
⑦　多画像の表示が可能な大型液晶ディスプレイを有すること。
⑧　3次元画像データの取得、再構成が可能であること。
⑨　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、医用画像管理システム（ＰＡＣＳ）、読影システムと接続可能であること。
⑩　装置および機器の運営管理に有用な周辺機器を備えていること。
⑪　装置の維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料の提供期限　平成30年7月6日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台5―7―1　国立研究開発法人国立循環器病研究センター　財務経理部契約第一係　辻本　真孝　電話06―6833―5012（内線）2982
メールアドレス　Sounding@ml.ncvc.go.jp
3　導入説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年5月22日から平成30年7月6日まで。
⑵　交付方法　2⑵に示す場所にて直接交付するか、若しくは電子メールにて交付する。電子メールでの交付を受ける場合は2⑵に示すメールアドレスに調達件名・会社名・担当者名・連絡先を入力し、参加希望の旨メール送信すること。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更する事があり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : X-ray Angiography system 1 set (should be installed catheter room ⑥ in the new center hospital)
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The system should a bi-plane system which is able to fluoroscopy and radiography from the front and the side
②　The system should be able to use cerebrovascular disease
③　The system should have a function to examine the whole body, from head to peripheral, without moving the patient
④　The system should carry a flat-panel detector which has high X-ray conversion efficiency and density resolution
⑤　The system should have a dedicated mode which is suited for taking high-definition images including unique devices for cerebral intravascular surgery
⑥　The system should have several type of dose reduce programs such as variable pulse rate digital fluoroscopy, automatic additional filter
⑦　The system should have a large LCD which is able to project multiple images
⑧　The system should be able to reconstruct by using 3D image data
⑨　The system should be able to connect with existing HIS, RIS, PACS, and diagnosis system
⑩　The bidder should provide peripheral equipment that is effective for managing machines and devices
⑪　The bidder should have an organization which provides quick user support and adequate maintenance of the system
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M 6, July 2018
⑹　Contact point for the notice : Masataka Tsujimoto, Finance and Accounting Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext. 2982

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年5月22日
国立研究開発法人
国立循環器病研究センター
理事長　小川　久雄
◎調達機関番号　820　◎所在地番号　27
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　血管撮影装置　1式（新センター　カテーテル室⑦に該当）
⑶　調達方法　購入等
⑷　導入予定時期　平成31年1月末日
⑸　調達に必要とされる基本的な要求要件
①　正面・側面の2方向から透視、撮影が可能なバイプレーンシステムであること。
②　循環器疾患に対する血管造影検査に対応可能なシステムであること。