6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Infusion Pump and Syringe pump 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Be able to adjust of infusion speed.
B　Be able to blood transfuse.
C　Be able to give IV solution and blood components. No extra setting is required for switching from IV solution to blood pumping, and vice versa.
D　Be able to show event logs.
E　Be able to continue to work by battery.
F　Have safety and communication functions.
⑸　Time limit for the submission of the requested material : 17 : 00　1 October, 2018
⑹　Contact point for the notice : Masaya Okuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 30 年 8 月 16 日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第５号
1　調達内容
⑴　品目分類番号　15、28
⑵　導入計画物品及び数量
高度急性期統合ネットワークシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年度10月以降
⑸　調達に必要とされる基本的な要求要件
A　平成31年度開院予定の中病棟及びｉＰＳ等臨床試験センターにおける、医療機器、医療IoTデバイス、モバイル端末を含む電子カルテシステム端末及びその他の情報通信機器が接続・通信可能であること。
B　本院全体の基幹ネットワークシステムである大容量・高速医療情報ネットワークシステムに接続し、一体として運用することが可能であること。
C　高度なセキュリティ及び通信制御機能を備えていること。
D　修理・保守・代替品の提供が可能であること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年10月１日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課契約掛　奥田　昌也　電話075―751―3025
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年８月16日から平成30年10月１日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年８月30日11時00分
⑵　開催場所　京都大学医学部附属病院会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 15, 28
⑵　Nature and quantity of the products to be purchased : Advanced Acute-Phase Integrated Network System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The procured system must connect to and communicate with medical devices, medical IoT devices, hospital information system terminals (including mobile devices) and information communication equipment introduced to Middle Ward and iPS Clinical Trial Center which will open in 2019.
B　The procured system must connect to the Integrated High-Speed Hospital Information Network System which is the backbone network system of the entire hospital, and the both systems must be possible to be operated as a single unit.
C　The procured system must have advanced security and communication control function.
D　The procured system can be repaired and can be maintained if necessary, and there is an appropriate substitute.
⑸　Time limit for the submission of the requested material : 17 : 00　1 October, 2018
⑹　Contact point for the notice : Masaya Okuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 30 年 8 月 16 日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第６号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
核医学診療用放射線管理システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年度９月以降
⑸　調達に必要とされる基本的な要求要件
A　平成31年３月に設置予定のサイクロトロン・ホットラボシステムの、Ｃ―11、Ｎ―13、Ｆ―18、Ga―68などの核種の標識薬剤合成および品質管理を行うための機器（合成装置、分注装置、分析装置など）および放射線防護システム（清潔作業を行うための空調機能も含む）を有すること。
B　核医学診療を行う部屋の放射線管理システムで、具体的には、ＲＩ管理区域内への入退室管理、ＲＩ管理区域内立入者の被ばく管理機能を備えていること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年10月１日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課契約掛　奥田　昌也　電話075―751―3025
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年８月16日から平成30年10月１日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年８月30日11時30分
⑵　開催場所　京都大学医学部附属病院会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : The Radiation Control System for Nuclear Medicine 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The instruments for synthesis of radiopharmaceuticals labeled with C-11, N-13, F-18 and Ga-68, and those for quality control of radiopharmaceuticals in addition to radiation protection systems including air condition control function. These systems are acquired to be compatible for the cyclotron and PET tracer laboratory system which is decided to be installed in March 2019.