6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Life Support Systems 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The instrument must include units such as Intra-aortic balloon pump (IABP), Percutaneous cardiopulmonary assist device system (PCPS), Electrocardiograph, Arterial pressure cardiac output meter, Dilution cardiac output meter, Indications must be in Japanese, clearly visible and easily operable.
B　Each unit must be compact, lightweight, easy to transport, and can be driven by a built-in or an external battery.
C　Intra-aortic balloon pump (IABP) must :
a　have sensitive responses to balloon catheters.
b　have auto time adjustment function for INF／DEF.
c　have a backup function in case of the disappearance of ECG synchronization signal.
D　Percutaneous cardiopulmonary assist device system (PCPS) must :
a　be able to drive its all the functions of centrifugal pump, electronic blender, and cold／hot water tank with a built-in battery.
b　be able to check the history of events such as warnings on a display.
E　Electrocardiograph must :
a　allow the recording of standard 12-lead resting electrocardiogram and also have the function to analyze the coronary artery responsible for the infarction and stenosis based on standard 12-lead electrocardiograms.
b　have an analytical guide function that explains the ground for the judgement of findings obtained through the analysis.
F　Arterial pressure cardiac output meter must :
a　be able to continuously measure cardiac output and central venous oxygen saturation using the line or the route of invasive arterial pressure measurement.
b　be able to display the parameters related to cardiac output such as stroke volume and body vascular resistance.
G　Dilution cardiac output meter must :
a　be able to continuously measure cardiac output, mixed venous blood oxygen saturation or central venous oxygen saturation, and right ventricular end-diastolic volume.
b　be equipped with a color display, and able to display the numerical value and trend graph of each parameter.
⑸　Time limit for the submission of the requested material : 17 : 00　7 December, 2018
⑹　Contact point for the notice : Masaya Okuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年10月22日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第10号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
内視鏡システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　気管支内視鏡に関連するスコープ、光源装置、画像構築装置、ディスプレイモニター、内視鏡消毒洗浄装置、内視鏡保管庫、その他周辺機器から構成され、診断、治療等の内視鏡診療全般に対応する最新機能を備えていること。
B　内視鏡画像、患者情報、所見レポートを記録し、データの保存管理及び業務支援を行うための画像サーバーシステム及び病院医療情報システムに、接続・連携できること。それにより、検査結果を迅速かつ効率的に共有できること。
2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年12月7日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課契約掛　奥田　昌也　電話075―751―3025
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年10月22日から平成30年12月7日まで。
⑵　交付場所　上記2⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年10月30日15時35分
⑵　開催場所　京都大学医学部附属病院会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Endoscopy system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　It must consist of scopes, light source devices, image construction devices, display monitors, endoscope cleaning／disinfection devices, endoscope storage cabinets, and other peripheral equipments, related to bronchoscope. Besides, it must have the latest features corresponding to endoscopic diagnosis and treatment in general.
B　It must be able to connect and cooperate with the image server system and the hospital medical information system which record endoscopic images, patient informations, endoscopy findings reports, and perform data storage management and operation support. Thereby, it must enable medical staffs to share the results of endoscopy quickly and efficiently.
⑸　Time limit for the submission of the requested material : 17 : 00　7 December, 2018
⑹　Contact point for the notice : Masaya Okuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成30年10月22日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第11号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
高度先進脳神経救急・集中治療管理システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　脳神経外科用手術顕微鏡、４Ｋ映像システム、脳波計、誘発電位検査装置、ワークステーション機能付画像診断装置、超音波ドプラ血流測定装置、グリオーマ検査機器、脳内圧モニタ、持続脳波モニタ、ビデオ喉頭鏡、自動血球計数装置、その他周辺機器から構成され、診断、処置、治療等の高度先進脳神経集中治療管理に対応する最新機能を有すること。