B　患者情報、所見レポートを記録し、データの保存管理及び業務支援を行うための画像サーバーシステム及び病院医療情報システムに、接続・連携できること。それにより、検査結果を迅速かつ効率的に共有できること。
C　脳神経外科用手術顕微鏡はより高い安全性と正確性を実現するために、シャープで鮮明な術野を提供する最新光学系、高解像度カメラ、そして優れた顕微鏡操作を実現するスタンドを有すること。また裸眼では観察できない術中蛍光機能（５－ＡＬＡ、indocyanine Green）を有し、術中において多角的な情報を通して適切な判断を行うことができること。
D　４Ｋ映像システムは、４Ｋ大型ディスプレイを有し、４Ｋレベルの動画、静止画を多画面化に提示できる機能を持つこと。また、それらを保存するに十分な記録装置を有すること。
E　脳波計は時定数が最大10秒まで設定可能であり、多チャンネルの入力にも対応すること。また、脳波データをＦＦＴ解析し、周波数表示が可能であること。
F　誘発電位検査装置は神経機能をモニタリングでき、測定チャンネルは32ch以上を有し、電気刺激部は電圧1000Ｖ／電流250mAに対応していること。また測定部のＡ／Ｄ分解能は18bit以上あること。
G　脳内圧モニタはＭＥＳマイクロセンサーを有しており、患者モニタと連動することが可能であること。
H　ワークステーション機能付画像診断装置、４Ｋ大型ディスプレイを有し、システム部分はカンファレンスに必要な各種映像を必要に応じて集約・切り替え表示が可能であること。また、それらを保存するに十分な記録装置を有すること。
I　細胞分析装置は、臨床的検査に対応できるレーザー・検出器構成を持ち、10カラー：３レーザー（Blue、Red、Violet）を備え持つ細胞分析装置であること。
J　超音波ドプラ血流測定装置は、頭部に固定し、持続的に経頭蓋超音波ドプラ測定ができること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成30年12月７日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課契約掛　奥田　昌也　電話075―751―3025
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成30年10月22日から平成30年12月７日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成30年10月30日16時00分
⑵　開催場所　京都大学医学部附属病院会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : A highly advanced medical system for neurological emergencies and intensive care management 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system needs highly advanced medical availability for diagnosis, treatment and management, which is composed of a surgical microscope for neurosurgery, a 4K video recording system, a electroencephalograph system, a evoked potential recording device, an diagnostic imaging apparatus with workstation, an ultrasonic Doppler blood flow meter, a fluorescence activated cell sorting analysis system, intracranial pressure monitoring devices, continuous electroencephalograph monitor devices, a video laryngoscope, an Fully automated blood cell counting device, and peripheral devices.
B　The system needs a server system to support medical managements by recording and analyzing patient's information and findings, and to connect to our hospital medical information system. It can present all information rapidly and effectively.
C　The surgical microscope for neurosurgery needs an advanced optical, imaging video imaging system and framework to support safe and precise surgery. It needs a fluorescence (5―ALA, indocyanine Green) imaging system, a link system to navigation device and analyzing system to treat various medical informasiton.
D　The 4K video recording system needs multichannel control functions for 4K movies and 4K images with 4K-large display. It needs recording system for 4K movies and images.
E　The electroencephalograph system needs special functions to set time constant to a maximum of 10 seconds and to treat multichannel recording. It can analyze electroencephalograph data with FFT method and present frequency data.
F　The evoked potential recording device needs neurological function monitoring function with more than 32 channels and 1000V／250mA stimulation. It need A／D resolution with more than 18 bit.
G　The intracranial pressure monitoring devices needs MES micro-censor and connection to patient medical information system.
H　The diagnostic imaging apparatus with workstation needs 4K large display and function to control image presentation. It needs adequate data recording system.
I　The cell analyzing system needs fluorescence activated cell sorter with laser (Blue, Red, Violet) and detector composed of 10 colors and 3 lasers.
J　The ultrasonic doppler blood flow meter needs continuous transcranial Doppler with its head fixation.
⑸　Time limit for the submission of the requested material : 17 : 00　7 December, 2018
⑹　Contact point for the notice : Masaya Okuda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 30 年 10 月 22 日
国立大学法人京都大学長　山極　壽一
◎調達機関番号　415　◎所在地番号　26
○第 12 号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
患者用ベッド　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成31年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　一般病棟向け電動ベッド及び高次集中治療室向けＩＣＵベッドから構成されること。
B　一般病棟向け電動ベッドは、ベッドフレーム、ベッドサイドレール、オーバーベッドテーブル、及びナースコール中継ユニットなどから構成されていること。必要に応じて背上げ、膝上げ及び高低の操作ができること。
C　高次集中治療室向けＩＣＵベッドは、ベッドフレーム、デジタルスケール、ベッドサイドテーブル、オーバーベッドテーブル、ＩＶポール及び酸素ボンベホルダーなどから構成されていること。必要に応じて、座位姿勢、背上げ、膝上げ及び高低の操作ができること。
D　ベッドサイドレールやヘッドフレームは、衛生管理がしやすい素材であること。