資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成30年12月17日

国立大学法人山口大学長　岡　　正朗

◎調達機関番号　415　◎所在地番号　35

○第３号

1　調達内容

⑴　品目分類番号　15、28

⑵　導入計画物品及び数量　ナースコールシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成31年度６月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　各ナースステーションに設置されるＮＣＳ親機は視認性に優れたものを用意し、タッチパネルによる操作を可能とすること。

Ｂ　ナースコールによる呼び出し時には、呼び出した患者をわかりやすく表示し、呼び出し用途（緊急呼び出し、一般呼び出し、排泄介助要請、脱落・断線､離床センサー、ドアセンサー、生体情報監視システムからの警報など）を明示すること。

Ｃ　看護に必要な情報（患者氏名、診療科、担当看護師など)をＨＩＳと連携すること。

Ｄ　ナースコールによるＰＨＳへの呼び出しは８台以上を可能とすること。

Ｅ　ＰＨＳシステムは本院で導入済みの交換機システム（沖電気工業製 DISCOVERY neo）に増設すること。

Ｆ　本院で導入済みの電子カルテシステムと連携すること。

Ｇ　各病室前に設置する表示灯には、ＨＩＳから取得した患者情報のうち、本院が指定する情報を表示すること。また、患者ごとに表示情報を設定できること。

Ｈ　各ベッドに設置するナースコール呼び出しスイッチは、患者の状態により複数の呼び出し方法（ボタン、息、声、手を近づける、舌を触れる等）を１台で切替可能であること。

Ｉ　ナースコールによる呼び出し履歴を記録し、分析可能なようにデータの取り出しを可能にすること。

Ｊ　ＮＣＳの保守体制は24時間365日体制とすること。

Ｋ　機器導入時の付帯工事では病棟閉鎖は行わないこととし、病室内での作業は必要最小限とすること。また、作業計画を事前に提出し、本院の許可を得ることとし、作業について本院から指示があった場合にはその内容に従うこと。

Ｌ　無停電電源装置は、サーバ、ＮＣＳ親機用に設置すること。

Ｍ　導入するＰＨＳは既存の内線固定電話と通話が可能なこと。

Ｎ　システムの都合よりも、本院の運用を重視したＮＣＳ及び作業計画を提案すること。

Ｏ　医療安全を考慮し、患者間違いや操作間違いが発生しにくい仕組みを提供すること。

Ｐ　厚生労働省の「医療情報システムの安全管理に関するガイドライン」の最新版に準拠したシステムを提供すること。

Ｑ　ＮＣＳに不備、不具合、セキュリティの問題が見つかった場合、ＮＣＳをバージョンアップする等速やかに対応すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成31年１月31日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒755―8505 山口県宇部市南小串１－１－１　国立大学法人山口大学医学部管理運営課契約管理係　熊崎　大樹　電話0836―22―2031

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成30年12月17日から平成31年１月31日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成31年１月９日15時00分

⑵　開催場所　山口大学医学部附属病院会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 15, 28

⑵　Nature and quantity of the products to be purchased : Nurse Call System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　the Master unit installed at each nurse station should prepare one with excellent visibility and enable operation with the touch panel.

B　when calling by a nurse call, display the calling patient in an easy to understand manner and clearly indicate the calling purpose (emergency call, general call, excretion assistance request, dropout/disconnection, bed leaving sensor, door sensor, warning from the Integrated Vital Sign Monitoring and Recording System, etc.) .

C　communicate with the HIS the information necessary for nursing (such as patient name, department name, nurse in charge) .

D　calling to PHS by nurse call should make more than 8 calls possible.

E　the PHS system should be added to the switch system (DISCOVERY neo, Oki Electric Industry Co. , Ltd.) already installed at our hospital.

F　cooperate with the electronic medical record system already installed at our hospital.

G　for indication lamps installed in front of each hospital room, display information designated by the hospital among patient information acquired from HIS. Also, it is possible to set display information for each patient.

H　the nurse call call switch installed in each bed should be able to switch multiple invocation methods (button, breath, voice, close the hand, touch the tongue, etc.) depending on the condition of the patient with one unit.

I　record the call history by the nurse call and make it possible to retrieve the data so that it can be analyzed.

J　the maintenance system of NCS shall be organized 24 hours a day, 365 days a year.

K　in the incidental work at the time of introduction of NCS, it is decided not to close the ward and work within the hospital room should be minimized. Also, submit a work plan in advance, obtain permission from our hospital, and follow instructions given by our hospital about the work.

L　the uninterruptible power supply shall be installed for the server, NCS base unit.

M　the PHS to be introduced must be able to talk with existing extension fixed telephone.

N　to propose NCS, work plan that emphasizes the operation of our hospital than the convenience of the system.

O　considering medical safety, to provide a mechanism that makes patient mistakes and mistakes less likely to occur.

P　provide a system that complies with the latest edition of the Ministry of Health, Labor and Welfare Guidelines for Safety Management of Medical Information Systems.

Q　incomplete, troubles, security problems in NCS should be addressed promptly, such as upgrading NCS.

⑸　Time limit for the submission of the requested material : 17 : 00　31 January, 2019

⑹　Contact point for the notice : Hiroki Kumasaki, Procurement, Accounting Division, School of Medicine, Yamaguchi University, 1―1―1 Minamikogushi Ube-shi 755―8505 Japan, TEL 0836―22―2031

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

平成30年12月17日

総務省大臣官房会計課企画官　中村　英男

◎調達機関番号　012　◎所在地番号　13

1　調達内容

⑴　品目分類番号　67、29

⑵　購入等特定役務及び数量　遠隔方位測定設備センサ等回線一括提供の請負　一式