6　Summary
(1)　Classification of the products to be procured : 22, 31
(2)　Nature and quantity of the products to be purchased : X-ray angiography System 1 Set
(3)　Type of the procurement : Purchase
(4)　Basic requirements of the procurement :
A　The angiography device is a flat panel detector (FPD). Be able to see and shoot.
B　Cone beam CT is possible and 3D image reconstruction is possible at high speed.
C　Equipped with the latest applications such as applications that enhance stent rendering Doing things.
D　It has dose reduction and protection functions for patients and workers, and displays dose in real time.Equipped with functions that can be used.
E　In-hospital hospital information system network (HIS／RIS／PACS／video network) Be able to connect to a server, radiation dose management system, etc.).
F　As a facility that provides highly advanced medical care, it can be used for research and development.
G　Perform construction to install the system.
(5)　Time limit for the submission of the requested material : 17 : 00, 22 November, 2019
(6)　Contact point for the notice : Makoto Nomoto, Procurement Section, Accounting Division, Chiba University Hospital, 1―8―1 Inohana Chuo-ku Chiba-shi 260―8677 Japan, TEL 043―226―2236

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和元年10月7日
国立大学法人千葉大学事務局長　松浦　晃幸
◎調達機関番号　415　◎所在地番号　12
○第７号
1　調達内容
(1)　品目分類番号　22、31
(2)　導入計画物品及び数量
ＩＶＲ―ＣＴシステム　一式
(3)　調達方法　購入等
(4)　導入予定時期　令和２年度12月以降
(5)　調達に必要とされる基本的な要求要件
Ａ　血管撮影装置は天井走行式で大視野フラットパネルディテクタ（ＦＰＤ）であること。
Ｂ　ＣＴ装置は80列以上の自走式マルチスライスＣＴであること。
Ｃ　２Ｄ３Ｄフュージョンアプリケーションなど最新のアプリケーションを搭載していること。
Ｄ　画質向上と被ばく低減が行える最新機能を有すること。
Ｅ　放射線防護・管理および、特殊検査に伴う付属機器・部品を備えること。
Ｆ　院内の病院情報システムのネットワーク（ＨＩＳ・ＲＩＳ・ＰＡＣＳ・動画ネットワークサーバ・放射線線量管理システム等）と接続できること。
Ｇ　高度先進医療を提供する施設として、研究・開発に利用できること。
Ｈ　システムを設置するための工事を行うこと。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1)　資料等の提供期限　令和元年11月22日17時00分（郵送の場合は必着のこと。）
(2)　提供先　〒260―8677　千葉市中央区亥鼻１－８－１　千葉大学医学部附属病院管理課用度係　野本　　誠　電話043―226―2236
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1)　交付期間　令和元年10月７日から令和元年11月22日まで。
(2)　交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1)　開催日時　令和元年10月９日15時00分
(2)　開催場所　千葉大学医学部附属病院セミナー室１
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
(1)　Classification of the products to be procured : 22, 31
(2)　Nature and quantity of the products to be purchased : Interventional Radiology Computed Tomography System 1 Set
(3)　Type of the procurement : Purchase
(4)　Basic requirements of the procurement :
A　The angiography device shall be a ceiling traveling type and a large visual field flat panel detector (FPD).
B　The CT device is a self-propelled multi-slice CT with 80 rows or more.
C　Equipped with the latest applications such as 2D3D fusion applications Being.
D　Have the latest functions to improve image quality and reduce exposure.
E　Provide accessory equipment and parts for radiation protection／management and special inspection.
F　Hospital information system network (HIS／RIS／PACS／video network) Be able to connect to a server, radiation dose management system, etc.).
G　As a facility that provides highly advanced medical care, it should be usable for research and development.
H　Perform construction to install the system
(5)　Time limit for the submission of the requested material : 17 : 00, 22 November, 2019
(6)　Contact point for the notice : Makoto Nomoto, Procurement Section, Accounting Division, Chiba University Hospital, 1―8―1 Inohana Chuo-ku Chiba-shi 260―8677 Japan, TEL 043―226―2236

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和元年10月7日
国立大学法人千葉大学事務局長　松浦　晃幸
◎調達機関番号　415　◎所在地番号　12
○第８号
1　調達内容
(1)　品目分類番号　22、31
(2)　導入計画物品及び数量
ＭＲＩシステム　一式
(3)　調達方法　購入等
(4)　導入予定時期　令和２年度12月以降
(5)　調達に必要とされる基本的な要求要件
Ａ　ＭＲＩシステムの静磁場強度は1.5Ｔとして、２台を構成内に含むこと。
Ｂ　撮像時間の短縮が行える圧縮センシング技術を有すること。
Ｃ　金属インプラントによるアーチファクトの低減が行える機能および局所選択励起技術を有すること。
Ｄ　映像、音楽、照明の色彩等で検査内の雰囲気を変え、患者のストレスを軽減できる設備を有すること。
Ｅ　特殊検査に伴う付属機器・部品を必ず備えること。
Ｆ　院内の病院情報システムのネットワーク（ＨＩＳ・ＲＩＳ・ＰＡＣＳ）と接続できること。
Ｇ　高度先進医療を提供する施設として、研究・開発に利用できること。
Ｈ　システムを設置するための工事を行うこと。
2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1)　資料等の提供期限　令和元年11月22日17時00分（郵送の場合は必着のこと。）
(2)　提供先　〒260―8677　千葉市中央区亥鼻１－８－１　千葉大学医学部附属病院管理課用度係　野本　　誠　電話043―226―2236
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1)　交付期間　令和元年10月７日から令和元年11月22日まで。
(2)　交付場所　上記２(2)に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
(1)　開催日時　令和元年10月９日14時00分
(2)　開催場所　千葉大学医学部附属病院セミナー室１
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。