6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Integrated Rapid Clinical Laboratory System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
① It should be able to evaluate the degree of stenosis of the lower limb arteries. ② It should be able to record 24-hour blood pressure. ③ It should be able to assist nerve block and catheter insertion with ultrasound images. ④ It should be able to analyze microbial components using plasma specimens. ⑤ It should be able to prepare blood collection tubes automatically. ⑥ It should be able to centrifuge blood samples and automatically wash red blood cells. ⑦ It should be able to photograph the gross image of organs and process hard tissues. ⑧ It should be able to perform fundus fluorescein angiography and OCT simultaneously.
⑸　Time limit for the submission of the requested material : 17 : 15　30 March, 2020
⑹　Contact point for the notice : Yuki Yamanaka, Procurement and Contracts Ⅱ, Management and Accounting Division, National University Corporation Ehime University School of Medicine, Shitsukawa Toon-shi Ehime Pref. 791―0295 Japan, TEL 089―960―5214

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和２年２月12日
国立大学法人長崎大学長　河野　　茂
◎調達機関番号　415　◎所在地番号　42
○第８号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　超音波検査システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和２年度２月以降
⑸　調達に必要とされる基本的な要求要件
A　本院既存の電子カルテシステム（NEC製）（以下、「電子カルテ」という。）と連携しオーダー情報を取得できること。
B　本院既存の電子カルテ端末から超音波画像管理ソフトウェアを起動して画像参照できること。
C　電子カルテと検査進捗情報を同期できる機能を有していること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和２年３月30日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒852―8501 長崎市坂本１－７－１　長崎大学病院経営管理課物品担当　清田千恵美　電話095―819―7218
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和２年２月12日から令和２年３月30日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和２年３月３日16時00分
⑵　開催場所　長崎大学病院多目的指導室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Ultrasound inspection system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The ability to start Ultrasonic imaging software from a electronic medical record and refer to an imaging.
B　The ability to obtain information in cooperation with electronic medical records.
C　The ability to synchronize information on the progress of the electronic medical record and the ultrasonic inspection system.
⑸　Time limit for the submission of the requested material : 17 : 00　30 March, 2020
⑹　Contact point for the notice : Kiyota Chiemi, Supplies Section, Management Division, Nagasaki University Hospital, 1―7―1 Sakamoto Nagasaki-Shi 852―8501 Japan, TEL 095―819―7218

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和２年２月12日
国立大学法人鹿児島大学
契約担当役理事　山木　宏明
◎調達機関番号　415　◎所在地番号　46
○第１号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　術中ＭＲＩ診断システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和２年度２月以降
⑸　調達に必要とされる基本的な要求要件
A　手術に有用なＭＲＩ画像を撮影することが可能であり、かつ十分な開口部を持ったＭＲＩであること。
B　ＭＲＩの維持費が比較的安価であり、現在使用している手術用具のほとんどを使用できること。
C　ＭＲＩに対応した手術台を擁すること。
D　術中ＭＲＩ画像を直接読み込むことができ、かつ反映させることのできるナビゲーションシステムを擁すること。
E　ＭＲＩにノイズをもたらすことのない、ＭＲＩ対応の麻酔器を有すること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和２年４月２日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒890―8520 鹿児島市桜ケ丘８―35―１　鹿児島大学病院経理調達課調達係　島名　冬美　電話099―275―5045
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和２年２月12日から令和２年４月２日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和２年３月16日16時00分
⑵　開催場所　鹿児島大学病院管理棟病院多目的室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : A complete set of intraoperative MRI system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Should be a magnet resonance imaging (MRI) system which could give valuable images for operation and have enough width of window to use during operation.
B　Should be an MRI system which could maintain in lower expense and use almost operation instruments which are now being used.
C　Should be a system which could have operation table to compatible to intraoperative MRI.
D　Should be a system which could have neuro-navigation system to import images from MRI system directly and use immediately.