(4) Basic requirements of the procurement :

A The system is the infrastructure environment to promote education and research activities of our department. The system must be configured by educational computer system, education and research support system, educational network system, and operations management system.

B The educational computer system must support the ability for our lecture courses, seminars, and self-learning. In addition, the system must be configured by computers, network facilities, printing equipment, authentication system, backup system, and software, etc.

C The education and research support system supports educational activities, and is composed from servers, classroom equipment, and network facilities.

D The network system must consist of research and education networks, networks to connect the campus network, network service systems, authentication management systems, security management systems, and backup devices. The system must have high-level performance for providing sufficiently communication and service level.

E The operations management system must have an integrated management capability, such as monitoring, recording and notifying the systems status that includes operation, utilization and errors. The system must also have an integrated and centralized user management.

F Sufficient security measures must be given to all the equipment and the applications.

G The system must be equipped with the manuals and documents needed for operating the entire system.

H Constant and quick support system against faults shall be available. Transition from the present system as well as daily operation after the transition also shall be supported.

(5) Time limit for the submission of the requested material : 17 : 00　31 March, 2020

(6) Contact point for the notice : Daisuke Takeshita, Section Chief for Contract and Purchase Section, Accounting at Koganei Accounting Office, Tokyo University of Agriculture and Technology, 2―24―16 Naka-cho Koganei-shi 184―8588 Japan, TEL 042―388―7005

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和２年２月17日

国立大学法人新潟大学長　牛木　辰男

◎調達機関番号　415　◎所在地番号　15

○第３号

１　調達内容

(1) 品目分類番号　22、31

(2) 導入計画物品及び数量　救命救急センター・ＩＣＵモニタリングシステム　一式

(3) 調達方法　購入等

(4) 導入予定時期　令和２年度３月以降

(5) 調達に必要とされる基本的な要求要件

Ａ　本院の中央診療棟、高次救命災害治療センター（高度救命救急センター）の生体情報モニタとして、ベッドサイドモニタ、搬送用モニタを備えているシステムであること。

Ｂ　ＩＣＵ、救急ＩＣＵの12ベッドについては、ベッドサイドモニタの操作が可能なスレーブモニターを備えていること。またスレーブモニターの操作はタッチパネルで行えること。

Ｃ　救急病棟については、ベッド移動や歩行可能な患者の独歩時にはベッドサイドモニタ本体からのワイヤレスにも対応しているシステムであること。

Ｄ　ＩＣＵ、救急ＩＣＵ、救急病棟、救急外来初療室・外傷室については、既存のセントラルモニタに接続ができ、既存の患者情報システムにバイタルデータ送信ができるシステムであること。

Ｅ　ベッドサイドモニタは既存のモジュール（血圧、体温など）を使用する事ができること。

Ｆ　ベッドサイドモニタはカラーワイド型液晶ディスプレイでタッチスクリーン、トラックボール、マウスによる操作が可能なこと。また感染防止のため、本体は冷却用ファンを内蔵していないこと。

Ｇ　ベッドサイドモニタはモジュール方式であり、統合モジュールはベッドサイドモニタに接続された状態で、セカンドディスプレイとして使用できるよう表示する機能を有し、アラーム消音、非観血血圧のスタート／ストップ操作を行う機能を有していること。

Ｈ　搬送用モニタはカラーディスプレイであり、スクリーンにはAntimicrobial Corning®　Gorilla®　Glassが採用されており、モニタ本体の設置状況に応じ縦向き及び横向きに回転表示ができること。

Ｉ　ベッドサイドモニタは波形データを出力し、サーバへデータを保存することができること。

２　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限　令和２年４月３日17時00分（郵送の場合は必着のこと。）

(2) 提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課契約係　野水　忠宏　電話025―227―0970

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

(1) 交付期間　令和２年２月17日から令和２年４月３日まで。

(2) 交付場所　上記２(2)に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

(1) 開催日時　令和２年２月26日14時00分

(2) 開催場所　新潟大学医歯学総合病院西診療棟３階　第３会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

(1) Classification of the products to be procured : 22, 31

(2) Nature and quantity of the products to be purchased : Emergency and Critical Care Center／ICU Patient Monitaoring System 1 set

(3) Type of the procurement : Purchase

(4) Basic requirements of the procurement :

A Patient Monitoring System includes the bedside monitor and the transport monitor for the Emergency and Critical Care medicine in the Central Wing.

B For 12 beds of ICU, Emergency ICU, the patient monitor can connect a touch slave monitor. In addition, it can be operated by touch.

C For emergency wards, the patient monitor supports wireless connection when the staff transfers the patient or the patient walks by him／herself.

D For ICU, Emergency ICU, Emergency ward, Emergency room／Trauma room, the patient monitor can be connected to the existing central monitor and send vital data to the existing clinical information system.

E The patient monitor can be connected existing modules (invasive blood pressure, temperature etc.).

F The bedside monitor has a color wide liquid crystal display that can be operated with a touch screen, trackball, and mouse. To prevent infection, it doesn't have a built-in cooling fan.

G The bedside monitor is a modular system, and the integrated module can display when it is connected to the bedside monitor. The integrated module also has a function to mute alarms, and start／stop measuring non-invasive blood pressure.