⑧　圧力モニタ、レベルセンサ、バブルセンサについては、アラーム報知のみと、ポンプとの連動が選択可能であること。
⑨　Ｐ／Ｓコントロール（ポンプ連動動作）が容易に行え、ブラッドカルディオプレジアも容易に行えること。
⑩　脱血流量に従って、ローラポンプ送血流量を調整する、流量連動機能が搭載されていること。
⑪　上限連動機能により、ターゲットとなる流量より高くならないように制御でき、人工肺陰圧のリスクを低減できること。
⑫　電子ブレンダユニットが使用可能で、酸素ガス濃度や流量など、データ出力可能であること。
⑬　システムベースに非常用電源を備えており、停電時のバックアップが構築されていること。
⑭　心筋保護装置「体外循環用システムＴＲＵＳＹＳ」との連動が可能であること。
⑮　アフターサービスケアにおいて、国内に自社工場を持っており、予備部品のストック、オーバーホール等が迅速に行われること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年３月19日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒460―0001 愛知県名古屋市中区三の丸４－１－１　独立行政法人国立病院機構名古屋医療センター企画課長　谷上　勇人　電話052―951―1111　内線2220
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年２月２日から令和３年３月19日まで。
⑵　交付場所　上記２⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Total cardiopulmonary system 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The system base can be configured to install up to 10 roller pumps and incorporate a dedicated centrifugal pump device
②　A roller pump is a stand-alone type allowing a versatile layout onto the system as desired. A roller pump head is available in three diameter sizes : ϕ150, ϕ100, and ϕ75 to set up the optimal system for neonate to adult
③　Occlusion shall be adjusted by dial operation with an occlusion locking mechanism
④　A pump remote controller (operation panel) for both roller and centrifugal pump is a stand-alone, enabling an installment on anywhere on the system base as well as any position on the system with a use of a special holder to complete the optimal system layout for each case
⑤　The system can be driven by a centrifugal pump. In addition, the safety operation can be assured with an auto-clamping function in a case of activated alarms such as backflow, bubble detection
⑥　Monitoring system is fully available and its system provides a high level of safety for the operation of the whole system. LCD touch panel installed must have an excellent visibility and operability
⑦　Not only digital numerical but also analog bar graph indication is available on pressure monitor, to get an easier notice to pressure changes like pulsation
⑧　When pressure monitor, level sensor, and bubble sensor detect unusual operation, it is selectable to receive alarm notification only or interlocking with the operating pump
⑨　P／S control (Primary-Secondary setting for 2 linked-pump operation) is easy to be set, leading to speedy blood cardioplegia infusion as well
⑩　The system must be operated interconnectedly by BF of roller pump and one of venous line
⑪　The upper limit function control not to be over the designated flow rate, reducing possible risk on negative pressure on oxygenator
⑫　An electronic blender unit must be used with the system and the data (oxygen gas concentration and flow rate) can be available in output
⑬　The system base is equipped with an emergency power supply in order to secure a backup system in case of a power failure like blackout
⑭　The system can be operated interconnectedly with a myocardial protection system called “TRUSYS”
⑮　The manufacturing company has its own factory in Japan, where spare parts are stored. Then after-service such as regular check-up, parts replacement and overhaul service is available promptly
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M. March 19, 2021
⑹　Contact point for the notice : Hayato Tanigami, Director, Accounting Division, National Hospital Organization Nagoya Medical Center, 4―1―1 Sannomaru, Naka-ku, Nagoya-shi, Aichi-ken, 460―0001, Japan. TEL 052―951―1111 ext. 2220

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和３年２月１日
国立大学法人東京大学総長　五神　　真
◎調達機関番号　415　◎所在地番号　13
○第10号
1　調達内容
⑴　品目分類番号　71、27
⑵　導入計画物品及び数量　ＨＰＣＩ共用ストレージ東拠点ストレージシステム機材2022年度保守　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和３年度３月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　導入説明書「東拠点ストレージシステム機材一覧」にて示す機材全てを保守対象機材とすること。
Ｂ　機材全体としての保守機関は、令和４年４月１日から令和５年３月月31日までであるが、機材の設置時期に応じて保守機関の開始日が異なるので留意すること。なお、令和４年12月29日から令和５年１月３日までの年末年始期間と土日祝日は対象外である。
Ｃ　ＤＤＮ社のストレージシステムの保守作業を実施することについて、該当メーカの了承を得ていること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年３月５日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒113―8658　東京都文京区弥生２―11―16　東京大学情報システム部情報戦略課会計チーム　志村　正規　電話03―5841―2705
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年２月１日から令和３年３月５日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年２月８日14時00分
⑵　開催場所　Zoomによるオンライン説明会
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the services to be procured : 71, 27
⑵　Nature and quantity of the services to be required : FY2022 maintenance of the storage system equipment at the East Hub of HPCI Shared Storage 1 Set
⑶　Type of the procurement : Require