6　Summary
⑴　Classification of the products to be procured : 14
⑵　Nature and quantity of the products to be rent : Kobe Academic Information System for Education and Research 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　Kobe Academic Information System for Education and Research supports the education and research activities for the approximately 36,000 university members.
B　The subsystems and services provided by this system includes the followings : Integrated user management system, User authentication system, Authentication federation system, Personal mail system, Education system, KAISER network system, etc.
C　Integrated user management system provides the user account management function for all university members and the management functions for the above services.
D　User authentication system and Authentication federation system provide the user authentication function to ICT systems in the university.
E　Personal mail system provides the electronic mail function for all university members.
F　The educational system, which should be used with the Bring Your Own Device (BYOD), provides a machine-independent educational support environment. BYOD is a program that students using their personal devices for educational purposes in Kobe University.
G　KAISER network system is used to connect the equipments of this system which is integrated with the other university ICT systems through the Kobe University Hyper Academic Network (KHAN).
⑸　Time limit for the submission of the requested material : 17 : 00　11 May, 2021
⑹　Contact point for the notice : Nagahama Yorihisa, Accounting and Procurement Division, Finance Department, Kobe University, 1―1 Rokkodai-cho Nada-ku Kobe-shi 657―8501 Japan, TEL 078―803―5404

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和３年４月６日
国立大学法人岡山大学長　槇野　博史
◎調達機関番号　415　◎所在地番号　33
○第１号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　手術映像支援システム　１式
⑶　調達方法　購入等
⑷　導入予定時期　令和３年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　手術映像記録・配信システム
a　術野カメラは４Ｋ映像対応であり、電動ズーム・旋回・回転・集音に対応していること。
b　手術室内の４Ｋモニタは天井懸垂式及び壁面埋込式であること。
c　手術室壁面に手術用内視鏡及び顕微鏡用の接続端子を設け、４Ｋ映像入出力の対応を行うこと。
d　ペンダントに生体情報モニタの接続端子を設けること。
e　映像切替システムは４Ｋ対応であり、切替・カメラ・録画を１台の操作用タッチパネルにて容易に操作できること。
f　麻酔記録システムから取得した手術オーダー情報を操作用タッチパネルに表示できること。また、本情報は既設ライブモニタリング端末にも共有し情報の表示が行えること。
g　既存のエンコーダー装置、手術映像記録サーバ及び病院情報システム（ＨＩＳ）と接続し、サーバ録画並びにＨＩＳ端末でのライブ映像及び記録映像の視聴について現行の運用を継承すること。
h　術野カメラ及び天井懸垂式モニタは既存のアームに設置を行うこと。
B　可搬型手術用顕微鏡
a　ヘッドアップサージャリーが行えること。
b　４Ｋ－３Ｄの高画質映像出力対応であること。
c　特殊光観察が行えること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年５月24日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒700―8558　岡山市北区鹿田町２－５－１　岡山大学病院経営・管理課調達グループ政府調達担当　中本　誠　電話086―235―7540
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年４月６日から令和３年５月24日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年４月９日10時00分
⑵　開催場所　岡山大学医学部管理棟小会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Operative Imaging Control System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Surgical video recording and distribution system
a　The operative field camera should be 4K video, and support electric zoom／pan／tilt／rotation／sound collection.
b　The 4K display in the operating room should be ceiling suspension type and wall embedding type.
c　The video signal terminals for surgical endoscopes and microscopes on the wall should support 4K video input and output.
d　The video signal terminals on the pendant can be utilized as the connection to the patient monitors.
e　The video switching system should be 4K compatible, and using a touch panel controller should facilitate the operation of the system.
f　For the touch panel for operation, surgery order information obtained from anesthesia recording system should be displayed on the existing live monitoring terminal.
g　Connect to the existing encoder device and surgical video recording server, and inherit the current operation for server recording and live distribution to the hospital information system (HIS), playback management of recorded video, and cooperation.
h　The operative field camera and the ceiling suspension type display should be mounted on the existing ceiling arm.
B　Portable surgical microscope
a　Heads-up surgery can be performed.
b　4K－3D video output for high image quality is available.
c　The microscope must offer special light observation mode.
⑸　Time limit for the submission of the requested material : 17 : 00　24 May, 2021
⑹　Contact point for the notice : NAKAMOTO Makoto, Procurement Section, Management Division, Okayama University Hospital, 2―5―1 Shikata-cho Kita-ku Okayama-shi 700―8558 Japan, TEL 086―235―7540