・The spectrometer should have an internal computer enabling measurement even during the lost of remote connection and／or hang up of host computer.

c　The following functionalities are required for the solid-state NMR probe :

・The probe with low ^{19}F background should be equipped.

・The MAS probe which is capable of temperature control in the range of −100℃ to 200℃ or higher and high precision measurement simultaneously should be equipped.

d　The following functionalities are required for the liquid-state NMR probe :

・The system needs to have the automated tuning system.

・The probe which is capable of temperature control in the range of −100℃ to 150℃ or higher should be equipped.

・The probe and magnetic amplifier which is capable of generating magnetic field gradient should be equipped.

C　300 MHz nuclear magnetic resonance system

a　Spectrometer should be equipped with the following functionalities :

・The measurement control and analysis system should be fully controlled from several remote access points.

・The spectrometer should have an internal computer enabling measurement even during the lost of remote connection and／or hang up of host computer.

・The spectrometer should be compatible with the superconducting magnet (JEOL NM−Z101371SCM30) which has been already installed in this university. The connection and performance check should be also performed.

b　The following functionalities are required for the liquid-state NMR probe :

・The probe which is capable of temperature control in the range of −100℃ to 150℃ or higher should be equipped.

・The probe and magnetic amplifier which is capable of generating magnetic field gradient should be equipped.

⑸　Time limit for the submission of the requested material : 17 : 00　24 May, 2021

⑹　Contact point for the notice : Shigehiro Akiyama, Procurement Services Office, Finance Department, Tohoku University, 2−1−1 Katahira Aoba-ku Sendai-shi 980−8577 Japan, TEL 022−217−4869

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和３年４月21日

国立大学法人東京大学総長　藤井　輝夫

◎調達機関番号　415　◎所在地番号　13

○第３号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　遠隔オンラインビデオ通話による臨床情報評価システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和３年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　オンラインビデオ通話が可能であり、施設側一被験者側の双方で、ビデオ映像転送のタイムラグが少なく、円滑な会話が可能であること。

Ｂ　オンラインビデオ通話のデータをクラウド上に記録・保管可能であること。

Ｃ　臨床情報をウェブブラウザより記入しクラウド上に保存可能であること。

Ｄ　タブレット端末などを用いた臨床指標の電子的取得を可能とするアプリケーションとのオンラインビデオ通話・クラウドデータ保管の連携が可能であり、ＡＰＩ仕様が公開されていること。

Ｅ　データベースシステムが、ＨＩＰＡＡなどの医療情報の取り扱いに関する十分な安全性を備えていること。

Ｆ　クラウドに保管されたデータに対してユーザがアクセスしデータ取得できること、およびその権限設定ができること。

Ｇ　使い方の講習を参画機関に対して行うこと、アフターサポートやメンテナンスの対応が可能であること。

Ｈ　ＥＤＣ機能等との連携など、機能拡張性が十分に備えられたシステムであること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和３年６月７日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒113−8655 東京都文京区本郷７−３−１　東京大学医学部附属病院事務部管理課経理チーム　福代　美樹　電話03−5800−8616

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和３年４月21日から令和３年６月７日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和３年４月28日10時00分

⑵　開催場所　東京大学医学部附属病院管理・研究棟２階第３会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Clinical information evaluation system based on remote online video communication 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Online video communication must be possible. To ensure smooth converation, the time lag between the facility side and the subject side should be kept to minimun.

B　Data gathered through online video communication can be recorded and stored in a cloud-based server.

C　Clinical information can be entered through a web browser and stored in the cloud-based server.

D　Online video communication data and cloud data storage can be linked to applications that enable electronic acquisition of clinical information using tablet devices, etc. The API specifications for the system must also be publishable.

E　The database system has adequate security for the handling of medical information (e.g. HIPAA compliance).

F　Users must be able to access and retrieve data stored in the cloud server, and must be able to set access permissions.

G　Training on how to use the system should be provided to facilities participating in the project. After-sales support and maintenance should also be available.

H　The system must have sufficient functional expandability, including linkage with EDC functions.

⑸　Time limit for the submission of the requested material : 17 : 00　7 June, 2021

⑹　Contact point for the notice : Miki Fukuyo, Secretariat Procurement Division, The University of Tokyo Hospital, 7−3−1 Hongo Bunkyo-ku Tokyo 113−8655 Japan, TEL 03−5800−8616

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和３年４月21日

大学共同利用機関法人

情報・システム研究機構長　藤井　良一

◎調達機関番号　416　◎所在地番号　13

○第１号―情報研

１　調達内容

⑴　品目分類番号　71、29

⑵　導入計画物品及び数量

令和３年度学術情報ネットワーク（ＳＩＮＥＴ６）用相互接続環境　一式

⑶　調達方法　購入等

⑷　導入予定時期　令和４年度３月以降