d　冷媒を入れるチャンバーは－196℃～０℃の範囲で温度調節が可能なこと。
e　液体窒素の貯蔵容器にベークアウトの機構を備えていること。
C　グロー放電表面改質装置
a　大気を利用するチャンバーと化合物の蒸気を利用するチャンバーを備えること。
b　安全に化合物の蒸気供給が行えること。
c　チャンバー間でクロスコンタミを防ぐ機構を備えていること。
d　操作がタッチスクリーンで行えること。
e　大気圧解放の速度コントロールが可能なこと。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年６月７日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒819―0395 福岡市西区元岡744 国立大学法人九州大学財務部調達課役務調達係　大木富守郎　電話092―802―2367

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年５月６日から令和３年６月７日まで。
⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年５月14日13時30分
⑵　開催場所　九州大学病院キャンパス　総合研究棟２階　205号室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 24
⑵　Nature and quantity of the products to be purchased : High acceleration voltage cryo-electron microscope 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　High acceleration voltage cryo-electron microscope
a　The electron gun is a cold cathode field emission type electron gun with an acceleration voltage of 200kV.
b　The sample introduction method is a cartridge introduction method, and it has a function to automatically transport the frozen sample into the column.
c　Sample loading and unloading can be performed while maintaining the vacuum state of the sample transfer chamber by equipping the sample transfer mechanism with an airlock valve.
d　Must have an in-column energy filter.
e　The amount of electron beam dose can be continuously modulated by a high-speed electrostatic shutter.
f　Equipped with an direct electron detection type camera, the number of effective pixels of the detector is 4,000×4,000 pixels or more, and the frame rate is 1500 images／second or more in full frame.
B　Automatic rapid freezing device
a　The temperature of the environmental chamber can be adjusted in the range of ＋4℃ to ＋60℃.
b　Must have a mechanism to humidify the environmental chamber.
c　The operation must be a touch screen or a foot switch.
d　The temperature of the chamber that holds the refrigerant can be adjusted in the range of －196℃ to 0℃.
e　The liquid nitrogen storage container must have a bakeout mechanism.
C　Glow discharge surface modifier
a　Provide a chamber that uses the atmosphere and a chamber that uses the vapor of the compound.
b　Safe vapor supply of compounds.
c　Must have a mechanism to prevent cross contamination between chambers.
d　Operation can be performed on the touch screen.
e　The speed of atmospheric pressure release can be controlled.
⑸　Time limit for the submission of the requested material : 17 : 00　7 June, 2021
⑹　Contact point for the notice : OHKI Toshiro, Service Procurement Section, Procurement Division, Finance Department, Kyushu University, 744 Motooka Nishi-ku Fukuoka-shi 819―0395 Japan, TEL 092―802―2367

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和３年５月６日
国立大学法人鹿児島大学
契約担当役代理財務部長　饗場　　厚

◎調達機関番号　415　◎所在地番号　46

○第１号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　全身用ＭＤＣＴ診断システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和３年度２月以降
⑸　調達に必要とされる基本的な要求要件
A　逐次近似画像再構成法を有すること。
B　造影剤自動注入装置を有すること。
C　ＣＴ透視機能を有すること。
D　本システムによる診療を安全かつ円滑に行う上で、必要不可欠な附属周辺機器を有すること。
E　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、読影レポートシステム、及び画像ネットワーク（ＰＡＣＳ）、その他関連機器と接続可能であること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年６月25日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒890―8520 鹿児島市桜ケ丘８―35―１　鹿児島大学病院経理調達課調達係　島名　冬美　電話099―275―5045

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年５月６日から令和３年６月25日まで。
⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年６月８日10時00分
⑵　開催場所　鹿児島大学病院管理棟４階第一会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : The whole body MDCT diagnostic scan system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The MDCT system must be equipped with successive approximations image reconstruction method.
B　The MDCT system must be equipped with an automated contrast injector.
C　The MDCT system must be equipped with the CT fluoroscopy function.
D　The MDCT system must have peripheral devices that facilitate smooth and safe operation.
E　The MDCT system must be able to connect with the existent HIS, RIS, image interpreting system, PACS and other related equipment.
⑸　Time limit for the submission of the requested material : 17 : 00　25 June, 2021
⑹　Contact point for the notice : SHIMANA Fuyumi, Procurement Section, Administrative Division, Kagoshima University Hospital, 8―35―1 Sakuragaoka Kagoshima-shi 890―8520 Japan, TEL 099―275―5045