2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年10月22日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学財務部経理課　嶋津　美穂　電話0776―61―3111　内線3145
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年９月１日から令和３年10月22日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年９年６日14時00分
⑵　開催場所　福井大学医学部管理棟中会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 14
⑵　Nature and quantity of the products to be purchased : Advanced Medical Information Network Integration Facilities 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　These facilities should be integrated network facilities that operate medical information systems, medical systems, medical devices, IP phones (VoIP), video conferencing, medical images and video distribution.
B　The network operated by these facilities must be able to use at least IEEE802.3, IEEE802.11, IEEE802.15.1, and IEEE802.15.4.
C　In these facilities, IPv4 and IPv6 can be used for the operation of the IP protocol.
D　The IEEE802.3 wired network of these facilities can be operated at a speed of 1 Gbps or higher.
E　The wireless networks such as IEEE802.11 and IEEE802.15 of these facilities shall have a function to prevent radio wave interference.
F　High-precision indoor positioning can be performed using the wireless network of these facilities.
G　These facilities should be facilities that can promote medical DX as an environment for Society 5.0 in medical care.
H　In these equipment, the connection of medical equipment, etc. shall be equipment that takes into consideration the priority and stability of information transmission.
I　These facilities must have network monitoring, equipment management, and automatic recovery support functions in the event of a failure.
⑸　Time limit for the submission of the requested material : 17 : 00　22 October, 2021
⑹　Contact point for the notice : Shimazu Miho, Procurement Section, Accounting Division, Finance Department, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3145

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和３年９月１日
国立大学法人島根大学長　服部　泰直
◎調達機関番号　415　◎所在地番号　32
○第１号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　放射線治療システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和４年度３月以降
⑸　調達に必要とされる基本的な要求要件
A　定位放射線照射、強度変調放射線治療（ＩＭＲＴ）などの高精度治療が可能であること。　B　治療装置本体に画像誘導放射線治療（ＩＧＰＴ）を可能とする照射位置照合システム及び関連装置を備えていること。　C　画像誘導放射線治療システムは、体表面、骨構造、腫瘍の位置情報による位置照合が可能なシステムであること。　D　動態追跡照射法が可能なシステムであること。　E　多発脳腫瘍に対する脳定位放射線治療が可能なシステムであること。　F　高精度治療の品質管理に対応した検証用ファントム、線量計、解析装置など関連システムを有すること。　G　既存のシステムと統合可能な放射線治療計画システム、治療情報管理システム及び治療支援システム等の周辺機器などを有すること。　H　患者情報保護のため、セキュリティ対策が施されていること。　I　本院の医用画像管理システム・放射線情報システム・病院情報システムと接続し、画像情報・オーダー情報等の通信ができること。　J　装置の故障に対して連絡体制が整備されており、迅速な障害復旧ができるサービス体制であること。　K　必要な研修及び訓練を継続的に提供すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和３年10月18日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒693―8501 出雲市塩冶町89―１　国立大学法人島根大学医学部会計課契約第二係　森山　　洋　電話0853―20―2043
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和３年９月１日から令和３年10月18日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和３年９月10日14時00分
⑵　開催場所　島根大学医学部本部棟４階第二会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Radiation therapy system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system should be designed to perform high precision radiotherapy, such as stereotactic radiotherapy (SRT) and intensity modulated radiotherapy (IMRT). B　The system should be equipped with imaging devices especially for dynamic targeting image-guide radiation therapy (IGRT). C　The image-guided radiation therapy system should be possible to collate based on body surface, bone structure, and tumor position information. D　The system should have a respiratory synchronizing or gating apparatus. E　The system should be possible to perform a brain stereotactic radiation therapy for multiple brain tumors. F　The system should have peripheral devices and accessories, including verification phantoms, dosimeters and analyzing systems, necessary to perform quality assurance protocols supporting high precision radiotherapy. G　The system should include