2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑹の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和６年３月14日午後５時（郵送の場合は必着のこと。）
⑵　提出先　〒305―8601 茨城県つくば市観音台２－１－９　農林水産技術会議事務局筑波産学連携支援センター総務課用度係　斉藤いずみ　電話029―838―7217
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和６年１月22日から令和６年２月９日
⑵　交付場所　上記２⑵に同じ。
⑶　交付方法　上記３⑵交付場所にて交付及び調達ポータル上にてダウンロード可能。
（https://www.p-portal.go.jp/pps-web-biz/UAA01/OAA0101）
4　説明会の開催　本公表に基づく導入説明会を導入説明書交付後別途開催する。
⑴　開催日時
①　令和６年２月９日午後１時30分
②　令和６年２月９日午後２時30分
③　令和６年２月９日午後３時30分
⑵　開催場所　農林水産技術会議事務局筑波産学連携支援センター本館３階入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 14
⑵　Nature and quantity of the products to be lease :
⒜　Scientific Computing System, 1 Unit
⒝　Network Service System, 1 Unit
⒞　Network Library System, 1 Unit
⑶　Type of the procurement : lease
⑷　Purpose of the procurement : The purpose of this procurement is to update the research information and computer systems for research institutions related to the Ministry of Agriculture, Forestry and Fisheries for research promotions.
⒜　The scientific computing system consists of supercomputer system to provide large scale parallel processing, machine learning, scientific applications, numerical database.
⒝　The network service system is responsible for user authentication management, provision of access routes from the Internet, and provision of operational infrastructure for scientific computing system and network library system.
⒞　The network library system consists document information and the books information.
⑸　Time-limit for the submission of requested material : 5 : 00 P.M., 14 March, 2024
⑹　Contact point for the notice : SAITO Izumi, Procurement section, General Affairs Division, Tsukuba Business-Academia Cooperation Support Center, Agriculture, Forestry and Fisheries Research Council Secretariat, 2―1―9 kannondai Tsukuba city Ibaraki 305―8601 Japan. TEL 029―838―7217

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和６年１月22日
国立大学法人東北大学理事　植木　俊哉
◎調達機関番号　415　◎所在地番号　04
○第13号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　超音波骨量測定装置　八式　外
⑶　調達方法　購入等
⑷　導入予定時期　令和６年度９月以降
⑸　調達に必要とされる基本的な要求要件
A　超音波骨量測定装置　八式
ａ）測定部、操作部、結果表示部及びプリンタを備えた一体型であること。
ｂ）超音波透過法による骨内伝搬速度と超音波反射法による骨幅測定を併用した測定ができること。
ｃ）超音波伝搬速度及び骨幅の他に、年代別の検査結果の偏差、骨梁面積率及び骨年齢判定などのデータが検査現場で記録できること。
ｄ）計測した検査数値を電子的に保存できること。
B　聴力検査機器　八式
ａ）本体は可搬できること。
ｂ）カラー液晶画面を有すること。
ｃ）以下の検査機能を有すること。
・標準純音聴力検査
・選別聴力検査
・幼児に対する聴力検査
ｄ）外部とデータ通信できること。
ｅ）音圧レベルが80ｄB以上になると警告表示される機能を有すること。
C　心電計　五式
ａ）本体に液晶画面、記録器を備え、ＡＣ100Ｖ及び内蔵バッテリでの駆動ができること。
ｂ）12誘導心電図解析機能を有し、心電図波形に解析所見を付加できること。さらに10電極法測定から、導出18誘導の演算導出ができること。
ｃ）心電図波形及び解析所見を検査現場で記録できること。
ｄ）ｃ）に示した計測した心電図波形及び解析所見を電子的に保存できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和６年３月８日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒980―8577 仙台市青葉区片平２－１－１　東北大学財務部調達課調達第一係長　大友　将充　電話022―217―4869
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和６年１月22日から令和６年３月８日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和６年１月29日10時00分
⑵　開催場所　オンライン開催とする。（詳細は、導入説明書による。）
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Ultrasonic Done Measurement Device 8 Sets and Others
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Ultrasonic bone measurement device
ａ）It must be an integrated type that includes a measurement section, operation section, result display section, and printer.
ｂ）Measurement is possible by combining intraosseous propagation velocity using ultrasonic transmission method and bone width measurement using ultrasonic reflection method.
ｃ）In addition to speed of sound on ultrasound velocity and bone width, data such as deviation of examination results by age, trabecular bone area ratio, and bone age determination can be recorded at the examination site.
ｄ）The measured inspection values that shown above c) must be able to be stored electronically.
B　Audiometric system
ａ）The main unit must be portable.
ｂ）Must have a color LCD screen.
ｃ）Must have the following inspection functions.
・Standard pure tone audiometry.
・Screening audiometry.
・Audiometry for infants.
ｄ）Ability to communicate data externally.
ｅ）Must have a function that displays a warning when the sound pressure level exceeds 80 dB.