４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和７年４月21日14時00分
⑵　開催場所　金沢大学学術メディア創成センター２階プレゼンテーション室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 15, 28
⑵　Nature and quantity of the products to be rent : Kanazawa University Academic Integrated Network System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement : This system is an information and communication infrastructure that supports academic education and research at our university. To provide safe and high-speed Internet communication and intra-campus communication to more than 20,000 devices, it consists of core switch group, branch switch group, access switch group, data center switch group, security system group and server device group.
A．The core switch group will connect and forward communication between the Internet, firewall, data center and any branch switches. This group will also provide router functions such as VRF (virtual routing and forwarding) and dynamic routing (BGP, OSPF etc.).
B．The branch switch group will be placed at each building and aggregate access switches placed at each floor of the building and be connected to core switches.
C．The access switch group will be placed at each floor of buildings and aggregate user devices connected via LAN outlets, and provide security functions such as terminal authentication, broadcast suppression functions.
D．The data center switch group will constitute the network in our university data center. It aggregates the server devices in the data center, and forwards its traffic.
E．The security system group will consist of firewall, sandbox, log analyzer, etc., and provides any network security functions such as anti-virus, application control, intrusion prevention, cloud access control, anti-spam, anti-phishing, etc.
F．The server device group is the virtual server environment which will provide network management functions such as monitoring server, name server, authentication server, log server.
G．The system will be protected from faults, power failures and security risks to ensure stable operation. In case of system faults, quick and accurate maintenance services will be required.
H．Management and operation of the system must be done with labor-saving, energy-saving and space-saving.
⑸　Time limit for the submission of the requested material : 17 : 00　20 May, 2025
⑹　Contact point for the notice : Masashi Kobayashi, Procurement management Section, Financial Management Division, Finance Department, Kanazawa University, Kakuma-machi Kanazawa-shi 920―1192 Japan, TEL 076―264―6162

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
令和７年４月16日
国立大学法人浜松医科大学
理事　三沼　仁
◎調達機関番号　415　◎所在地番号　22
○第１号
１　調達内容
⑴　品目分類番号　24
⑵　導入計画物品及び数量　創薬支援分析システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　令和８年３月頃
⑸　調達に必要とされる基本的な要求要件
A　精密質量分析計　１式
a　電場型フーリエ変換による高分解能質量部と非対称イオンミラーによる高分解能質量分析部で構成されていること。
b　調整可能な加熱エレクトロスプレーイオン化（ＨＥＳＩ）プローブが内蔵されていること。
c　質量測定範囲は、m／z 40～8,000の範囲を含むこと。
d　装置内部には、ポジティブモードおよびネガティブモードでのリアルタイムでのロックマス機能を有するキャリブラントソースを装着していること。
e　マトリックス支援レーザー脱離イオン化（ＭＡＬＤＩ）による質量分析イメージングが可能なイオン源システムを備えること。
B　ナノフロー液体クロマトグラフ　１式
a　送液設定範囲は１nL／min～100μL／minを１nL間隔で設定できること。
b　試料の注入量精度は0.05～2.00μLの間で最大5.0％ＲＳＤ以下であること。
c　「A　精密質量分析計」と「B　ナノフロー液体クロマトグラフ」は、「C　装置制御・データ解析部」を介して制御し、連動させることができる機能を有すること。
C　装置制御・データ解析部　１式
a　制御用コンピュータシステムと解析用コンピュータシステムを備えること。
b　解析用コンピュータシステムは、以下のソフトウェアを備えること。
・ＭＳ及びＭＳ／ＭＳのスペクトル解析ができる機能を有するソフトウェア。
・プロテオーム解析ソフトウェア。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和７年５月19日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒431―3192 浜松市中央区半田山１―20―１　浜松医科大学会計課用度係　中島　大智　電話053―435―2553
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和７年４月16日から令和７年５月19日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和７年４月23日11時30分
⑵　開催場所　浜松医科大学管理棟２階第二会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 24
⑵　Nature and quantity of the products to be purchased : Drug Discovery Support Measurement System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　High-Resolution Mass Spectrometer 1 unit
a　The system shall consist of a high-resolution mass analyzer based on electrostatic Fourier transform and a high-resolution mass analyzer utilizing an asymmetric ion mirror.
b　The system shall include an adjustable heated electrospray ionization (HESI) probe.
c　The system shall have a mass measurement range covering m／z 40 to 8,000.
d　The system shall be equipped with a calibrant source that enables real-time lock mass functionality in both positive and negative ion modes.
e　The system shall be equipped with an ion source system capable of mass spectrometry imaging by matrix-assisted laser desorption／ionization (MALDI).