5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

(1)　Classification of the products to be procured : 22, 31

(2)　Nature and quantity of the products to be purchased : 3-Tesla MRI System 1 Set

(3)　Type of the procurement : Purchase

(4)　Basic requirements of the procurement :

A　The MRI system has a static magnetic field strength of 3 Tesla.

B　In order to acquire high-speed and high-quality images, it should be equipped with a specialized imaging coil for brain imaging.

C　It should feature high magnetic field homogeneity and high gradient performance.

D　In addition, the obtained images can be stored in the medical image (DICOM) server.

E　The peripheral equipment (image processing device, etc.) to perform the examination safely and efficiently.

(5)　Time limit for the submission of the requested material : 17 : 00　27 April, 2026

(6)　Contact point for the notice : Takashi Fukumoto External Funding Section, Accounting Division, Kyoto University Administration Office (Letters), Main Campus, 36―1 Yoshida-honmachi, Sakyo-ku, Kyoto-shi 606―8501 Japan, TEL 075―753―6905

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和８年３月11日

国立大学法人九州大学総長　石橋　達朗

◎調達機関番号　415　◎所在地番号　40

○第７号

1　調達内容

(1)　品目分類番号　22、31

(2)　導入計画物品及び数量

手術器材搬送システム　一式

(3)　調達方法　購入等

(4)　導入予定時期　令和８年度３月以降

(5)　調達に必要とされる基本的な要求要件

①　手術器材管理システム　１式

②　滅菌管理システム　１式

③　自動搬送ロボット管理システム　１式

2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1)　資料等の提供期限　令和８年４月27日17時00分（郵送の場合は必着のこと。）

(2)　提供先　〒819―0395　福岡市西区元岡744　国立大学法人九州大学財務部調達課特定調達係　迫田　正浩　電話092―802―2363

E-Mail：zactokutei@jimu.kyushu-u.ac.jp

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

(1)　交付期間　令和８年３月11日から令和８年４月27日まで。

(2)　交付場所　電子媒体にて交付するため、交付を希望する者は上記２(2)に記載のメールアドレスに連絡すること。

4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

(1)　Classification of the products to be procured : 22, 31

(2)　Nature and quantity of the products to be purchased : Surgical Instrument Transport System 1 Set

(3)　Type of the procurement : Purchase

(4)　Basic requirements of the procurement :

①　Surgical Instrument Management System 1 set

②　Sterilization Management System 1 set

③　Automated Guided Vehicle (AGV) Management System 1 set

(5)　Time limit for the submission of the requested material : 17 : 00　27 April, 2026

(6)　Contact point for the notice : SAKODA Masahiro, Specified Procurement Section, Procurement Division, Finance Department, Kyushu University, 744 Motooka Nishi-ku Fukuoka-shi 819―0395 Japan, TEL 092―802―2363

E-Mail : zactokutei@jimu.kyushu-u.ac.jp

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和８年３月11日

国立大学法人長崎大学長　永安　　武

◎調達機関番号　415　◎所在地番号　42

○第７号

1　調達内容

(1)　品目分類番号　22、31

(2)　導入計画物品及び数量　密封小線源治療システム　一式

(3)　調達方法　購入等

(4)　導入予定時期　令和８年度11月以降

(5)　調達に必要とされる基本的な要求要件

A　導入装置は高エネルギー治療棟の密封小線源治療室に設置すること。

B　当院既存の画像ネットワークおよび放射線治療情報システムと接続し、治療計画データ・画像データの送受信が可能であること。

C　密封小線源治療に対応した治療計画装置を備えること。

D　国際標準に準拠した高線量率密封小線源治療が実施できること。

E　治療中の安全性を確保するためのインターロック機構、線源位置精度管理機能、緊急停止機能を備えること。

F　当院既存の治療ワークフローと整合性を持ち、治療計画から照射までの一連の工程が円滑に行えること。

2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1)　資料等の提供期限　令和８年４月27日17時00分（郵送の場合は必着のこと。）

(2)　提供先　〒852―8501　長崎市坂本１―７―１　長崎大学病院経理調達課物品担当　本多　克也　電話095―819―7218

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

(1)　交付期間　令和８年３月11日から令和８年４月27日まで。

(2)　交付場所　上記２(2)に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

(1)　開催日時　令和８年３月17日10時30分

(2)　開催場所　長崎大学病院多目的研修室（中央診療棟４Ｆ）

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

(1)　Classification of the products to be procured : 22, 31

(2)　Nature and quantity of the products to be purchased : High-Dose-Rate Brachytherapy System 1 Set

(3)　Type of the procurement : Purchase

(4)　Basic requirements of the procurement :

A　The HDR brachytherapy system shall be installed in the designated brachytherapy treatment room.

B　The system shall be connected to the existing Radiology Information System (RIS) and Picture Archiving and Communication System (PACS) to enable the exchange of treatment and imaging data.

C　The system shall include a dedicated brachytherapy treatment planning system.

D　The system shall be capable of performing high-dose-rate (HDR) brachytherapy in accordance with international standards.

E　The system shall include safety interlock mechanisms, source position accuracy management, and emergency stop functions.

F　The system shall integrate with the existing clinical workflow from treatment planning to irradiation.

(5)　Time limit for the submission of the requested material : 17 : 00　27 April, 2026

(6)　Contact point for the notice : HONDA Katsuya, Accounting and Procurement Division, Nagasaki University Hospital, 1―7―1 Sakamoto Nagasaki-Shi 852―8501 Japan, TEL 095―819―7218