2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　令和８年７月13日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒770―8503 徳島市蔵本町３丁目18番地の15　国立大学法人徳島大学経理部蔵本会計課調達係長　三好　敏子　電話088―633―9556
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　令和８年６月11日から令和８年７月13日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　令和８年６月17日11時00分
⑵　開催場所　徳島大学医学部第五会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 24
⑵　Nature and quantity of the products to be purchased :　Ultra-High-Sensitivity and High-Speed Mass Spectrometry System 1set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Mass Spectrometer
a　Must consist of a high-resolution mass analyzer based on electrostatic field Fourier trans form and a high-resolution mass analyzer using an asymmetric ion mirror.
b　Must support electrospray ionization (ESI).
c　Must have a function to automatically control the ion population inside the mass spectrometer.
d　The mass range must include m／z 40～6,000 for the electrostatic field Fourier transform analyzer and m／z 40～6,000 for the asymmetric ion mirror analyzer.
e　Scan speed must be at least 40 Hz for the electrostatic field Fourier transform analyzer and at least 270 Hz for the asymmetric ion mirror analyzer.
f　The electrostatic field Fourier transform analyzer must have a resolution of at least 480,000 (FWHM) at m／z 200. The asymmetric ion mirror analyzer must have a resolution of at least 80,000 (FWHM) at m／z 524, and at least 100,000 (FWHM) at m／z 138 when switching modes.
g　Must have an ion source compatible with a nano LC column that integrates emitter, column, and column heater.
B　Nano flow Liquid Chromatograph
h　The wetted parts must be made of non-metallic, biocompatible materials suitable for proteomics analysis.
i　Must support binary high-pressure gradients.
j　System pressure tolerance must be at least 150 MPa.
k　The nano flow pump must support a flow rate range from 1 nL／min to 100 μL／min.
l　Must include a degasser capable of online degassing of mobile phases.
m　The autosampler must accommodate at least four well plates.
n　The sample temperature control range must include 4～40° C.
C　LCMS Control Computer System
o　The OS must provide functionality equivalent to or higher than Microsoft Windows 11.
p　The CPU must have performance equivalent to or greater than Intel 12-core 4.9 GHz.
q　Memory must be at least 32GB.
r　Must include a hard disk drive with a physical capacity of at least 4TB.
s　Must include a liquid crystal color display with a diagonal size of at least 27 inches.
t　Must include software capable of controlling both the mass spectrometer and the nano flow liquid chromatograph.
⑸　Time limit for the submission of the requested material : 17 : 00　13 July, 2026
⑹　Contact point for the notice : Toshiko Miyoshi Procurement Assistant Manager, Kuramoto Accounting Division, Accounting Department, National University Corporation, Tokushima University, 3―18―15 Kuramoto-cho　Tokushima-shi　770―8503 Japan, TEL 088―633―9556

## 意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和８年６月11日
個人情報保護委員会事務局総務課長
戸梶　晃輔
◎調達機関番号　025　◎所在地番号　13
1　調達内容
⑴　品目分類番号　71、27
⑵　購入等特定役務及び数量　第三期監視・監督システムの運用・アプリケーション保守業務　一式
2　意見の提出方法
⑴　意見の提出期限　令和８年７月１日正午（郵送の場合は必着のこと。）
⑵　提出先　〒105―0001 東京都港区虎ノ門２―２―３　虎ノ門アルセアタワー12階　個人情報保護委員会事務局総務課会計係　連絡先：電話03―6457―9619
3　仕様書案の交付
⑴　交付期間　令和８年６月11日から令和８年７月１日まで
⑵　交付場所　上記２⑵に同じ
4　Summary
⑴　Classification of the services to be procured : 71, 27
⑵　Nature and quantity of the services to be required : Operation and application maintenance for the Third Phase Kanshikantoku System, 1 set
⑶　Time-Limit for the submission of comments : 12 : 00 p.m, 1, July, 2026
⑷　Contact point for the notice : Accounts subsection, General Affairs Division, Secretariat of Personal Information Protection Commission, 2―2―3 Toranomon Minato-ku, Tokyo 105―0001 Japan, TEL 03―6457―9619.

## 意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
令和８年６月11日
支出負担行為担当官
関東信越国税局総務部次長　平賀　純二
◎調達機関番号　015　◎所在地番号　11
1　調達内容
⑴　品目分類番号　133
⑵　購入等物品及び数量　令和８事務年度「申告案内コールセンター」の運営業務委託　一式
2　意見の提出方法
⑴　意見の提出期限　令和８年７月２日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒330―9719 埼玉県さいたま市中央区新都心１―１さいたま新都心合同庁舎１号館　関東信越国税局総務部会計課経費第一係　杉本日楠子　電話048―600―3111　内線2146
3　仕様書案の交付
⑴　交付期間　令和８年６月11日から令和８年７月２日まで。
⑵　交付場所　上記２⑵に同じ。
4　Summary
⑴　Classification of the services to be procured : 133
⑵　Nature and quantity of the services to be required : Operation of call center for tax return guidance.
⑶　Time-limit for the submission of comments : 17 : 00　2 July 2026.
⑷　Contact point for the notice : SUGIMOTO Hinako Keihi Daiichi Section, Accounting Division, Kanto-Shinetsu Regional Taxation Bureau, 1―1 Shintoshin Chuo-ku Saitama city, Saitama 330―9719 Japan. TEL O48―600―3111 ex. 2146