⑸　調達に必要とされる基本的な要求要件

Ａ　生理機能検査システムは、生理検査データの収集・管理・参照が可能なシステムであり、院内ネットワークを介して電子カルテ端末から参照が可能な機能を有すること。

Ｂ　生理機能検査システムで管理される検査データは、単一のデータベースで一元管理する機能を有すること。

Ｃ　生理機能検査システムは、ユーザーＩＤごとにアクセス権限(参照・編集・管理等)を設定でき、個人単位での権限制御が可能な機能を有すること。

Ｄ　生理機能検査システムは、生理機能検査システムと接続された既存の日本光電工業株式会社製の心電計、運動負荷試験装置、長時間心電図解析装置、脳波計および筋電図・誘発電位検査装置と検査オーダ情報の連携が可能な機能を有すること。

Ｅ　生理機能検査システムは、検査結果をレポート形式で作成でき、ＰＤＦ等の汎用形式で出力可能な機能を有すること。

Ｆ　生理機能検査システムは、過去の検査データとの比較表示が可能であり、同一患者の時系列評価を支援する機能を有すること。

Ｇ　脳波計は、生理機能検査システムと院内ネットワークにて接続が可能であり、検査オーダ情報の連携や検査結果の送信が可能な機能を有すること。

Ｈ　脳波計は、多チャネル脳波計測が可能であり、臨床検査に十分な分解能およびサンプリング性能を有すること。

Ｉ　脳波計は、脳波の周波数を解析する機能を有すること。

Ｊ　脳波計は、ビデオカメラの接続が可能であり、脳波検査中の映像を表示・保存する機能を有すること。

Ｋ　筋電図・誘発電位検査装置は、生理機能検査システムと院内ネットワークにて接続が可能であり、検査オーダ情報の連携や検査結果の送信が可能な機能を有すること。

Ｌ　筋電図・誘発電位検査装置は、弁別比や雑音に強いアンプを搭載しており、各種神経伝導検査や聴覚誘発反応検査、体性感覚誘発電位検査などの誘発電位検査に対応した検査装置であること。

Ｍ　筋電図・誘発電位検査装置は、電気刺激、音刺激、視覚刺激が可能な機能を有すること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　令和８年８月７日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒606―8507 京都市左京区聖護院川原町54　京都大学医学部附属病院経理・調達課　契約掛長　植田　忠紘　電話075―751―3025

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　令和８年６月22日から令和８年８月７日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　令和８年７月６日10時00分

⑵　開催場所　京都大学医学部附属病院外来棟５階　会議室Ｂ

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : The physiological function test system 1 set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The physiological function test system is capable of collecting, managing, and viewing physiological test data, and must have the ability to view it from electronic medical record terminals via the hospital network.

B　Test data managed by the physiological function test system must be centrally managed in a single database.

C　The physiological function test system must have the ability to set access permissions (view, edit, manage, etc.) for each user ID, allowing for individual permission control.

D　The physiological function test system must have the ability to link test order information with existing Nihon Kohden electrocardiographs, exercise stress testing devices, long-term electrocardiogram analyzers, electroencephalographs, and electromyography／evoked potential testing devices that are connected to the physiological function test system.

E　The physiological function test system must have the ability to create test results in report format and output them in a standard format such as PDF.

F　The physiological function test system must have the ability to display and compare test data with past test data, supporting time-series evaluation of the same patient.

G　The EEG device must be capable of connecting to the physiological function test system via the hospital network and have the functionality to share test order information and transmit test results.

H　The EEG device must be capable of multi-channel EEG measurement and have sufficient resolution and sampling performance for clinical testing.

I　The EEG device must have the functionality to analyze EEG frequencies.

J　The EEG device must be capable of connecting to a video camera and have the functionality to display and save images during EEG testing.

K　The EMG／evoked potential testing device must be capable of connecting to the physiological function test system via the hospital network and have the functionality to share test order information and transmit test results.

L　The EMG／evoked potential testing device must be equipped with an amplifier that is resistant to discrimination ratios and noise, and be compatible with various evoked potential tests such as nerve conduction tests, auditory evoked response tests, and somatosensory evoked potential tests.

M　The EMG／evoked potential testing device must have the functionality to deliver electrical, acoustic, and visual stimulation.

⑸　Time limit for the submission of the requested material : 17 : 00　7 August, 2026

⑹　Contact point for the notice : Tadahiro Ueda, Contract Section, Accounting and Procurement Division, Kyoto University Hospital, 54 Kawahara-cho Shogoin Sakyo-ku Kyoto-shi 606―8507 Japan, TEL 075―751―3025

意見招請に関する公示

次のとおり調達役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

令和８年６月22日

総務省統計局長　永島　勝利

◎調達機関番号　012　◎所在地番号　13

１　調達内容

⑴　品目分類番号　72

⑵　購入等件名及び数量　2027年経済構造実態調査及び2027年事業所・企業照会の実施等業務の請負　一式

２　意見の提出方法

⑴　意見の提出期限　令和８年７月13日16時00分迄（郵送の場合は必着のこと。）

⑵　提出先　〒162―8668 東京都新宿区若松町19―１総務省第二庁舎　総務省統計局総務課支出負担行為係　電話03―5273―1122
E-mail　g-futankoui@soumu.go.jp

３　仕様書案の交付等

⑴　交付期間　公示日から意見の提出期限まで

⑵　交付場所　政府電子調達（ＧＥＰＳ）システムへアクセスし入手すること。
https://www.p-portal.go.jp/pps-web-biz/

４　仕様書案の説明会

説明会に参加しようとする者は、開催日時の前日までに２⑵へ説明会参加の旨をメールで登録すること。

⑴　開催日時　令和８年６月30日14時00分

⑵　開催場所　総務省第二庁舎１階105号室