5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Magnetic Resonance Imaging System 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
1）The system should have main magnetic field of 3.0 Tesla.
2）The system should be convenient and safe for patients, and be easy to operate.
3）Floor plan should have high space utility, and should be easy for patients and operators to access the system.
4）The system should have high capability in morphologic and functional data acquisition in central nervous system, cardiovascular system, and other organs.
5）The system should have high capability to process, and analyze the data.
6）The system should have high flexibility for upgrading.
7）The system should have communication with other computer and network systems.
8）The system and users should be supported by appropriate and prompt maintenance services.
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M August 31, 2009
⑹　Contact point for the notice : Tamio Hirohata, Director, Accounting Division, National Cardiovasucular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu 565―8565 Japan. TEL06―6833―5012 ext. 2213

○第10号③（№10③）

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
フルデジタル血管撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成21年度第４・四半期
⑸　調達に必要とされる基本的な要求要件
1）頭部領域、胸腹部領域、四肢末梢血管領域の検査および治療に幅広く適応できること。
2）バイプレーンデジタル透視、バイプレーンデジタル撮影が可能であること。
3）透視画像および撮影画像が優れていること。
4）血管回転撮影および３次元表示が可能で、十分なデータ保管機能を有すること。
5）軟部組織断面を表示できるＣＴ機能を備え、解析が可能であること。
6）ＤＩＣＯＭによるデジタル画像観察・保管が可能であること。
7）撮影画像のオンラインでの観察・解析が可能なこと。
8）当センターのネットワークに接続し、患者基本情報の受信、画像データの送受信ができること。
9）患者、術者の被曝軽減対策が施されていること。
10）操作性、患者や術者の快適性と安全性および空間効率に配慮した設計であること。
11）その維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法
上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成21年８月31日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台５－７－１　国立循環器病センター会計課長　廣畑　民夫　電話06―6833―5012　内線2213
3　説明書の交付
本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成21年７月15日から平成21年８月31日まで。
⑵　交付場所　上記２⑵に同じ。
4　その他
この導入計画の詳細は導入説明書による。
なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Full Digital Angiographic System 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
1）The angiographic system should be applied for catheter examinations and intervensional procedures for cerebrovascular disease, peripheral arterial disease, and other vascular disease.
2）The biplane-angiographic system should show us high performance for digital fluoroscopy and angiography.
3）The biplane-angiographic system should enable us to obtain high quality fluoroscopic and angiographic images.
4）The angiographic system should enable us to perform rotational angiography procedure in combination with 3D reconstruction image. In addition, the system should also have much memory for data storage.
5）The angiographic system should show us cross-sectional, soft tissue CT image, and also enables us the analysis of the acquired data.
6）The angiographic system should have ability for viewing, recording DICOM volume data.
7）The angiographic machine should provide us viewing and analysis of 3D volume on intra-network system in National Cardiovascular Center.
8）After the access to the intra-network in National Cardiovascular Center, the angiographic system should have a capability to accept the patient information as well as post-processing data.
9）Using the angiographic system, radiation exposure of the patient, doctors, and co-medical stuffs during the procedure should be reduced.
10）The angiographic system should provide us safety, comfortableness during the interventional procedure. In addition, the system should be suitable for the indicated angiographic room in terms of size and layout.
11）The angiographic system and users should be supported by appropriate and prompt maintenance services.
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M August 31, 2009
⑹　Contact point for the notice : Tamio Hirohata, Director, Accounting Division, National Cardiovasucular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu 565―8565 Japan. TEL 06―6833―5012 ext. 2213

○第10号④（№10④）

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
サイクロトロン・ＰＥＴ薬剤合成装置　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成21年度第４・四半期
⑸　調達に必要とされる基本的な要求要件
1）循環器疾患の検査に広く適用できる高性能サイクロトロンおよび合成装置であること。
2）Ｃ―11、Ｎ―13、Ｏ―15およびＦ―18の安定供給が可能なシステムであること。
3）循環器疾患の診療と併せて、臨床研究も同時に行えるシステムであること。
4）Ｏ―15ガス、Ｆ―18ＦＤＧ等の診療用標識薬剤合成装置を備えていること。
5）Ｃ―11、Ｎ―13、Ｏ―15、Ｆ―18等の研究用標識薬剤合成装置を備えていること。
6）信頼性が高く、迅速かつ安全な照射および合成の実施が可能であること。
7）優れた操作性と簡便性を有する装置であること。
8）ＰＥＴ薬剤の品質管理および放射線防護に必要な周辺機器を備えていること。
9）装置および機器の運営管理に有用な周辺機器を備えていること。
10）装置の維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法
上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成21年８月31日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台５－７－１　国立循環器病センター会計課長　廣畑　民夫　電話06―6833―5012　内線2213