J　The instrument should be equipped with a field emission type electron source for Auger Electron Spectroscopy.

K　The instrument should be equipped with a heat and cool accessory, to enable samples to be either heated or cooled while in the XPS analysis position in the sample analysis chamber.

L　The instrument should be equipped with a heat and cool accessory, to enable samples to be either heated or cooled while in the sample entry chamber.

M　The instrument should be equipped with a gas reaction cell accessory in the sample entry chamber.

N　An accessory to enable a sample, which is prepared in a chamber isolated from atmosphere such as a glove box, to be loaded into the instrument without exposing to the atmosphere should be provided.

O　Sample exchange while acquiring data should be possible.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 27, August, 2009

⑹　Contact point for the notice : Haga Masafumi, Procurement Section, Finance Department, Tohoku University, 2―1―1 Katahira Aoba-ku Sendai-shi 980―8577 Japan, TEL 022―217―4869

## 資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成21年7月27日

国立大学法人富山大学事務局長　鹿野　芳郎

◎調達機関番号　415　◎所在地番号　16

**○第１号**

1　調達内容

⑴　品目分類番号　24

⑵　導入計画物品及び数量
ヘリウム液化システム　一式

⑶　調達方法　購入等

⑷　導入予定時期
平成21年度３月以降

⑸　調達に必要とされる基本的な要求要件

A　ヘリウム液化システムとして既存の設備を活用し、能力を上回るものであること。（液体窒素予冷時液化能力、40Ｌ／hr以上の能力）

B　供給者は高圧ガス保安法を遵守したシステムの設計に責任をもつこと。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成21年８月28日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒930―8555 富山市五福3190　国立大学法人富山大学財務部経理グループ契約チーム　藤井　栄吉　電話076―445―6055

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成21年７月27日から平成21年８月28日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成21年８月５日14時00分

⑵　開催場所　国立大学法人富山大学事務局入札室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 24

⑵　Nature and quantity of the products to be purchased : Helium liquefaction system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　In the new system the performance higher than that of the present system is required. (more than 40 L／h with liq. N2 precooling.) The system must be suited to the present system.

B　Supplier is responsible for designing the system that obeys the High Pressure Gas Safety Law.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 28, August, 2009

⑹　Contact point for the notice : Eikichi Fujii, Accounting Group, Financial Affairs Division, University of Toyama, 3190 Gofuku Toyama-shi 930―8555 Japan, TEL 076―445―6055

## 資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成21年７月27日

国立大学法人　山梨大学長　前田秀一郎

◎調達機関番号　415　◎所在地番号　19

**○第２号**

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量
迅速検査・治療システム　一式

⑶　調達方法　購入等

⑷　導入予定時期
平成21年度２月以降

⑸　調達に必要とされる基本的な要求要件
⒜最先端マルチスライスＣＴ撮影装置を有すること　⒝新生児から成人までに対応できる従量・従圧式人工呼吸器装置と非侵襲陽圧人工呼吸器装置を有すること　⒞ＡＥＤ機能を備える除細動器装置を有すること　⒟温度制御方法が選択できる閉鎖式保育器装置を有すること　⒠緊急時でも迅速に使用できる人工心肺装置・補助循環装置を有すること　⒡ＣＨＤＦ、ＣＨＤ、ＣＨＦを施行できる血液浄化用装置を有すること　⒢ＶＣＶ、ＰＣＶ、ＶＳＶ、ＰＳＶ、ＳＩＭＶモードに対応が可能な人工呼吸器装置を有すること　⒣結果判定が迅速で精度が高い血液ガス分析装置を有すること　⒤急変にも迅速・簡潔に対応できる据置型ベッドサイドモニタ装置と移動用ベットサイドモニタ装置を有すること　⒥Ｘ線Ｃアーム装置での撮影が可能な高機能ＩＣＵベッドを有すること　⒦低体温療法での使用ができる体温管理システムを有すること　⒧速やかな診断と最適な患者管理が可能な連続心拍出量計を有すること　⒨活性化凝固時間を測定できる血液凝固計を有すること

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成21年９月11日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒409―3898 山梨県中央市下河東1110番地　山梨大学医学部管理課病院契約グループ　大澤　正弘　電話055―273―9224

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成21年７月27日から平成21年９月11日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成21年８月10日15時00分

⑵　開催場所　山梨大学医学部キャンパス内管理棟２階小会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Fast testing and emergency treatment system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

⒜　The system contains state-of-the-art multislice CT. ⒝　The system contains artificial respirator device. ⒞　The system contains defibrillator device that has AED function. ⒟　The system contains closed system incubator device that can select method of controlling temperature. ⒠　The system contains artificial heart and lung machine that can be promptly used even in emergency. ⒡　The system contains hemocatharsis device that can enforce CHDF, CHD, and CHF. ⒢　The system contains artificial respirator device to be enforceable of VCV, PCV, VSV, PSV, and the SIMV mode is contained. ⒣　The system contains blood gas apparatus to be able to do prompt analysis. ⒤　The system contains fixed mount type bed side monitor de-