◎調達機関番号　017　◎所在地番号　13

○第 26 号（No.26）

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量

入院患者監視システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成21年度第四・四半期

⑸　調達に必要とされる基本的な要求要件

①　ナースコール親機（以下「親機」という）はデスクトップ型とし、タッチパネル操作機能を有すること。

②　ナースコール呼出時は、親機画面に患者情報をポップアップし、患者氏名、ベッド番号、呼出種別（ナースコール、スタッフコール、トイレコール等）、診療科名、担当医師名、担当看護師名等の情報を表示する機能を有すること。

③　ナースコール呼出時は、ハンディナース（ＰＨＳ）画面に、患者氏名、ベッド番号、呼出種別等の情報を表示する機能を有すること。

④　病院情報システムの電子カルテシステムとの連動により、患者情報をナースコールシステムに取り込む機能を有すること。

⑤　生体情報監視モニターとの連動により、患者容態の急変をナースコールシステムに自動で知らせる機能を有すること。

⑥　電話交換システムとの連動により、ハンディナース（ＰＨＳ）で内線通話をする機能を有すること。

2　資料及びコメントの提供方法

上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成21年９月14日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒104―0045 東京都中央区築地５－１－１　国立がんセンター中央病院会計第一課長　竹内　茂男　電話03―3542―2511　内線2150

3　説明書の交付

本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成21年７月31日から平成21年９月14日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催

本公表に基づく導入説明会を下記のとおり開催する。

⑴　開催日時　平成21年８月11日15時00分

⑵　開催場所　国立がんセンター管理棟会議室

5　その他

この導入計画の詳細は導入説明書による。

なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : In-patient monitoring system 1 set

⑶　Type of the procurement : purchase

⑷　Basic requirements of the procurement :

①　The nurse call main phone (Hereafter, it is said, “Main phone”) must assume the desktop system, and have the touchpanel operation function.

②　Possess the function to pop up the patient information on the main phone screen, and to display the patient name, the bed number, the call type (nurse call, staff call, and rest room call, etc.), the diagnosis and treatment department names, the doctor in charge names, and information by charge nurse at the nurse call.

③　Have the function to display information on the patient name, the bed number, and the call type, etc. on handynurse (PHS) screen at the nurse call.

④　Possess the function to take patient information into the nurse call system by synchronization with the electronic clinical record system of the hospital synthesis information system.

⑤　Possess the function to inform the nurse call system of the sudden change of the patient condition by the automatic operation by synchronization with the living body information watch monitor.

⑥　Possess the function to talk over the telephone extension by handynurse (PHS) by synchronization with the telephone switching system.

⑸　Time-limit for the submission of the requestrd material : 5 : 00 P.M. September 14th, 2009

⑹　Contact point for the notice : Takeuti Shigeo, Director, Accounting Section 1, National Cancer Center Chuo Hospital, 5―1―1, Tsukiji, Chuo-ku, Tokyo, 104―0045, Japan. TEL 03―3542―2511 ext. 2150

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 21 年７月 31 日

支出負担行為担当官

国立がんセンター中央病院運営部次長

平田　強

◎調達機関番号　017　◎所在地番号　13

○第 27 号（No.27）

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量

全身ＭＲＩ装置　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成21年度第四・四半期

⑸　調達に必要とされる基本的な要求要件

①　磁場強度が3.0Ｔで、広い撮像視野で磁場均一性が良いマグネットシステムであること。

②　誘電効果による感度ムラと熱吸収比を低減させる技術が搭載されていること。

③　複数のｂ値収集による拡散強調画像の撮像が可能であること。

④　ＲＦコイルは検査効率の良いシステム構成であること。

⑤　病院情報システムとの間で画像（ＤＩＣＯＭ3.0）、患者検査情報及び実施等の入出力ができること。

2　資料及びコメントの提供方法

上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成21年９月14日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒104―0045 東京都中央区築地５－１－１　国立がんセンター中央病院会計第一課長　竹内　茂男　電話03―3542―2511　内線2150

3　説明書の交付

本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成21年７月31日から平成21年９月14日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催

本公表に基づく導入説明会を下記のとおり開催する。

⑴　開催日時　平成21年８月11日15時30分

⑵　開催場所　国立がんセンター管理棟会議室

5　その他

この導入計画の詳細は導入説明書による。

なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Magnetic resonance tomogram system 1 set

⑶　Type of the procurement : purchase

⑷　Basic requirements of the procurement :

①　The magnet with magnetic field 3.0T should be a uniform magnetic field for imaging of wide field of view.

②　The system should include technology to reduce local specific absorption rate (SAR) and overcome image degradation by the dielectric effect.

③　Able to image acquisition of diffusion-weighted imaging with more than three kinds of b values within a sequence.