⑤　独立行政法人理化学研究所　神戸研究所
〒650―0047 兵庫県神戸市中央区港島南町２－２－３
・研究推進部経理課
電話078―306―3015　Fax078―306―3122
⑥　独立行政法人理化学研究所　仙台支所
〒980―0845 宮城県仙台市青葉区荒巻字青葉519―1399
・仙台研究推進室
電話022―228―2111　Fax022―228―2122
⑦　独立行政法人理化学研究所　名古屋支所
〒463―0003 愛知県名古屋市守山区大字下志段味字穴ヶ洞2271―130
・名古屋研究推進室
電話052―736―5850　Fax052―736―5854

招　請

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成21年12月10日

国立国際医療センター戸山病院運営局次長　福田　隆

◎調達機関番号　017　◎所在地番号　13

○第34号（No.34）

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量

①　高精度ビーム強度変調放射線治療システム　一式

②　ＰＥＴサイクロトロン＆ＲＩ合成システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成22年度第３・四半期

⑸　調達に必要とされる基本的な要求要件

①－１　多岐にわたる強度変調照射機能を有していること。

①－２　ＭＶＣＴ等による画像誘導放射線治療（ＩＧＲＴ）が可能であること。

①－３　治療計画、ＱＡ（検証）、最新のＣＴ装置による画像誘導放射線治療、照射のプロセスを統合したシステムであること。

①－４　マルチコリメーターを利用したヘリカル照射が可能であること。

②－１　サイクロトロンは陽電子（ポジトロン）放出核種を製造するために陰荷電粒子を用いる負イオン型のサイクロトロンであり２ポート同時照射ができること。

②－２　標識化合物自動合成装置は、Ｆ－18標識ＦＤＧの合成、Ｃ－11標識ヨウ化メチル・メチオニン酢酸・コリン、Ｆ－18標識ＦＬＴ等の多目的な合成ができること。合成装置は取扱が容易で、かつ高収率・短時間に安定して合成ができ、低ランニングコストであること。

②－３　合成装置はホットセルに格納され、作業者の被爆を防護できること。

②－４　標識化合物の純度等の安全性を確認できること。

②－５　必要とする放射能測定器及び関連する品質管理機器や検査装置などの周辺機器を備えること。

2　資料及びコメントの提供方法

上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成22年１月29日17時15分（郵送の場合は必着のこと。）

⑵　提出先　〒162―8655 東京都新宿区戸山１－21－１　国立国際医療センター戸山病院会計第一課長　佐藤　星斗　電話03―3202―7181　内線2050

3　説明書の交付

本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成21年12月18日から平成22年１月29日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を下記のとおり開催する。

⑴　開催日時

①　平成21年12月18日14時00分

②　平成21年12月18日14時30分

⑵　開催場所　院内会議室

5　その他

この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased :

①　Highly Accurate Intensity Modulated Radiothrapy use 1 set.

②　PET cyclotron & RI synthesis system use 1 set.

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

①－１　This system shall have various function such as IMRT.

①－２　This system shall allow IGRT function such as MVCT.

①－３　All processes of Treatment planing, Quality Assurance, CT Guided Randiation Therapy and Treatment Delivery by Newly CT Equipment shall be possible on this system.

①－４　This system shall allow Helical pattern delivery, under guidance by Multi-Leaf Collimator.

②－１　Cyclotron system is negative ion type with dual particle to produce positoron emitting nuclides, and has function of dual port simultaneously.

②－２　The synthesizers shall be for synthesis of F―18 labeled FDG and for synthesis of C―11 labeled compound such as methyliodide, methionine, acetate, choline and F―18 labeled compound FLT etc. The synthesizers must be for easy operation, high yield, short synthesis time and low running cost.

②－３　The synthesizers must be installed in hot cells to avoid the radiation exposure for the operators.

②－４　Safety of radio-pharmaceuticals, such as purity, etc. must be confirmed.

②－５　Necessary devices for measuring radioactivity, quality check and inspection equipment and peripheral equipment shall be insalled.

⑸　Time-limit for the submission of the requested material : 5 : 15 P.M January 29, 2010

⑹　Contact point for the notice : Seito Sato, Director, Accounting Division 1, International Medical Center of Japan Toyama Hospital, 1―21―1 Toyama Shinjuku-ku, Tokyo 162―8655 Japan. TEL 03―3202―7181 ext. 2050

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

平成21年12月10日

総務省大臣官房会計課企画官　福田　進吉

◎調達機関番号　012　◎所在地番号　13

1　調達内容

⑴　品目分類番号　71、27

⑵　購入等特定役務及び数量　IPv6対応に向けたテストベッドによる実証実験に係る請負　一式

2　意見の提出方法

⑴　意見の提出期限　平成22年１月４日午後５時（郵送の場合は必着のこと。）

⑵　提出先　〒100―8926 東京都千代田区霞が関二丁目１番２号　総務省大臣官房会計課契約第二係　担当　森　　康二　電話03―5253―5132

3　仕様書案の交付

⑴　交付期間　平成21年12月10日から平成22年１月４日まで

⑵　交付場所　上記２⑵に同じ

4　Summary

⑴　Classification of the services to be procured : 71, 27

⑵　Nature and quantity of the services to be required : Experiment in testbed for adapting to IP version 6　1 set

⑶　Time limit for the submission of comments : 5 : 00 PM 4 January 2010

⑷　Contact point for the notice : Koji Mori, Accounts Division, Minister's Secretariat, Ministry of Internal Affairs and Communications, 1―2 Kasumigaseki 2-chome, Chiyoda-ku Tokyo 100―8926 Japan. TEL 03―5253―5132