② C-arm should equip either one or two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary-direction.
③ The flat-panel detector should be especially designed to fulfill the interventional angiography.
④ The digital imaging processing system should have supporting and analysis function for intravascular treatment.
⑤ The system should equip the radiation exposure reduction unit for patient, operator, and staff.

(4) Gamma Camera System
① The system should be multiple digital gamma camera.
② The system should have function of SPECT and Whole body acquisition.
③ The system should equip the fusion software and quantitative application-replete.

(5) X-ray fluoroscope Radiography System
① The system is available in fluoroscopic radiography.
② The system should capable of recording dynamic images.
③ The system should be capable of positioning the bed position to horizontal.
④ The system should equip the flat-panel detector.

(6) General Radiographic System
① The system should equip x-ray tube.
② The system should equip a table for-standing and lying position, in addition, radiography of whole body couldable to perform speedily and safely.
③ The system should equip the flat-panel detector.

(7) High Radiationtherapy System
① The system should be capable to multipurpose treatment.
② The system should equip multi-leaf collimator with high shielding performance.
③ The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiationtherapy.

5 Time-limit for the submission of the requested material : 5 : 00 P.M. March 18, 2010

6 Contact point for the notice : Katsumi Miyashita, Manager, Facilities Planning Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2―5―21 Higashigaoka Meguro-ku Tokyo-to 152―8621, Japan. TEL 03―5712―5071

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成22年2月1日

国立大学法人三重大学長　内田　淳正

◎調達機関番号　415　◎所在地番号　24

○第３号

1　調達内容
(1)　品目分類番号　15、28
(2)　導入計画物品及び数量
病院内情報ネットワークシステム　一式
(3)　調達方法　購入等
(4)　導入予定時期
平成22年度11月以降
(5)　調達に必要とされる基本的な要求要件
A　本学医学部附属病院内において平成23年度に稼動予定の新総合医療情報システムを十分機能させる事が可能な高速で信頼性のある情報ネットワークシステムであること。
B　基幹ネットワークは10ギガビット以上、支線ネットワークは１ギガビット以上の通信速度を有する高速大容量ネットワークシステムであること。
C　本学医学部附属病院内のファイアーウォールを介して、本学のキャンパスＬＡＮにセキュリティを確保して接続できること。
D　病院内に不正接続防止の認証機能を有するＩＥＥＥ802.11ｎ規格対応無線ＡＰを設置し、ユビキタス環境で情報ネットワークを利用可能にすること。また、人・機器の位置情報を確認できる位置検出管理機能を有すること。
E　ネットワーク監視機能を有し、ネットワーク機器および通信を常時監視し、機器の異常や通信障害を自動的に通知できること。
F　24時間365日の一次対応を含めた保守作業を行うことが可能な保守体制を有すること。

2　資料及びコメントの提供方法　上記１(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1)　資料等の提供期限　平成22年３月４日17時00分（郵送の場合は必着のこと。）
(2)　提供先　〒514―8507 津市栗真町屋町1577
三重大学財務部病院再開発調達プロジェクトチーム　丹羽　　章　電話059―231―5716

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1)　交付期間　平成22年２月１日から平成22年３月４日まで。
(2)　交付場所　上記２(2)に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。
(1)　開催日時　平成22年２月８日15時00分
(2)　開催場所　三重大学事務局会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
(1)　Classification of the products to be procured : 15, 28
(2)　Nature and quantity of the products to be purchased : The Integrated High Speed Hospital Network System 1 Set
(3)　Type of the procurement : Purchase
(4)　Basic requirements of the procurement :
A　The system needs to be able to make to operate sufficiently the new integrated medical information system of Mie University Hospital expected to start from 2011.
B　The system needs to be the large-capacity and the high-speed Network system, which has a 10Gbps backbone network, and 1Gbps connection to terminals, in order to deliver images for medical diagnosis and video of medical care.
C　The network can be connected with the campus LAN through the existing fire wall system.
D　The system needs to have wireless access points at various places in the hospital, and be able to operate the new integrated medical information systems. The Wireless LAN needs to be available for the IEEE802.11n standard, and have an authentication function for preventing unauthorized connections. In addition, the system needs to be the sensor network that is able to collect the location information of persons and goods.
E　The system must have a monitoring function which maintains constant surveillance of the network and notifies the administrator of failure of the equipment and of unusual data traffic on the network.
F　The system maintenance including the first correspondence should be possible during 24 hours throughout the year.
(5)　Time limit for the submission of the requested material : 5 : 00 PM 4, March, 2010
(6)　Contact point for the notice : Akira Niwa, University Hospital Redevelopment Project Team, Finance Department, Mie University, 1577 Kurimamatiya-cho Tsu-shi 514―8507 Japan, TEL 059―231―5716

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成22年2月1日

国立大学法人長崎大学長　片峰　　茂

◎調達機関番号　415　◎所在地番号　42

○第８号

1　調達内容
(1)　品目分類番号　14
(2)　導入計画物品及び数量
原爆被爆者データベース用電子計算機システム　一式