3）Pre-analytical device shall be equipped de-stopper function, aliquot function, labeling function, and sorting function.
4）The system shall be an integrated system for sample and information handling that is capable for various laboratory processes such as preparation for clinical tests, analyzing samples and making reports. (The system shall be ready for present analyzing and hospital information system.)
5）It shall be equipped ultra pure water device for operation.
6）It shall be operated on 24 hours a day, every day.

② Washer／Disinfection and Sterilization System 1 Set
1）To provide high-pressure steam sterilization at an optional temperature of 115～135℃, by using the existing steam facilities.
2）To be able to sterilize with ethylene oxide gas from 20% mixed gas cylinder.
3）To make exhausted gas from ethylene oxide gas sterilizer harmless.
4）The all processes of sterilization must be controlled by the computer.
5）The sterilizer must be swing-double-door type.
6）To have full automatic washing, disinfecting and drying functions with 1 chamber.
7）To provide ultrasonic washing function.
8）To be able to disinfect with hot water.
9）To have automatic rack transferring system in and between washing, disinfecting and drying processes.
10）To provide hydrogen peroxide plasma sterilization.
11）The all processes of hydrogen peroxide plasma sterilization must be controlled by the computer.
12）To be able to disinfect corrugated tubes and masks, etc with hot water.
13）To provide with the air gun and the water gun in two-tank-type sink.
14）To be able to sterilize with the 100% ethylene oxide gas cartridge.
15）To make exhausted gas from ethylene oxide gas sterilizer harmless.
16）To provide washing and dehydrating function that have various programs and be able to desinfect with hot water.
17）The system must have efficient transferring and storage functions of sterilized objects.

⑸ Time-limit for the submission of the requested material : 5 : 00 PM 26, April, 2010
⑹ Contact point for the notice : Hiroshi Ijiri, Procurement Section, Accounting Division, Yokohama Rosai Hospital, 3211 Kozukue-cho Kohoku-ku Yokohama-shi Kanagawa-ken 222―0036 Japan, TEL 045―474―8111

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 22 年 3 月 11 日
国立大学法人弘前大学
医学部附属病院長　花田　勝美
◎調達機関番号　415　◎所在地番号　02

○第７号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
救急患者診断用マルチスライス型ＣＴ撮影装置　一式
⑶　調達方法　購入等
⑷　導入予定時期
平成22年度２月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　空間分解能に優れた全身のＸ線ＣＴ画像が得られること。
Ｂ　濃度分解能に優れた全身のＸ線ＣＴ画像が得られること。
Ｃ　時間分解能に優れた全身のＸ線ＣＴ画像が得られること。
Ｄ　全身の等方性のボリュームデータを撮影可能であること。
Ｅ　心電図同期撮影が行えること。
Ｆ　画像情報を他のワークステーションに転送可能であること。
Ｇ　ディテクターは64列以上であること。
Ｈ　病院情報システムおよび放射線情報システムと連動すること。
Ｉ　画像情報のフィルムレス化に対応できること。
Ｊ　ＤＩＣＯＭ対応で十分な容量のデータサーバを有すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成22年４月26日15時00分（郵送の場合は必着のこと。）
⑵　提供先　〒036―8563 弘前市本町53　弘前大学医学部附属病院経理調達課　経理調達グループ　調達担当　山口　富雄　電話0172―39―5197
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成22年３月11日から平成22年４月26日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成22年３月26日14時30分
⑵　開催場所　弘前大学医学部附属病院小会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Multi-detector diagnostic CT system for emergency patients 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system should provide cross-sectional images of the whole body with excellent spatial resolution.
B　The system should provide cross-sectional images of the whole body with excellent contrast resolution.
C　The system should provide cross-sectional images of the whole body with excellent temporal resolution.
D　The system should be capable of isotropic volume CT scan of the whole body.
E　The system should be capable of an ECG-gated CT scan.
F　The system should be capable of transmission of scanned images for any workstations.
G　The system should have equal to more than 64-row detectors.
H　The system should work in cooperation with hospital information system (HIS) and radiology information system (RIS).
I　The system should work in a filmless environment including picture archiving and mmunication system (PACS).
J　The system should have data server for DICOM with enough volume.
⑸　Time limit for the submission of the requested material : 3 : 00 PM 26, April, 2010
⑹　Contact point for the notice : Tomio Yamaguchi, Accounting Procurement Group, Accounting Procurement Division,Hirosaki University Hospital, 53 Hon-cho Hirosaki-shi 036―8563 Japan, TEL 0172―39―5197

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 22 年 3 月 11 日
国立大学法人東京大学総長　濱田　純一
◎調達機関番号　415　◎所在地番号　13

○第８号

1　調達内容
⑴　品目分類番号　14
⑵　導入計画物品及び数量
データ統合・情報融合コアシステムの拡張システム　一式
⑶　調達方法　購入等
⑷　導入予定時期
平成22年度２月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　データ公開用サーバ、データ統合・解析サーバ拡張システム、大容量磁気ディスクアレイ装置、大容量磁気テープライブラリ装置、管理ソフトウェア、インタフェースアダプタ、その他関連機器から構成されるものとする。
Ｂ　大量の地球環境関連データを保存・解析・公開するのに十分な演算性能・記憶容量・Ｉ／Ｏ性能を有すること。
Ｃ　データ公開用サーバは現有のディスクアレイ装置を接続し、利用可能であること。
Ｄ　大容量磁気ディスクアレイ装置は省エネルギー運転を実現する機能を有すること。
Ｅ　運用管理が容易で、セキュリティ管理等を十分考慮したものであり、高い耐故障性、信頼性、可用性を有すること。
Ｆ　システムの障害を未然に防止し、また障害発生時には即時に対応できる体制を有すること。