6　Summary
⑴　Classification of the products to be procured : 26
⑵　Nature and quantity of the products to be purchased : Fabrication and evaluation system for organic thin-film solar cells 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Thin-film fabrication and photocurrent evaluation system
①　The vacuum evaporation system must have at least one chamber for organic film and one chamber for metal film, respectively.
②　The vacuum evaporation system and the glove box must be connected so that the substrate can be transferred between them without exposure to air.
③　The action spectrum for monochromatic light and the photovoltaic performance under AM1.5 solar simulator can be measured in the glove box.
B　Measurement system for excited states
①　The system must equip a pump-probe system to measure the transient photo-absorption spectrum.
②　The spectrum must be measured in the wavelength region from 400nm to 1400nm.
③　The system must include a femtosecond laser pulse system as a light source.
C　Evaluation system for fine structure of thin films
①　The system must include a scanning probe microscope system in order to observe surface profile of the organic thin films.
②　The system must equip an optical microscope whose magnification must be over 1000.
③　The sample of 30㎜ square can be set up on the system.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 17, May, 2010
⑹　Contact point for the notice : Hidekatsu Ono, Procurement Section, Accounting Division, Yamagata University Faculty of Engineering, 4―3―16 Jonan Yonezawa-shi 992―8510 Japan, TEL 0238―26―3009

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成22年4月16日
国立大学法人金沢大学長　中村　信一
◎調達機関番号　415　◎所在地番号　17
○第１号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
ＦＰＤ型Ｘ線撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期
平成22年度３月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　ＦＰＤ型Ｘ線テレビ装置は、高精細フラットパネルディテクタを搭載し、Ｘ線透視と撮影において、様々な画像取得機能、画像処理機能及び画像転送機能を有すること。
Ｂ　ＦＰＤ型Ｘ線撮影システムは、フラットパネルディテクタを搭載し、天井走行式Ｘ線管を有し、様々な画像取得機能、画像処理機能及び画像転送機能を有すること。
Ｃ　ＣＲ型Ｘ線撮影システムは、立位タイプ撮影台（ＩＰカセットを保持）を有し、Ｂシステムの Ｘ線管によりＸ線撮影ができる機構であり、各サイズのカセッテを６枚ずつ有すること。また２台のＩＰ処理機と１台の画像観察ワークステーションを有すること。
Ｄ　それぞれの装置において最新研究に対応できるローデータ転送機能を有すること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成22年６月１日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒920―8640 金沢市宝町13―１ 金沢大学宝町地区事務部会計課調達係長　西嶋　浩司　電話076―265―2803
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成22年４月16日から平成22年６月１日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成22年４月23日15時00分
⑵　開催場所　金沢大学医薬保健学域保健学類棟１号館会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : FPD type X-ray radiography system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　For the FPD type X-ray fluoroscopy system, the following parts and functions should be included.
a　high resolution flat-panel X-ray detector
b　flexible image acquisition modes
c　effective image processing and image transfer functions
B　For the FPD type X-ray radiography system, the following parts and functions should be included.
a　high resolution flat-panel X-ray detector
b　overhead X-ray tube crane system.
c　flexible image acquisition modes
d　effective image processing and image transfer functions
C　For the CR type X-ray radiography system, the following parts should be included.
a　upright stand for IP-cassette folding which is used for X-ray imaging by a combination with the X-ray tube mentioned in above B section.
b　six IP-cassettes for each size.
c　two IP-processing systems
d　one image viewing workstation
D　All systems should have a raw-data transfer function which contributes to advanced analysis.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 1, June, 2010
⑹　Contact point for the notice : Koji Nishijima, Procurement Section, Accounting Division, Takara-machi Area, Kanazawa University, 13―1 Takara-machi Kanazawa-shi 920―8640 Japan, TEL 076―265―2803

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成22年4月16日
国立大学法人神戸大学理事　竹園　充夫
◎調達機関番号　415　◎所在地番号　28
○第２号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
人工心肺システム　一式
⑶　調達方法　購入等
⑷　導入予定時期
平成22年度12月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　８台のローラーポンプかつ、１台の遠心ポンプが搭載できること
Ｂ　ポンプのオクルージョンはクリックオクルージョンであり、１クリックが0.015㎜の精度であること
Ｃ　全てのポンプがバッテリーにて駆動できること
Ｄ　非常用バッテリーの残量及び使用可能時間が表示されること
Ｅ　心筋保護装置機能を有すること
Ｆ　圧力モニターを５チャンネル以上有すること
Ｇ　ローラーポンプの種類が２種類以上あること
Ｈ　ディスプレイパネルの入替えが動作中にも可能で各パネルが共有できること
Ｉ　ポンプの増設が容易なこと
Ｊ　データ通信可能なこと
Ｋ　ＦＤＡの認可が取得済みであること