② 得られたＣＴ画像からＳＰＥＣＴ画像の吸収補正が行え、単独で診断可能なＣＴ画像とＳＰＥＣＴ画像の重ね合わせにより機能情報と形態情報が同時に把握できること。
③ 本体収集コンソールの他、核医学用解析コンソールを有し、全身のＳＰＥＣＴ検査が可能なこと。
④ 全身用コンピュータ断層撮影装置は16スライス以上のマルチスライスＣＴ装置であること。
⑤ 全身の検査に対し高速で撮影が可能なこと。
⑥ 被ばく低減に配慮されていること。
⑦ 本院の医療情報システムと接続すること。
⑧ 読影環境の整備の為、本院の画像保管装置の拡張を行うこと。

2 資料及びコメントの提供方法 上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに資料等の提供を招請します。
⑴ 資料等の提供期限 平成22年９月13日14時00分（現地時間） （郵送の場合は必着のこと）
⑵ 提出先
① 〒984―8560 宮城県仙台市若林区大和町２―29―１ 東日本電信電話株式会社 東北病院 企画担当 佐藤 学
電 話：（022）236―5711
ファックス：（022）238―7987
なおＮＴＴ東日本ホームページ http://www.ntt-east.co.jp/purchase/inquiry/index.html記載窓口においても問い合わせを行います。

3 Summary
⑴ Classification of the products to be procured : 22, 31
⑵ Name and quantity of the products to be purchased :
SPECT・CT System (1 set) and Multi-Slices CT System (1 set)
⑶ Type of the procurement : purchase
⑷ Introductory schedule time : March, 2011
⑸ Basic requirements of the procurement :
① Variable dual head SPECT・CT system with a diagnostic CT (2 slices or more).
② Fusion image of the independently obtained CT and SPECT images should allow simultaneous view of the functional information and the anatomical information as well as support attenuation correction.
③ Acquisition and processing console are configured separately and allow support for whole body SPECT studies.
④ Whole body CT system has to be the multi slice CT more than 16 slices.
⑤ CT system has to be able to acquire with high speed for whole body.
⑥ Low dose features must be supported.
⑦ The systems has to be connected to the medical information system now working at our hospital.
⑧ The upgrade of the image server has to be done at our hospital for the image viewing environment update.
⑹ Time-limit for the submission of the requested material : 2 : 00PM Sep 13, 2010 (Local Time)
⑺ Contact point for the notice :
a. TOHOKU MEDICAL CENTER NTT EC 2―29―1, Yamatomachi, Wakabayashi-ku, Sendai, Miyagi 984―8560 Japan
TEL : 81⑵2)236―5711
FAX : 81⑵2)238―7987
Attn. : Manabu Sato
b. Overseas contact points described on the following URL. http://www.ntt-east.co.jp/purchase/inquiry/index.html

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成22年７月28日
契約担当役
独立行政法人労働者健康福祉機構
香川労災病院長 多田 慎也
◎調達機関番号 590 ◎所在地番号 37
○第34号
1 調達内容
⑴ 品目分類番号 17、31
⑵ 購入等物品及び数量 胃部集団検診車一式
2 意見の提出方法
⑴ 意見の提出期限 平成22年８月20日15時（郵送の場合は必着のこと。）
⑵ 提供先 〒763―8502 香川県丸亀市城東町３丁目３－１ 香川労災病院 会計課契約係 高田 雅基 電話0877―23―3111
3 仕様書案の交付
⑴ 交付期間 平成22年７月28日から平成22年８月20日まで。
⑵ 交付場所 上記２⑵に同じ。
4 その他 本公示の調達物品は、通常の仕様書案についての意見招請手続きをとった場合、納入期限を十分に確保することが困難となるため、入札公告までの期間及び意見の提出期間を短縮した。
5 Summary
⑴ Classification of the products to be procured : 17, 31
⑵ Nature and quantity of the products to be purchased : Mass stomach X-Ray clinical van 1 Set
⑶ Time-limit for the submission of comments : 3 : 00 PM 20, August, 2010
⑷ Contact point for the notice : Masaki Takata Procurement Section, Accounting Division, Kagawa Rosai Hospital 3―3―1 Joutou-cho Marugame-shi Kagawa-ken 763―8502 Japan TEL 0877―23―3111

意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。
平成22年７月28日
独立行政法人 日本貿易振興機構
副理事長 中富 道隆
◎調達機関番号 579 ◎所在地番号 13
1 調達内容
⑴ 品目分類番号 30
⑵ 購入等特定役務及び数量 ＷＡＮおよびインターネット接続サービス一式
2 意見の提出方法
⑴ 意見の提出期限 平成22年８月27日17時00分（郵送の場合は必着のこと）
⑵ 提出先 〒107―6006 東京都港区赤坂１―12―32アーク森ビル６階 日本貿易振興機構 企画部情報システム課 電話（03）3582―5320
3 仕様書案の交付
⑴ 交付期間 平成22年７月28日から平成22年８月25日まで
⑵ 交付場所 上記２⑵に同じ
4 Summary
⑴ Classification of the services to be procured : 30
⑵ Nature and quantity of the services to be required : WAN and Internet Network systems 1 set
⑶ Time-limit for the submission of comments : 5 : 00 p.m. Aug 27th, 2010
⑷ Contact point for the notice : Information systems division, planning department, Japan External Trade Organization, ARK Mori Building 6F, 12―32 Akasaka 1-Chome, Minato-ku, Tokyo 107―6006 Japan. TEL : 81―3―3582―5320