movement instructions, Nursing instructions, Instructions of DPC, critical path, Name of disease instructions etc. The section system ; Medical affairs accounts, DPC request, automatic second coming reception system, medical examination and treatment costs management, credit management, compounding of medicines support (prescription inspection, compounding of medicines machinery connection, medicine information, injection compounding of medicines) medicine management guidance, human-blood-based manufactured blood, laboratory deter mination support (inspection such as biochemistry／blood／serum／the outside order, drawing blood pipe preparations, a drawing blood reception system, bacteria, pathological examination connection etc) endoscopy support (endoscope image, report management etc), accounts-related (financial accounting, administrative accounting) nursing support (nursing care plan, nursing record, nursing memo), rehabilitation, the system which manages the lunch section. This system assumes to be Multi-vendor-system which integrates the control, reference and management of all our facility's business in.

② This system will provide the Communication tools and Data formats under international standard specification, such as newest DICOM or HL7 protocols to exchange the patient besic information, the clinical data and the image information smoothly.

③ This system would possess functions to manage the long-term and the comprehensive clinical information of the patients independently from the database for business use.

④ This system would possess enough data processors, terminal units and networks to attain requirements mentioned this requirement.

⑤ This system would have interfaces for the medical equipment and system used in each section of our facility.

⑥ The system would be indispensable for the hospital continuous service must be automated for 24 hours. It can be applied safely and easily. And this system fully must take measures to cope with the system failure.

⑦ The access permission of the data corresponding to a user class must be to set up. The protection of the privacy and security countermeasures are sufficient.

⑧ This system must be based on the newest technology to support hospital management and medical care task. And the operational environment and quickresponse ability can cope with our actual medical care. The secondary utilization of the data must be possible.

⑸ Time-limit for the submission of the requested material : 5 : 00 P.M December 15, 2010

⑹ Contact point for the notice : Taku Shiomi, Director, Accounting Division, National Hospital Organization Osaka National Hospital, 2―1―14 Houenzaka Chuoku Osaka-shi Osaka-fu 540―0006, Japan. TEL 06―6942―1331 ext. 2325

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成22年10月1日

国立大学法人名古屋大学

財務担当理事　高橋　　誠

◎調達機関番号　415　◎所在地番号　23

○第３号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ＮＩＣＵ・ＧＣＵシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成23年度第２・四半期以降

⑸　調達に必要とされる基本的な要求要件

Ａ　新生児モニタリングシステム

ａ　本システムの生体情報モニタリングシステムは、ベッドサイドモニタ及び各種検査、治療機器等から常時生体情報を収集・表示するとともにサーバーにデータを蓄積し、二次的処理が可能であること。

ｂ　セントラルモニターを有し、連続的集中監視・不整脈監視が行えること。また、サーバーに蓄積された元及び加工データの表示・記録が可能であること。

ｃ　将来の機能拡張に対応が可能であること。

Ｂ　情報マネジメントシステム

ａ　本システムは、オンライン及びオフラインで患者データを入力することによりＮＩＣＵ業務（入退室管理、各種文書の作成等）の支援を可能とし、収集されたデータは、サーバーに保存できること。

ｂ　現有の医療情報システム、管理システム、ＮＩＣＵ患者管理マネジメントシステムとリンクが可能であり、基本データ及び臨床検査等の読み込み、表示やプリントアウトが可能であること。

Ｃ　シーリングペンダント

各種医療機器（シリンジポンプ、輸液ポンプ、ベッドサイドモニタ、情報用のＰＣ等）を載せるシーリングペンダントを各ベッドに一式設置し、夜間照明システムを有すること。

Ｄ　保育器システム

本システムは、閉鎖型保育器と、開放型保育器、搬送用保育器を有すること。

Ｅ　気管支・喉頭内視鏡

観察、診断、処置、検査を行うための内視鏡を有すること。

Ｆ　重症患者を管理するために必要な以下に記載する各種外部医療機器を有すること。

ａ　超音波診断装置

ｂ　脳波モニタシステム

ｃ　新生児人工呼吸器

ｄ　ＭＲＩ対応人工呼吸器

ｅ　クリーンベンチ

ｆ　光線治療器

Ｇ　安定で迅速な保守・支援体制、ユーザー利用日の支援を十分行える体制であること。

2　資料及びコメントの提供方法　上記1⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成22年11月19日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒466―8560 名古屋市昭和区鶴舞町65　名古屋大学医学部・医学系研究科経理課契約室長　安田　浩明　電話052―744―2799

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成22年10月１日から平成22年11月19日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成22年10月８日14時00分

⑵　開催場所　名古屋大学医学部附属病院中央診療棟３階会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : NICU・GCU system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Neonatal monitering system

a　This system should be able to continuously collect bio-information data derived from patient monitors and various clinical & laboratory testing devices, store these data in the server system and process data.

b　Central monitoring system should be included and be able to survey patient information and to analyze arrhythmia continuously. This system should be able to display and record patient original and processd data stored in the server.

c　This system should has sufficient flexibility and potentiality for system upgrading in the future.