(7) Positoron Emission Tomography-CT
① The PET scanner should be combined with multi-slice X-ray computed tomography.
② Image fusion as PET＋CT, and PET＋MRI should be able to proceed.
(8) Digital X-rays equipment to mammography
① The system should be digital X-rays equipment to mammography.
② The system must be able to digital stereotactic mammography.
③ The system should equip the high resolution monitor system for mammography diagnosis.
(9) High Radiationtherapy System
① The system should be capable to multipurpose treatment.
② The system should equip multi-leaf collimator with high shielding performance.
③ The system should be designed safely and user-friendliness operability, in addition, conformable to high-precision radiationtherapy.
5 Time-limit for the submission of the requested material : 5 : 00 P.M. March 22, 2011
6 Contact point for the notice : Yuji Noda, Manager, Facilities Planning Division, Department of Facilities Planning, National Hospital Organization Headquarters, 2―5―21 Higashigaoka Meguro-ku Tokyo-to 152―8621, Japan. TEL 03―5712―5071

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成23年2月4日
独立行政法人国立病院機構相模原病院長　秋山　一男
◎調達機関番号　597　◎所在地番号　14
○第 04601 号(№.04601)
1 調達内容
(1) 品目分類番号　22、31
(2) 導入計画物品及び数量　血管連続撮影装置　1式
(3) 調達方法　購入等
(4) 導入予定時期　平成23年度第2・四半期
(5) 調達に必要とされる基本的な要求要件
① フラットパネルディテクター搭載型シングルプレーン血管撮影装置であること。
② 多目的な手技に対応可能な装置であること。
③ 操作性およびランニングコストに優れた装置であること。
④ 画像情報等はＰＡＣＳ管理とし、ＲＩＳ及びＨＩＳとのネットワーク接続が可能なこと。
⑤ 大視野且つ高精細なＦＰＤ（フラットパネルディテクター）を搭載し、自動回転機能を有すること。
⑥ 被ばく低減機能を有し、高画質な透視画像・撮影画像が迅速に提供されること。
⑦ 高速・高画質なＣＴ-likeイメージ機能を有することでＩＶＲ環境を支援すること。
⑧ ３Ｄワークステーションを有し、多種の画像解析ソフトを有すること。
2 資料及びコメントの提供方法　上記1(2)の物品に関する一般的な参考資料及び同(5)の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
(1) 資料等の提供期限　平成23年3月25日17時00分（郵送の場合は必着のこと。）
(2) 提出先　〒252―0392 神奈川県相模原市南区桜台18―1　独立行政法人国立病院機構相模原病院企画課長　亘　　松男　電話042―742―8311　内線4100
3 説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
(1) 交付期間　平成23年2月4日から平成23年3月25日まで。
(2) 交付場所　上記2(2)に同じ。
4 その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5 Summary
(1) Classification of the products to be procured : 22, 31
(2) Nature and quantity of the products to be purchased : Single plane angiography with a flat panel detector 1set
(3) Type of the procurement : Purchase
(4) Basic requirements of the procurement :
① Single plane angiography with a flat panel detector
② The system should be multifunctional.
③ The system should be user-friendly and be cost-effective.
④ The imaging data should be controlled under PACS and be connected to RIS and HIS.
⑤ The system should mount a flat panel detector that has a large viewing field, a high-definition imager and an automatic rotating system.
⑥ The system should have irradiation reducing functions, providing high quality fluorographic and radiographic images without any delays.
⑦ The system should provide the high-speed and high-quality CT-like imaging functions, assisting IVR environment.
⑧ The system should mount a 3D-work station, possessing a various types of image analyzing software.
(5) Time-limit for the submission of the requested material : 5 : 00 P.M March 25, 2011
(6) Contact point for the notice : Matsuo Watari, Director, Accounting Division, National Hospital Organization sagamihara National Hospital, 18―1 Sakuradai Minami-ku Sagamihara-shi Kanagawa-ken 252―0392, Japan. TEL 042―742―8311 ext.4100

意見招請に関する公示

次のとおり調達物品の仕様書案の作成が完了しましたので、仕様書案に対する意見を招請します。
平成23年2月4日
沖縄振興開発金融公庫　会計役　許田　盛秀
◎調達機関番号　119　◎所在地番号　47
1 調達内容
(1) 品目番号　71、27
(2) 調達物品及び数量　業務用端末機器等の調達　一式
2 意見の提出方法
(1) 意見書の提出期限　平成23年2月25日午後5時
(2) 提出先　〒900―8520 沖縄県那覇市おもろまち1―2―26　沖縄振興開発金融公庫　庶務部庶務課　高良　直樹　電話098―941―1700
3 仕様書案の交付
(1) 交付期間　平成23年2月4日から平成23年2月25日
(2) 交付場所　上記2(2)に同じ
4 Summary
(1) Classification of the products to be procured : 71, 27
(2) Nature and quantity of the products to be required : Information system 1 set
(3) Time-limit for the submission of comments : 5 : 00 PM February 25, 2011
(4) Contact point for the notice : Takara Naoki General Service Division, The Okinawa Development Finance Corporation. 1―2―26 Omoromachi, Naha City, Okinawa, 900―8520 Japan, TEL098―941―1700

随意契約

随意契約に関する公示

次のとおり随意契約について公示します。
平成23年2月4日
契約責任者
独立行政法人国立印刷局
財務担当理事　大野　雅人
◎調達機関番号　562　◎所在地番号　13
1 調達内容
(1) 品目分類番号　12
(2) 購入等件名及び数量
券面検査装置改造　1式
2 随意契約の予定日　平成23年2月25日
3 随意契約によることとする「政府調達に関する協定」の規定上の理由　ｂ「排他的権利の保護」及びｄ「互換性」
4 随意契約を予定している相手方の名称
日本電気㈱
5 担当部局　〒105―8445 東京都港区虎ノ門二丁目2番4号　独立行政法人国立印刷局財務部調達業務担当　佐瀬　正孝　電話03―6893―4619