◎調達機関番号　415　◎所在地番号　12
○第４号
１　調達内容
⑴　品目分類番号　４、32
⑵　導入計画物品及び数量
循環器系疾患診断治療システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成23年度10月以降
⑸　調達に必要とされる基本的な要求要件
①　Ａｎｇｉｏ／ＣＴは大視野平面検出器での透視および撮影が可能であること。
②　ＣＴ装置は16列以上の自走式マルチスライスＣＴであること。
③　心臓カテーテル装置は平面検出器を搭載し、同時二方向より透視及び撮影が可能であること。
④　回転ＤＳＡ撮影ができ、高速に３Ｄ画像再構成が可能であること。
⑤　デジタル画像処理機能及びＱＣＡ解析機能などを装備すること。
⑥　患者及び従事者の被ばく低減機能を有し、リアルタイムで線量を表示できる機能を装備していること。
⑦　外部画像記憶装置及び透視画像録画システムを装備すること。
⑧　ＤＩＣＯＭ3.0に準拠した画像出力に対応し、当院画像ネットワークサーバに出力できること。
⑨　放射線情報システム（ＲＩＳ）との通信が可能であること。
⑩　循環器画像ネットワークサーバに出力できること。
⑪　その他循環器対応型の周辺機器（造影剤自動注入器、ポリグラフ装置、Ｘ線防護用具、及びＸ線測定機器等）を装備すること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年４月18日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒260―8677　千葉市中央区亥鼻１－８－１　千葉大学医学部附属病院管理課用度係　小川代枝生　電話043―226―2236
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成23年３月１日から平成23年４月18日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成23年３月15日14時00分
⑵　開催場所　千葉大学医学部附属病院第一会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 4, 32
⑵　Nature and quantity of the products to be purchased : CT Angio System for Interventional Radiology 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The Angio／CT system shall have the capability of fluoroscopy and radiography with the large size flat panel detector.
②　The CT system shall have a self propelled MDCT having 16 or more detector rows.
③　The cardiac angiography system should be a bi-plane angiographic system equipped with flat panel detectors and the system shall have a capability of fluoroscopy and radiography from simultaneous 2 direction.
④　The system shall have the capability of high speed reconstruction of 3D image from rotational DSA.
⑤　This system shall have the capability of digital image-processing function and QCA analysis function.
⑥　The system shall have the function of reducing exposure dose for the operators and patients and display the dose information in real time.
⑦　The system shall have the external and fluoro image recoding system.
⑧　The digital image output from equipment is based on DICM3.0. To have interfacesconnectable with the image servers that are already installed in Chiba University Hospital.
⑨　Correspond to the communication function with a radiation information system.
⑩　This system should be able to transmit the image data to the existing Cardiac imagenetwork severs.
⑪　The system shall have the peripheral equipment (contrast medium injecting device, Polygraph, Protection kit of radiation, Dose meter).
⑸　Time limit for the submission of the requested material : 5 : 00 PM 18, April, 2011
⑹　Contact point for the notice : Yoshio Ogawa, Procurement Section, Accounting Division, Chiba University Hospital, 1―8―1 Inohana Chuo-ku Chiba-shi 260―8677 Japan, TEL 043―226―2236

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 23 年３月１日
国立大学法人新潟大学長　下條　文武
◎調達機関番号　415　◎所在地番号　15
○第４号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　歯科用チェアーユニット　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成23年度３月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　歯科用チェアーユニット本体を有すること。
Ｂ　エアータービンハンドピースを有すること。
Ｃ　マイクロモーターを有すること。
Ｄ　マイクロモーター用ハンドピースを有すること。
Ｅ　３ウェイシリンジを有すること。
Ｆ　バキュームシリンジを有すること。
Ｇ　超音波スケーラーを有すること。
Ｈ　超音波スケーラーハンドピースを有すること。
Ｉ　無影灯を有すること。
Ｊ　含水装置を有すること。
Ｋ　オペレーティングスツールを有すること。
Ｌ　診療に必要な操作が支障なく行えること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年４月18日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒951―8520 新潟市中央区旭町通１番町754番地　新潟大学医歯学総合病院管理運営課契約係　宮路　真一　電話025―227―0970
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成23年３月１日から平成23年４月18日まで。
⑵　交付場所　上記２⑵に同じ。
４　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成23年３月８日11時00分
⑵　開催場所　新潟大学医歯学総合病院管理棟４階大会議室
５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
６　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Dental Chair Unit 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　Dental chair shall be provided.
B　Air turbine hand-piece shall be provided.
C　Micro motor instrument shall be provided.
D　Micro motor hand-piece shall be provided.
E　3-way syringe shall be provided.
F　Vacuum syringe shall be provided.
G　Ultrasonic scaler shall be provided.