6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Angiography system interconnected between operation theater and section of radiology 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　Common denominator eyes
A　Fluoroscopy and serial radiography are available
B　Promptly obtain the patient information from hospital information system (HIS) and easily transfer them to the angiography system
C　Confirm the image and information obtained from the angiography through the intra-hospital medical image management system (PACS)
D　Perform the various radiography and treatment while observing an image of the other modality including CT MRI and ultrasonography with a monitor in the laboratory at real time
②　Angiography system of operation theater
E　Take the placement enough for the cardiovascular surgery using an extracorporeal circulation support system in the operating room
F　The free space is enough and available for other general operation when a new angiography system is not used
G　Cone-Beam CT is available
③　Cardiac angiography system of section of radiology
H　Perform Fluoroscopy and Angiography corresponding to the diagnostic images of various heart disease
I　Obtain the adequate function for the various endovascular treatment against heart and peripheral disease
J　Have a proper function which acquire bidirectional image simultaneously
⑸　Time limit for the submission of the requested material : 5 : 00 PM 11, May, 2011
⑹　Contact point for the notice : WAKISAKA Kazunori, supplies I Section, Financial Division, Hokkaido University Hospital, Kita 14 Nishi 5 Kita-ku Sapporo-shi 060―8648 Japan, TEL 011―706―7401

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成23年3月28日
国立大学法人　山梨大学長　前田秀一郎
◎調達機関番号　415　◎所在地番号　19

○第３号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　磁気共鳴断層撮影システム（３Ｔ）　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成23年度10月以降
⑸　調達に必要とされる基本的な要求要件
A　超伝導マグネットで3.0Ｔ以上の磁場強度を有すること。
B　中枢神経機能画像、血流解析、拡散強調画像などの高度で先進的な診断解析が可能であること。
C　全身に対する撮影や広範囲の撮影が可能であること。
D　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、読影システムおよび医用画像管理システム（ＰＡＣＳ）と接続可能であること。
E　システムを設置するための工事を含むこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年５月13日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒409―3898 山梨県中央市下河東1110番地　山梨大学医学部管理課　仲嶋　宏治　電話055―273―9224
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成23年３月28日から平成23年５月13日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成23年４月４日14時00分
⑵　開催場所　山梨大学医学部キャンパス内管理棟２階小会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : magnetic resonance imaging system (3T) 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must have the magnetic field strength of 3.0 T.
B　The system should include high performance applications, including neural functional image, perfusion analysis, diffusion imaging.
C　The system must allow for the scanning of whole body.
D　The system should have an interface to communicate with the existing hospital information system (HIS), radiology information system (RIS), reporting system, picture archiving and communication system (PACS).
E　The cost of construction and installation should be included.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 13, May, 2011
⑹　Contact point for the notice : Kouji Nakazima, Faculty of Medicine Accounting Division, University of Yamanashi, 1110 Simokato Chuo-shi Yamanashi 409―3898 Japan, TEL 055―273―9224

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成23年3月28日
国立大学法人徳島大学長　香川　　征
◎調達機関番号　415　◎所在地番号　36

○第９号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　磁気共鳴断層装置（ＭＲＩ 3.0Ｔ）システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成23年度３月以降
⑸　調達に必要とされる基本的な要求要件
A　本システムは、磁気共鳴断層装置システム及び必要な画像処理・解析システムを含んでいること。
B　磁気共鳴断層装置システム及び画像処理・解析システムは、以下の要件を満たすこと。
ａ　優れた組織コントラスト及び時間分解能を有し、迅速・正確な画像解析や処理が可能であること。静磁場強度は3.0Ｔ以上であること。
ｂ　操作性が良く、高精度で高速な三次元画像処理や潅流解析等の４次元画像処理が可能であること。
ｃ　多様な代謝及び機能情報の取得技術が搭載されており、proton以外の他核種にも対応していること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年５月13日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒770―8503 徳島市蔵本町２―50―１　国立大学法人徳島大学病院経理調達課調達係　大坂　誠一　電話088―633―9559