B　リモートアフターローディング式高線量率密封小線源治療システムは、以下の要件を満たすこと。
a　高線量率密封小線源を用いて遠隔操作によりがんの放射線治療ができること。
b　密封小線源治療が適用可能な全ての臓器がんに対応できること。
c　高度な照射精度が保証されること。
C　３次元放射線治療計画システムは、以下の要件を満たすこと。
a　デジタルＸ線テレビシステムや既設の統合画像診断システムからオンラインで取得したＸ線写真、コンピュータ断層像、核磁気共鳴画像を利用して３次元放射線治療計画ができること。
b　３次元放射線治療計画に基づき、リモートアフターローディング式高線量率密封小線源治療システムを操作し高精度放射線治療ができること。
c　治療計画結果を既設の病院放射線部情報システムにオンラインで転送し保存できること。
D　デジタルＸ線テレビシステムは、以下の要件を満たすこと。
a　任意の角度のＸ線写真とコンピュータ断層写真が撮影できること。
b　操作性が良く優れた分解能を有し、迅速に正確な診断ができること。
c　取得した画像をオンラインで既設の統合画像診断システムに転送し保存できること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年５月13日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒770―8503　徳島市蔵本町２―50―１　国立大学法人徳島大学病院経理調達課調達係　大坂　誠一　電話088―633―9559

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成23年３月28日から平成23年５月13日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成23年４月５日16時30分
⑵　開催場所　国立大学法人徳島大学病院入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Remote afterloading high-dose-rate brachytherapy system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system should be composed of remote afterloading high-dose-rate brachytherapy system, three-dimensional radiation therapy planning system, and digital x-ray television system.
B　Remote afterloading high-dose-rate brachytherapy system should fulfill the conditions described below.
a　It should be able to provide remote afterloading brachytherapy of cancer using high-dose-rate sealed radioactive source.
b　It should enable to deliver high-dose-rate brachytherapy for every type of cancer that becomes a candidate of brachytherapy.
c　It should have a quality assurance system to maintain a highly precise radiation therapy.
C　Three-dimensional radiation therapy planning system should fulfill the conditions described below.
a　It should be able to create three-dimensional radiation therapy planning based on the radiography, computed tomography, and magnetic resonance imaging transferred on line from digital x-ray television system and existent picture archiving and communication system.
b　It should control remote afterloading high-dose-rate brachytherapy system and provide highly precise radiation therapy based on three-dimensional radiation therapy planning.
c　It should be able to transfer planning data to existent radiation oncology information system on line and save data in it.
D　X-ray television system should fulfill the conditions described below.
a　It should provide radiography at variable angle and computed tomography.
b　It should provide excellent patient access, rotational scans with precision quickly, and high resolution imaging to enable quick accurate diagnosis.
c　It should transfer imaging data to existent picture archiving and communication system and save data in it.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 13, May, 2011
⑹　Contact point for the notice : Seiichi Osaka, Supplies, University Hospital, Accounting and Supply Affairs Section , National University Corporation, The University of Tokushima, 2―50―1 Kuramoto-cho Tokushima-shi 770―8503 Japan, TEL 088―633―9559

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 23 年３月 28 日
国立大学法人愛媛大学長　柳澤　康信
◎調達機関番号　415　◎所在地番号　38
○第１号
１　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量
ＣＴアンギオシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成23年度12月以降
⑸　調達に必要とされる基本的な要求要件
①　同時に二方向の血管造影撮影が行えること。
②　血管造影撮影とＣＴ撮影が同一の寝台で行えること。
③　血管造影によるデジタル透視、デジタル撮影機能を有すること。
④　多断面高速ヘリカルスキャンによるＣＴ撮影が可能であること。
⑤　高精細な三次元立体画像を構築できること。
⑥　本システムによる診療を安全かつ円滑に行う上で、必要不可欠な附属周辺機器を有すること。
⑦　本院既存の放射線科情報システム（横河電機（株）製ＲＩＳ）とMWM接続し、患者属性情報及びオーダー番号（アクセッション番号）が取得できること。
⑧　本院既存の医用画像保存通信システム（富士フイルム（株）製ＰＡＣＳ）への画像転送機能を有すること。
２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成23年５月16日17時15分（郵送の場合は必着のこと。）
⑵　提供先　〒791―0295 東温市志津川　国立大学法人愛媛大学医学部経営管理課調達第二チーム　細川　富生　電話089―960―5154
３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成23年３月28日から平成23年５月16日まで。
⑵　交付場所　上記２⑵に同じ。