4）Ventricular Assist System must have a backup solenoid valve.
5）NCVC-assisted artificial heart pump can be operated.
6）Ventricular Assist System must be driven by an internal battery in case of power outage.
7）Operation switches must be a LC touch panel.
8）ICU／PICU department system should be consisted with client(s) for clinical infomation and server(s) for storage.
9）ICU department system has the functionalities for acquiring existing patient monitor(s) and charting onto anesthetic chart automatically.
10）ICU／PICU department system has the functionalities for retrieving and statistics with clinical database.
11）Clinical information management system has the functionalities for communication with HIS regard to patient demographics and accounting information in terms of anesthetic reimbursement.
12）Patient information system should be consisted with bedside monitor(s) and central station(s).
13）Patient monitor has the functionalities for measuring and monitoring vital signs parameters i. e. ECG, Respiratory, APB, NIBP, SpO2 and Temp simultaneously.
14）Central station has the functionalities for managing all bedside monitor(s) in each operational theater(s) remotely.
15）Anesthesia record system must renew the existing hardware of the system.

⑸　Time-limit for the submission of the requested material : 5 : 00 P.M July 15, 2011

⑹　Contact point for the notice : Aoi Shin, General Affairs Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujisirodai, Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext.2982

○第２号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　手術用顕微鏡　1式

⑶　調達方法　購入等

⑷　導入予定時期　平成23年度第３・四半期

⑸　調達に必要とされる基本的な要求要件

1）倍率変倍機構の総合倍率は12倍以上であること。
2）対物レンズの作業距離は200㎜～400㎜の範囲で無段階で変更できること。
3）対面手術にも対応していること。
4）オートバランス機能を有すること。
5）キセノン光源を採用していること。
6）アイピース内で内視鏡の映像を観察できること。
7）ナビゲーションシステムと接続可能なこと。
8）スタンドストッパーを有すること。
9）術中の血管観察用モジュールを搭載していること。
10）術中の腫瘍観察用モジュールを搭載していること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成23年７月15日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒565―8565 大阪府吹田市藤白台５―７―１　独立行政法人国立循環器病研究センター　総務課契約係　新　　葵　電話06―6833―5012　内線2982

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成23年５月31日から平成23年７月15日まで。

⑵　交付場所　上記２⑵に同じ。

4　その他　この導入計画の詳細は導入説明書による。

なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Surgical Operation Microscope 1 set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

1）Observed by more than 12 × total magnification.
2）Variable working distance range of more than 200㎜ to 400㎜.
3）Possible for face to face procedure.
4）Autobalance function to balance the microscope.
5）Integrated Xenon illumination system.
6）Possible to inject the endoscopic viewinto the eyepiece.
7）Possible to connect with the navigation system.
8）Integrated locking pedal in the floorstand.
9）Intraoperative blood vessel observation is possible.
10）Intraoperative tumor observation is possible.

⑸　Time-limit for the submission of the requested material : 5 : 00 P.M July 15, 2011

⑹　Contact point for the notice : Aoi Shin, General Affairs Division, National Cerebral and Cardiovascular Center, 5―7―1 Fujisirodai, Suita-shi, Osaka-fu, 565―8565 Japan. TEL 06―6833―5012 ext.2982

○第３号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　注射薬自動払出装置　1式

⑶　調達方法　購入等

⑷　導入予定時期　平成23年度第３・四半期

⑸　調達に必要とされる基本的な要求要件

1）定期的な処方の払い出しは２時間以内で作業を完了できる能力があること。
2）機器の設置スペースが狭いので、１ユニットの機器で定期的な処方の払い出しが行える能力があること。
3）定期的な処方を払い出している途中に緊急の処方のオーダーが来た場合、緊急の処方を優先的に払い出す機能を持っていること。
4）無菌室で混合する必要がある処方を事前に払い出すことができる機能を持っていること。
5）50,000個の薬品を払い出したと仮定した場合、薬品を１個も破壊しない性能であること。
6）2012年に導入を予定している【電子カルテシステム】と接続できること。
7）薬品管理のために十分なセキュリティー対策ができていること。
8）薬品を搬送するカートを必要な台数用意できること。
9）装置に搭載できる薬品はアンプル、バイアル及び100㎖輸液であること。
10）薬品をトレーに払い出し、トレーを自動的にワゴンに格納できる機能がある。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成23年７月15日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒565―8565 大阪府吹田市藤白台５―７―１　独立行政法人国立循環器病研究センター　総務課契約係　新　　葵　電話06―6833―5012　内線2982

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成23年５月31日から平成23年７月15日まで。

⑵　交付場所　上記２⑵に同じ。

4　その他　この導入計画の詳細は導入説明書による。

なお、本公表内容は予定であり、変更することがあり得る。

5　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Automated preparation system of injectable drugs 1 set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

1）The equipment has an ability to complete a dispensary of a regular prescription within two hours.
2）Due to the limitation of the installation space of the equipment, it has the ability that a regular prescription can be despensed with the equipment of one unit.