2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年３月19日17時15分（郵送の場合は必着のこと。）
⑵　提供先　〒565―0871 吹田市山田丘２―15 大阪大学医学部附属病院管理課専門職員 宮田　静子　電話06―6879―5118
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年２月１日から平成24年３月19日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年２月９日13時30分
⑵　開催場所　大阪大学医学部・同附属病院共通棟３階入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Whole body magnetic resonance imaging system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must have a homogenous high magnetic field (1.5 T or more) with a superconducting magnet. The system also must have a gradient coil system which enables fast data acquisition.
B　The system must be equipped with multi-channel phased array coil systems, dedicated coil systems with high resolution, and a whole body coil system.
C　The system must be able to acquire MRI and MRA in any part of the body. The system also must have advanced applications.
D　The system must have equipments which can monitor physiological conditions and appearance of a patient.
E　Functional analysis and magnetic resonance spectroscopy must be possible with the system. The system also must be equipped with a dedicated workstation which enables image interpretation and data analysis.
F　An image network system with image servers must be constructed. The system must be able to communicate with the existing hospital information system (HIS) and radiology information system (RIS). The system must be equipped with high-speed data processing, output and transfer of DICOM images.
G　An automatic injection system of contrast medium must be equipped.
H　Phantoms must be equipped for the image quality control.
I　The system must have a remote maintenance system.
⑸　Time limit for the submission of the requested material : 5 : 15 PM 19, March, 2012
⑹　Contact point for the notice : MIYATA Shizuko, Procurement Section, Accounting Division, Osaka University Hospital, 2―15 Yamadaoka Suita-shi 565―0871 Japan, TEL 06―6879―5118

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 24 年２月１日
国立大学法人大阪大学総長　平野　俊夫
◎調達機関番号　415　◎所在地番号　27

○第５号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　全身用磁気共鳴画像診断装置（３Ｔ）　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度３月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　超伝導マグネットで均一性に優れた高磁場（3.0テスラ以上）を有し、高速撮像が可能な傾斜磁場システムであること。
Ｂ　多チャンネルフェーズドアレイコイルシステム、高解像度のコイル及び全身対応可能なコイルを有すること。
Ｃ　均一なＲＦ送信を行えること。
Ｄ　全身のＭＲＩ／ＭＲＡ撮像の他、先進的アプリケーションを有すること。
Ｅ　患者の生理的状態及び外観を監視するシステムを有すること。
Ｆ　機能解析、磁気共鳴スペクトロスコピーが可能で、専用の画像診断及び解析ワークステーションを有すること。
Ｇ　画像サーバを持つ画像ネットワークシステムが構築でき、病院情報管理システム（ＨＩＳ）及び放射線部情報システム（ＲＩＳ）と接続し、必要な情報の受け渡し通信ができること。さらに、高速データ処理能力を備えＤＩＣＯＭでの高速画像出力・転送が可能であること。
Ｈ　造影剤自動注入装置を有すること。
Ｉ　精度管理用ファントムを有すること。
Ｊ　ＭＲＩ専用シリンジポンプを有すること。
Ｋ　リモートメンテナンスシステムを有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年３月19日17時15分（郵送の場合は必着のこと。）
⑵　提供先　〒565―0871 吹田市山田丘２―15 大阪大学医学部附属病院管理課専門職員 宮田　静子　電話06―6879―5118
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年２月１日から平成24年３月19日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年２月９日13時45分
⑵　開催場所　大阪大学医学部・同附属病院共通棟３階入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Whole body magnetic resonance imaging system (3 Tesla) 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system must have a homogenous high magnetic field (3.0 T or more) with a superconducting magnet. The system also must have a gradient coil system which enables fast data acquisition.
B　The system must be equipped with multi-channel phased array coil systems, dedicated coil systems with high resolution, and a whole body coil system.
C　Uniform RF transmission is required.
D　The system must be able to acquire MRI and MRA in any part of the body. The system also must have advanced applications.
E　The system must have equipments which can monitor physiological conditions and appearance of a patient.
F　Functional analysis and magnetic resonance spectroscopy must be possible with the system. The system also must be equipped with a dedicated workstation which enables image interpretation and data analysis.
G　An image network system with image servers must be constructed. The system must be able to communicate with the existing hospital information system (HIS) and radiology information system (RIS). The system must be equipped with high-speed data processing, output and transfer of DICOM images.
H　An automatic injection system of contrast medium must be equipped.
I　Phantoms must be equipped for the image quality control.
J　The system must be equipped with an IV pump system which can be used in high magnetic field.
K　The system must have a remote maintenance system.