B　本デジタル交換機はアナログ系回線、デジタル系回線、ＩＳＤＮ64回線を接続できる機能を有し、既存の電話機を収容出来る（既設同等品との取替も可能とする）とともに、将来内線実装宇治地区2100回線、熊取地区800回線、犬山地区400回線まで容易に増設できるネットワーク架構成であること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成24年４月９日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒606―8501 京都市左京区吉田本町　京都大学情報部情報推進課財務掛　廣瀬幸一　電話075―753―7403

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年３月６日から平成24年４月９日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成24年３月19日11時00分

⑵　開催場所　京都大学学術情報メディアセンター北館会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 15, 28

⑵　Nature and quantity of the products to be purchased : Digital PBX 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The product must be a digital PBX which has connectivity to analog, digital phone lines, INS 64 lines and private PHS (Personal Handy-phone System) base stations. It must have unified message function, short message access from PHS mobile terminals, and be capable of additional IP phone interfaces. It must be capable of connecting with PHS terminals, IP phone terminals and dual mode terminals. The product includes the terminal equipment recognized under "Telecommunication Business Law Article 53."

B　The product must be a digital PBX which has connectivity to analog, digital phone lines, and INS 64 lines. It must be capable of accommodating existing telephone equipment (may be replaced with the equivalent) and have the capability to easily increase its capacity to accommodate up to Uji Campus 2,100 lines, Kumatori Campus 800 lines and Inuyama Campus 400 lines.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 9, April, 2012

⑹　Contact point for the notice : Kouichi Hirose, Accounting Section, Information Management Department, Kyoto University, Yoshida-Hommachi Sakyo-ku Kyoto-shi 606―8501 Japan, TEL 075―753―7403

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 24 年３月６日

国立大学法人広島大学

分任契約担当職　病院長　茶山　一彰

◎調達機関番号　415　◎所在地番号　34

○第３号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ハイブリッド手術対応型血管Ｘ線撮影システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成24年度９月以降

⑸　調達に必要とされる基本的な要求要件

A　脳血管障害、循環器疾患など全身の診断・治療に対応した正側同時２方向撮像可能なシステムであること。

B　基幹となる血管造影装置は患者へのアプローチを妨げないために、アームの退避が可能なこと。

C　患者を移動することなく、頭部から下肢までの全身を検査可能な機能を有すること。

D　Ｘ線変換効率、濃度分解能が高いフラットパネルディテクタ搭載装置であること。

E　パルスレート可変ディジタル透視、自動付加フィルタ挿入など、複合的に被曝低減可能なプログラムを有すること。

F　本システムを用いて開頭・開胸・開腹手術が実施できること。

G　高精細３次元ワークステーションを備え、ＩＶＲ（ＴＡＶＩ、心筋焼灼術、脳血管内手術など）を支援するシステムを考慮すること。

H　本システムの円滑な運営のため、必要な周辺機器を考慮すること。

I　放射線部情報システムと接続し、患者属性、撮影情報など必要事項を受け渡す機能を有すること。

J　リモートメンテナンスサービスが可能であり、保守サービス体制が確立されていること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成24年４月23日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒734―8551 広島市南区霞一丁目２番３号　広島大学病院財務グループ主査（物品主担当）　曽根　淳一　電話082―257―5025

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年３月６日から平成24年４月23日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成24年３月13日16時00分

⑵　開催場所　広島大学病院外来棟２Ｆ会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Hybrid Operation-directed X-ray Angiographic System 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Biplane system for diagnosis and treatment of whole body including cerebrovascular and cardiovascular

B　Evacuation of the main arm is possible for the purpose of smooth approach to the patients

C　Function that can examine the whole body from the head to lower limbs without moving the patients

D　Device mounted with the flat panel detector which is high in X-rays exchange efficiency and density resolution

E　Radiation exposure decrease is compositely available including pulse rate variable digital investment and the automatic addition filter insertion

F　That we can perform craniotomy, thoracotomy, abdominal operation using this system

G　To consider a system supporting IVR (trans-aortic valve implantation, cardiac ablation, brain intravascular operation et.) with a high-definition three-dimensional work-station

H　To consider necessary peripheral equipment for the smooth administration of this system

I　Function to deliver necessary matters such as patients attribute, scan information with connected to information system in the radiation department

J　Maintenance service system being established including remote maintenance service

⑸　Time limit for the submission of the requested material : 5 : 00 PM 23, April, 2012

⑹　Contact point for the notice : Junichi Sone, Hiroshima University Hospital Finance Group, 1―2―3 Kasumi Minami-ku Hiroshima-shi 734―8551 Japan, TEL 082―257―5025