6　Summary
⑴　Classification of the products to be procured : 13
⑵　Nature and quantity of the products to be purchased : Patient meal cooking system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The facility／equipment shall be for producing and serving any diets meals to the patients.
B　The facility shall have Cook-Serve system.
C　The facility shall have temperature control system for kitchen, food storage and food storage equipment.
D　The facility shall be able to produce formula milk for newly-born infant and baby.
E　The facility shall be able to cook and serve individual／personal diet meal.
F　All equipment shall be of the energy-saving models.
G　It continues every day washes up to 600 place settings three times a day.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 12, April, 2012
⑹　Contact point for the notice : Toshiyuki Sasaki, Hospital Procurement Section 1, Accounting Division, The National University Shimane University Faculty of Medicine, 89―1 Enya-cho Izumo-shi 693―8501 Japan, TEL 0853―20―2548

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 24 年 3 月 12 日
国立大学法人九州大学事務局長　本木　章喜
◎調達機関番号　415　◎所在地番号　40
○第 10 号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　検体検査・病理診断総合システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度３月以降
⑸　調達に必要とされる基本的な要求要件
A　緊急検査システムは、ルーチン検査用機器との互換性や運用も考慮された生化学自動分析装置、自動血球分析装置、血液凝固線溶分析装置、免疫測定装置、血液ガス分析装置で構成されており、既存の検査情報システムと接続し連携できること。また、ＤＮＡシーケンサ、結核菌群同定装置、全自動血液培養検査装置、呼吸機能検査装置を有していること。
B　総合血液検査搬送システムは、自動血球分析装置、塗抹標本作成装置を搬送ラインで接続し、充分なバックアップがとれる処理能力を有すること。また、血液細胞分類装置とルーチン用血液凝固線溶分析装置を有し総合血液検査情報システムを構築し、既存の検査情報システムと接続し連携できること。
C　病理診断支援システムは、処理能力効率化を考慮したルーチン用及び迅速用自動固定包埋装置、パラフィン包埋装置、回転式ミクロトーム、感染防止対策用クリオスタット、自動免疫組織化学染色装置、ＨＥ染色ならびに特殊染色用自動染色装置より構成されること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年４月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒812―8581 福岡市東区箱崎六丁目10番１号　国立大学法人九州大学財務部調達課備品調達係　吉田　孝士　電話092―642―4327
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年３月12日から平成24年４月27日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年３月22日14時00分
⑵　開催場所　九州大学病院北棟２階共用会議室２
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Automation system for clinical laboratory tests and pathological diagnosis support system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The compatibility and the use with the apparatus for routine examinations are comprised of biochemistry analyzer, automatic blood cell analyzer, blood coagulation analyzer, automatic immunology analyzer, blood gas analyzer and urgently connect the laboratory system to existing laboratory information system and can cooperate. In addition, the following equipments are included. The DNA sequencer, tuberculosis mycol identification device, automatic blood culture device, and respiratory function test equipment.
B　The total blood test transport system should connect automatic blood cell analyzer and automatic blood cell staining equipment in a transportation line, and have the throughput which can take sufficient backup. Moreover, the thing which it has blood-cells classification equipment and blood coagulation analyzer for routines, a total blood test information system is built, and it connects with the existing laboratory information system, and can cooperate.
C　Pathological diagnosis support system have high processing capacity and efficiency. The system is composed of automatic fixed and embedded unit, tissue embedding unit, rotary microtome, infection protected cryostat, automated immunostaining stainer and automated slide stainers for HE and special stainings.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 27, April, 2012
⑹　Contact point for the notice : YOSHIDA Takashi, Equipment Procurement Section, Procurement Division, Kyushu University, 6―10―1 Hakozaki Higashi-ku Fukuoka-shi 812―8581 Japan, TEL 092―642―4327

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 24 年 3 月 12 日
国立大学法人大分大学
理事（財務・環境担当）　岩切　平治
◎調達機関番号　415　◎所在地番号　44
○第６号
1　調達内容
⑴　品目分類番号　４、31
⑵　導入計画物品及び数量　一般Ｘ線撮影デジタルシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度９月以降
⑸　調達に必要とされる基本的な要求要件
A　一般Ｘ線撮影デジタルシステム　１式
ａ　一般撮影装置　５式
ｂ　頭部撮影装置　１式
ｃ　パントモグラフィー装置　１式
ｄ　Ｘ線ＴＶ装置　１式
ｅ　ＲＡＬＳ用透視、撮影装置　１式
ｆ　回診用X線装置　５式　等
B　現状で行われている全ての一般撮影に対応できること。
C　本院で運用中の病院情報管理システム（ＨＩＳ）、放射線画像情報システム（ＲＩＳ）、画像保存通信システム（ＰＡＣＳ）に接続可能で、撮影業務の効率化、迅速化を図れること。
D　患者、術者、スタッフに対する被ばく低減機構を有すること。
E　システムとして安全で迅速かつ簡便な操作性を有すこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年４月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒879―5593 大分県由布市挾間町医大ケ丘１丁目１番地　大分大学医学・病院事務部経営管理課用度第一係　村田　典彦　電話097―586―5270