B　The progress of the entire blood collection system should be automatically announced and displayed to standby patients.

C　The patient's information and the items tested should appear on the display, upon the input of the patient's ID number at each blood collection booth.

D　The test tubes needed should be automatically supplied and labeled, upon receiving the patients' information, including the ID numbers, names, and test items.

E　Misidentification of patients or blood collection tubes should be prevented by matching patients and their tubes.

F　The enough room for safe blood collection should be kept, and the height of chairs and blood collection tables should be adjusted.

G　Blood sample tubes should be collected from all the booths, and transferred to the laboratory testing unit.

H　Laboratory Automation System (LA system) consists of Chemistry & Immunology Line, Hematology Line, Blood Sugar Line, Coagulation Line, Urinalysis Line, and Comprehensive Blood Collection System. All laboratory sample tests are to be performed in this system.

I　The test sample sorting and pretreatment line, automatic chemistry analyzers, automatic immunoassay analyzer, multi-item hematology analyzers, automatic smear preparation, coagulation analyzers, blood glucose and glycohemoglobin analyzers and sample storage sites should be equipped in this system, and the measurement should be performed sequentially. Furthermore, individual (off-line) automatic chemistry analyzers, multi-item hematology analyzers, coagulation analyzers, urine qualitative analyzers and automatic urinary sediment analyzers, a flow cytometers, erythrocyte sedimentation rate analyzers, blood gas analyzers, automatic fecal occult blood analyzers, genetic testig analyzers, automatic blood typing analyzers should also be included.

J　Control and regulation of laboratory analyzers and the sample sorting and transport, and the quality control of measurement results should be performed in this LA system. Control and integration of the information arising from test analyzers must be enabled. The LA system must be highly integrated and connected to existing Hospital Information System. Extreme care should be paid for the security of patients' information and the countermeasure against system collapse.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 1, May, 2012

⑹　Contact point for the notice : Kiyomi Kugaya, Procurement Section, Accounting Division, The University of Tokyo Hospital, 7―3―1 Hongo Bunkyo-Ku Tokyo 113―8655 Japan, TEL 03―5800―8615

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 24 年 3 月 16 日

国立大学法人香川大学長　長尾　省吾

◎調達機関番号　415　◎所在地番号　37

○第３号

１　調達内容

⑴　品目分類番号　22

⑵　導入計画物品及び数量　周産期科女性診療科およびＭＦＩＣＵ・部門専用サーバーシステム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成24年度９月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　超音波診断装置から送信された画像（静止画・動画）を閲覧できること。

Ｂ　既存の超音波診断装置（Voluson GE社）の解析機能が搭載されていること。

Ｃ　超音波診断装置で計測した結果を受信し、自動的にレポートに反映されること。

Ｄ　作成されたレポートは、電子カルテから閲覧可能なこと。

Ｅ　データのバックアップが可能なこと。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成24年５月１日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒761―0793 香川県木田郡三木町大字池戸1750―１　香川大学医学部管理課用度第二係　川田　保明　電話087―891―2028

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年３月16日から平成24年５月１日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成24年３月23日11時00分

⑵　開催場所　香川大学医学部管理棟中会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22

⑵　Nature and quantity of the products to be purchased : Exclusive use server system for MFICU, Perinetology and gynecology 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Complete access to the image (still and movie picture) which has taken by ultrasonograph.

B　It is able to analyze data which is in the existent ultrasonograph (Voluson, GE Co.) on the server system.

C　It receives results measured by ultrasonograph and reflect reports automatically.

D　The made reports on the server system can be browsed on the electronic medical record system.

E　It is easy to backup the data.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 1, May, 2012

⑹　Contact point for the notice : Yasuaki Kawada, Procurement Section, Accounting Division, Kagawa University, 1750―1 Oaza Ikenobe Miki-cho Kita-gun Kagawa 761―0793 Japan, TEL 087―891―2028

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 24 年 3 月 16 日

国立大学法人九州大学事務局長　本木　章喜

◎調達機関番号　415　◎所在地番号　40

○第 11 号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　高度診療支援システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成24年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　人工心肺システム　人工心肺システムは、人工心肺装置及びその関連装置を有すること。

Ｂ　補助循環システム　補助循環システムは、ＰＣＰＳ、ＩＡＢＰ、ＥＣＭＯ装置を有すること。

Ｃ　人工呼吸器システム　人工呼吸器システムは、成人～小児に対応する装置を有すること。また、非侵襲型の人工呼吸器を有すること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成24年５月２日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒812―8581 福岡市東区箱崎六丁目10番１号　国立大学法人九州大学財務部調達課備品調達係　吉田　孝士　電話092―642―4327

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年３月16日から平成24年５月２日まで。

⑵　交付場所　上記２⑵に同じ。