⑸　Time limit for the submission of the requested material : 5 : 00 PM 1, May, 2012
⑹　Contact point for the notice : Toshiyuki Kawamura, Contract Section, Finance Division, Kobe University, 1―1 Rokkodai-cho Nada-ku Kobe-shi 657―8501 Japan, TEL 078―803―5152

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成24年3月26日
国立大学法人徳島大学長　香川　　征
◎調達機関番号　415　◎所在地番号　36
○第３号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　処方・調剤支援トータルシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　本システムは全自動ＰＴＰ払出装置、調剤監査支援システム、全自動散薬分包機、薬袋印字機、投薬表示システムを含んでいること。
B　全自動ＰＴＰ払出装置は以下の要件を満たすこと。
ａ　１錠単位で端数分もカットし払出が出来ること。
ｂ　６分割トレイで運用でき、束払出装置を有していること。
ｃ　本装置は当院にて使用中のトーショー製処方解析システムと連動し運用が出来ること。
C　調剤監査支援システムは以下の要件を満たすこと。
ａ　処方箋に印字されているバーコードをモバイル端末で読むことができること。
ｂ　患者毎にどの薬剤を何錠（包）取り揃える必要があるか瞬時に表示できること。
D　全自動散薬分包機は以下の要件を満たすこと。
ａ　全自動散薬分包機の印字は全行印字熱転写方式であること。
ｂ　制御装置だけでなく、本体備え付けのタッチパネルからの入力でもでき、分包が行えること。
E　薬袋印字機は以下の要件を満たすこと。
ａ　薬袋印字装置からは処方箋、薬袋、薬剤情報が印刷され、患者毎にセットされること。
ｂ　薬袋の別包処理ができること。
F　投薬表示システムは以下の要件を満たすこと。
ａ　プラズマディスプレイ投薬表示盤の２台構成とすること。
ｂ　患者に対しての投薬状況表示（完了番号・待ち時間・先行番号・遅れ番号表示）テロップによるメッセージ表示が掲示出来る事。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年５月14日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒770―8503　徳島市蔵本町２―50―１　国立大学法人徳島大学病院経理調達課調達係　大坂　誠一　電話088―633―9559
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年３月26日から平成24年５月14日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年４月５日14時00分
⑵　開催場所　国立大学法人徳島大学病院入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Pharmacy Support System and equipment 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system and equipment must be included Full Automatic Blister Medicine Dispensing Machine, Administration Support System, Full Automatic Powder Medicine Dispensing Machine, Printing System for Prescription & Envelop and Notice Board System for Patients.
B　Full Automatic Blister Medicine Dispensing Machine must be fulfilled following specifications.
a　It can serve by one tablet and also cut odd tablet.
b　It can use 6 division tray and has the function of serving several medicine sheets at one time.
c　This equipment can be used in conjunction with prescription analysis system made by TOSHO that is used at our hospital.
C　Administration Support System must be fulfilled following specifications.
a　Handy (Mobile) terminal must be able to scan barcode that printed on prescription sheets.
b　The system must be able to show necessary information (for which patient, what and how many medicines) on the screen of handy (mobile) terminal.
D　Full Automatic Powder Medicine Dispensing Machine must be fulfilled following specifications.
a　Printing method of Full Automatic Powder Medicine Dispensing Machine must be full line thermal print.
b　The system must be able to be commanded to dispense by not only control PC, but also touch screen mounted on main body of the machine.
E　Printing System for Prescription and Envelop must be fulfilled following specifications.
a　Prescription sheets, envelops and drug information sheets must be printed by Printing System for Prescription and Envelop in patient specifically.
b　The system must be able to process to print special independent envelop.
F　Notice Board System for Patients must be fulfilled following specifications.
a　The system must implement 2 sets of plasma display panels as notice board.
b　The system must be able to display to patient for status of administration of medication (queue no. of ready to dispense, waiting time, preceding queue no. and delaying queue no.) and text messages.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 14, May, 2012
⑹　Contact point for the notice : Seiichi Osaka, Supplies, University Hospital, Accounting and Supply Affairs Section, National University Corporation, The University of Tokushima, 2―50―1 Kuramoto-cho Tokushima-shi 770―8503 Japan, TEL 088―633―9559

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成24年３月26日
国立大学法人徳島大学長　香川　　征
◎調達機関番号　415　◎所在地番号　36
○第４号
1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　ハイブリッド手術対応型Ｘ線血管撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　天井走行式シングルＣアームであること。
B　フラットパネルディテクター搭載型の多目的血管撮影装置であること。
C　患者及び従事者の被爆低減機能を有すること。
D　固定型手術台はカーボンテーブルトップを使用できる装置であること。
E　天吊式照明器具など周辺機器を装備すること。
F　当院の病院情報システムとの連携が可能であること。