⑸　調達に必要とされる基本的な要求要件
　A　このシステムは、各種の事務業務に使用することを目的とする。
　B　このシステムは、学内ＬＡＮに接続できること。
　C　既存の業務システム及びデータの資産が全て継承されること。
　D　システムの障害を未然に防止するための十分な保守体制及び障害発生時の迅速なサポート体制が提供できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年５月14日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒920―1192 金沢市角間町　金沢大学財務部財務管理課調達管理係長　辻田　恵　電話076―264―6162
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年４月11日から平成24年５月14日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年４月20日14時00分
⑵　開催場所　金沢大学総合メディア基盤センタープレゼンテーション室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 14
⑵　Nature and quantity of the products to be rent : Official Computer System 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
　A　The system is intended to be used for a variety of office work.
　B　The system should be connected to the campus LAN system.
　C　The new system must completely inherit the programs and data from the current system.
　D　System disaster should be completely avoided and should be immediately recovered from any troubles, if it could occur.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 14, May, 2012
⑹　Contact point for the notice : Megumi Tsujita, Procurement management Section, Financial management Division, Kanazawa University, Kakuma-machi Kanazawa-shi 920―1192 Japan, TEL 076―264―6162

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成24年４月11日
国立大学法人島根大学長　小林　祥泰
◎調達機関番号　415　◎所在地番号　32

○第１号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　眼科先進診断治療システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成24年度１月以降
⑸　調達に必要とされる基本的な要求要件
　A　光干渉断層計は、高精細かつ広角（９×９㎜程度）の眼底および前眼部観察および画像解析が可能な装置一式から構成されること。
　B　共焦点走査型レーザー検眼鏡は、高精細の眼底造影（インドシアニングリーンおよびフルオレセイン）検査および自家蛍光眼底観察が可能な装置一式から構成されること。
　C　超広角レーザー検眼鏡は、高精細かつ広角（200度程度）の眼底観察が可能な装置一式から構成されること。
　D　多局所網膜電位刺激装置は、眼底多局所の同時刺激、記録、解析が可能な装置一式から構成されること。
　E　無散瞳眼底カメラ装置は、ステレオ眼底写真撮影および緑内障画像解析が可能な装置一式から構成されること。
　F　硝子体白内障手術装置は、高回転（2500rpmを超えるもの）硝子体カッターおよび白内障超音波ハンドピースが使用可能な装置一式から構成されること。
　G　手術顕微鏡装置は、高精度の光学系を有する眼科手術顕微鏡及びＣＣＤカメラ、画像モニター一式から構成されること。
　H　内視鏡装置は、高精細３ＣＣＤイメージングシステム、眼内及び涙道ファイバースコープ、およびモニター一式から構成されること。
　I　角膜移植装置は、角膜内皮移植用マイクロケラトームおよび硬性小物一式から構成されること。
　J　前眼部撮影装置は、細隙灯に接続可能な高精度ＣＣＤカメラおよび電子カルテに接続可能なファイリングシステム二式から構成されること。
　K　自動視野計解析装置は、自動視野計に接続可能で、視野解析ソフトウェアを有する装置一式から構成されること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成24年５月30日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒693―8501 出雲市塩冶町89―１　国立大学法人島根大学医学部会計課契約第二係　福間　　彰　電話0853―20―2043
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成24年４月11日から平成24年５月30日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成24年４月17日14時00分
⑵　開催場所　島根大学医学部本部棟第二会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Advanced system in ophthalmic diagnosis and therapy 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
　A　One set of optical coherent tomography that capable of high-quality wide area (approx. 9×9 ㎜) fundus imaging, anterior segment imaging, and imaging analyses.
　B　One set of confocal scanning laser ophthalmoscopy that capable of high-quality fundus angiographies by indocyanine green and fluorescein, and fundus autofluorescence imaging.
　C　One set of super wide laser ophthalmoscopy that capable of high-quality wide area (approx. 200-degree) fundus imaging.
　D　One set of multifocal electroretinogram that capable of simultaneous multifocal fundus stimulation, recordings, and analyses.
　E　One set of non-mydriatic fundus came that capable of recordings and analyses of stereo fundus photography.
　F　One set of vitreoretinal／cataract surgery system that capable of usage of high-speed vitreoretinal cutter (>2500 rpm) and cataract ultrasound probe.
　G　One set of ophthalmic microscope system that equips high-quality optical system and is consisted of CCD camera and TV monitor.
　H　One set of ophthalmic endoscopy that is consisted of high-quality 3CCD imaging system, intraocular and intralacrymal fibers, and TV monitor.
　I　One set of keratoplasty system that consisted of microkeratome for corneal endothelial transplantation and surgical accessories.