資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成24年８月３日

独立行政法人国立病院機構

神奈川病院長　加勢田　静

◎調達機関番号　597　◎所在地番号　14

○第04703号（№04703）

１　調達内容

⑴　品目分類番号　71、27

⑵　導入計画物品及び数量　病院情報システム１式

⑶　調達方法　購入

⑷　導入予定時期　平成25年度第３・四半期

⑸　調達に必要とされる基本的な要求要件

①　本システムは、独立行政法人国立病院機構神奈川病院における診療業務及び経営管理業務を総合的に支援するためのもので、電子カルテ・オーダリングシステム、部門システム（医事会計、自動再来受付、調剤支援、看護支援、画像管理）、看護勤務管理、栄養管理、リハビリ管理、診察表示、自動精算の業務全般にわたり統合的に管理、参照、運営が行えるマルチベンダー指向のシステムであること。

②　本システムは、患者基本情報、診療情報、画像情報など円滑に交換でき、そのために必要な国際的な標準規格（最新のＤＩＣＯＭ、ＨＬ７等）に準拠した通信規格、データ形式を提供できること。

③　本システムは、業務用データベースとは独立した、患者の臨床データ、情報を長期的総合的に蓄積管理できる機能を有すること。

④　本システムは、要求機能を達成するために十分なデータ処理機器、端末、ネットワークを保有すること。

⑤　本システムは、各部門で使用している医療機器及びシステムに対するインターフェイスを有すること。

⑥　本システムのうち、病院業務に不可欠なシステムは24時間、無人にて自動稼働し安全で容易に運用できること。また、システム障害対策が十分になされていること。

⑦　利用者層に応じたデータのアクセス権が設定できること。また、プライバシーの保護及びセキュリティ対策が充分なされていること。

⑧　本システムは、病院経営、診療業務を高度に支援するため、最新の技術に基づくこと。また、操作性及び応答時間は十分実業務に対応できること。さらに各種データの二次利用が可能であること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成24年９月18日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒257―8585 神奈川県秦野市落合666―１　独立行政法人国立病院機構神奈川病院企画課長　熊澤　克人　電話0463―81―1771　内線2200

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年８月３日から平成24年９月18日まで。

⑵　交付場所　上記２⑵に同じ。

４　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

５　Summary

⑴　Classification of the services to be procured : 71, 27

⑵　Nature and quantity of the services to be required : Hospital Infomation System 1 set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

①　This system will be the integration system which supports medical and administrative service at National Hospital Organization Kanagawa National Hospital. This system will be composed of ordering system which includes Electric Medical Record System ; divisions' system which manages accounting, remuneration, pharmacy and managerial accounting, supply processing distribution, intra-hospital information system, nursing duty control, nutritional control, rehabilitation control, examination indicate, automatic payment, nutrition control and regional networks. This system assumes to be Multi-vendor-system which integrates the control, reference and management of all our facility's business in.

②　This system will provide the Communication tools and Data formats under international standard specification, such as newest DICOM or HL 7 protocols to exchange the patient besic information, the clinical data and the image information smoothly.

③　This System would possess functions to manage the long-term and the comprehensive clinical information of the patients indepently from the database for business use.

④　This system would possess enough data processors, terminal units and networks to attain requirements mentioned this requirement.

⑤　This System would have interfaces for the medical equipment and system used in each section of our facility.

⑥　The system would be indispensable for the hospital continuous servis must be automated for 24 hours. It can be applied safely and easily. And this system fully must take measures to cope with the system failure.

⑦　The access permission of the data corresponding to a user class must be to set up. The protection of the privacy and security countermeasures are sufficient.

⑧　This system must be based on the newest technology to support hospital management and medical care task. And the operational environment and quickresponseability can cope with our actual medicai care, The secondary utilization of the data must be possible.

⑸　Time-limit for the submission of the requested material : 5 : 00 P.M September 18, 2012

⑹　Contact point for the notice : Katsuhito Kumazawa, Director, Accounting Division, National Hospital Oganization Kanagawa National Hospital, 666―1 Ochiai Hadanoshi Kanagawa-ken 257―8585, Japan. TEL 0463―81―1771 ext. 2200

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成24年８月３日

独立行政法人国立がん研究センター

理事長　堀田　知光

◎調達機関番号　611　◎所在地番号　13

○第６号

１　調達内容

⑴　品目分類番号　22、31

⑵　購入等物品及び数量　ＩＶＲ―ＣＴシステム一式

⑶　調達方法　購入等

⑷　導入時期　平成24年度第４・四半期

⑸　調達物品に必要とされる基本的要求要件

①　ＩＶＲ―ＣＴシステムはＸ線アンギオ装置と多列Ｘ線ＣＴ装置を備え、共通の検査台においてシステム連携のもと検査と治療が行えること。

②　全身の画像を高精細に撮像でき、精度の高い診断と治療が行えること。

③　医療Ｘ線被ばくを低減する機構と機能を有すること。

④　病院情報システム、画像ファイリングシステム（ＤＩＣＯＭ3.0）との間で画像、患者検査情報、実施情報等の入出力が可能であること。

⑤　安定で迅速な保守支援体制、ユーザー利用時の支援を十分行える体制を有すること。

⑥　将来の機能拡充に対して柔軟な対応ができること。