E　The ceiling pendant of GCU has a medical gas outlet in one bed, and there are each multiple oxygen, air, absorption, power supply double outlets, and there is a monitor terminal more than one place.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 28, January, 2013

⑹　Contact point for the notice : Yoshida Yasushi, Procurement Section, Accounting Division, Kobe University Hospital, 7―5―1 Kusunoki-cho Chuo-ku Kobe-shi 650―0017 Japan, TEL 078―382―5120

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 24 年 12 月 11 日

国立大学法人神戸大学

契約担当役　理事　竹園　充夫

◎調達機関番号　415　◎所在地番号　28

○第 15 号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　胎児集中監視システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成25年度８月以降

⑸　調達に必要とされる基本的な要求要件

A　胎児集中監視システム・分娩監視装置から構成されること。

B　胎児・母体病態の監視が可能であること。

C　胎児・分娩監視装置からのデータ管理が可能であること。

D　リスクマネジメントのサポート機能を有すること。

E　電子カルテおよびモニターシステムとのスムーズな連動性を有すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年１月28日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒650―0017　神戸市中央区楠町７－５－１　神戸大学医学部管理課病院契約係　吉田　　靖　電話078―382―5120

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年12月11日から平成25年１月28日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成24年12月20日15時30分

⑵　開催場所　神戸大学医学部管理棟入札室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : The fetal monitoring central system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　This system consists of fetal centralized monitoring system and delivery monitoring system.

B　The monitoring of fetal and maternal fluctuation is possible.

C　The data management from maternal and fetal monitoring is possible.

D　It should have a support function of risk management.

E　It should have smooth correlation with an electric Karte system and a monitor system.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 28, January, 2013

⑹　Contact point for the notice : Yoshida Yasushi, Procurement Section, Accounting Division, Kobe University Hospital, 7―5―1 Kusunoki-cho Chuo-ku Kobe-shi 650―0017 Japan, TEL 078―382―5120

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成 24 年 12 月 11 日

国立大学法人大分大学理事　岩切　平治

◎調達機関番号　415　◎所在地番号　44

○第２号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　磁気共鳴画像診断装置　一式

⑶　調達方法　借入

⑷　導入予定時期　平成25年度５月以降

⑸　調達に必要とされる基本的な要求要件

A　超電導マグネットで3.0Ｔ以上の磁場強度を有し、高分解能撮影が可能なこと。

B　磁気共鳴断層撮影システムは、全身の検査が可能であり、中枢神経機能画像、灌流画像解析、拡散強調画像、ＭＲスペクトロスコピーなど先進的な画像検査が可能であること。

C　既存の病院情報システム（ＨＩＳ）、放射線情報システム（ＲＩＳ）、読影システムおよび医用画像管理システム（ＰＡＣＳ）と接続可能であること。

D　磁気共鳴断層撮影システムは、既設の放射線情報システム（ＲＩＳ）より患者および検査情報を取得できること。

E　安定した迅速な保守支援体制があり、ユーザ使用の時に十分な支援を行なうことができること。

F　システムを設置するための工事費用を含むこと。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年１月28日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒879―5593 大分県由布市挾間町医大ケ丘１丁目１番地　大分大学医学・病院事務部経営管理課用度第一係　村田　典彦　電話097―586―5270

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成24年12月11日から平成25年１月28日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成24年12月20日15時00分

⑵　開催場所　大分大学医学部管理棟中会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be rent : Magnetic Resonance Imaging System 1 Set

⑶　Type of the procurement : Rent

⑷　Basic requirements of the procurement :

A　The system must have the magnetic field strength of 3.0T or more with a superconducting magnet to provide high resolution images.

B　The system can perform MR examinations of the whole body and should have advanced performance applications, including neural functional image, perfusion analysis, diffusion weighted image, and magnetic resonance spectroscopy (MRS).

C　The system should have an interface to communicate with the existing hospital information system (HIS), radiology information system (RIS), radiology reporting system, and picture archiving and communication system (PACS).

D　The system should be able to acquire data of patients' information and of other examinations through existing radiology information system (RIS).

E　There should be the stable and prompt maintenance-support system which can provide sufficient on-site support in hospital. at the time of the user use.

F　The cost of construction to install the system should be included in the total cost of proposal.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 28, January, 2013

⑹　Contact point for the notice : Norihiko Murata, Procurement Section, Accounting Division, Oita University, 1―1 Idaigaoka Hasama-machi Yufu-shi Oita 879―5593 Japan, TEL 097―586―5270