◎調達機関番号　415　◎所在地番号　43

○第２号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　歯科口腔外科処置システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成25年度４月以降

⑸　調達に必要とされる基本的な要求要件

A　歯科口腔外科処置システム

a　歯科診察台、ライト、切削器具、含嗽スピットンなどで構成される歯科用ユニットであること。

b　電気手術器装置は、切開・凝固・バイポーラ機能を有し口腔内の処置が行えること。

c　歯科用Ｘ線装置は、高品質デジタルＸ線画像撮影ができる歯科用イメージングプレートデジタルＸ線画像システムを備えていること

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年５月７日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒860―8556　熊本市中央区本荘１－１－１　熊本大学医学部附属病院事務部財務・調達ユニット　上野　孝之　電話096―373―5938

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成25年３月21日から平成25年５月７日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成25年４月４日10時00分

⑵　開催場所　熊本大学医学部附属病院第一会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Dentistry and Oral Surgery treatment system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Dentistry and Oral Surgery treatment system

a　The system should consist of the dental care units,including a adjustable dental chair,a light,a cuspidor,and dental engines.

b　The Electrosurgical instrument system should have the functional modes for incision,coagulation, and bipolar attachment and enable us to treat in the oral cavity.

c　Dental X-ray equipments, they have a digital X-ray imaging system withdental imaging plates which enable us to obtain high-quality digital X-ray images.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 7, May, 2013

⑹　Contact point for the notice : Takayuki Ueno, Finance and Procurement Division, Hospital Administration Office, Kumamoto University, 1―1―1 Honjo, Chuo-ku, Kumamoto-shi 860―8556 Japan, TEL 096―373―5938

○第３号

1　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　眼科診断治療システム　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成25年度４月以降

⑸　調達に必要とされる基本的な要求要件

A　細隙灯顕微鏡検査装置

a　細隙灯顕微鏡は、眼球の状態を観察する器械でシステムテーブルを使うことにより効率よく診察が行えること。またコードレスＬＥＤ単眼倒像検眼鏡を使うことによりきめ細やかな眼底観察が行えること。また小型高性能カメラを使うことにより患者への説明が容易に行えること。

B　眼科用レーザ光凝固装置

a　低侵襲で従来より短時間で効率的に光凝固が行えること。

C　眼科検査装置

a　ＬＥＤ視力検査装置を使うことによりコスト削減、電球交換の手間を省けること。自動視野計は、視野検査の診断補助から経過観察まで臨床に有用な各種解析ソフトを備えていること。眼軸長測定装置は、光干渉式により眼内レンズ度数計算に必要な全てのパラメーターの測定が行えること。非接触式眼圧計は、目に空気を当てることにより角膜の形状の変化から眼圧の測定が行えること。非接触角膜内皮測定装置は、角膜の内皮細胞を自動撮影、マニュアル撮影、自動解析、マニュアル解析が行えること。

D　眼科画像ファイリング装置

a　院内データの一元管理、画像への直接描画によるインフォームドコンセント、レポート作成、検査機器からのデータ直接取込みなど、データの有効な管理・利用が可能な医療用データマネージメントシステムであること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年５月７日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒860―8556　熊本市中央区本荘１－１－１　熊本大学医学部附属病院事務部財務・調達ユニット　上野　孝之　電話096―373―5938

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成25年３月21日から平成25年５月７日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成25年４月４日14時00分

⑵　開催場所　熊本大学医学部附属病院第一会議室

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Ophthalmic diagnostic and therapeutic system 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　Slit-lamp microscopy system

a　Slit-lamp microscopy is an equipment by which we can observe an eye condition. System table enable us to see a patient efficiently ; LED monocular indirect ophthalmoscope enable us to see a fundus in detail ; small, high performance camera enable us to explain to a patient easily.

B　Ophthalmic laser photocoagulation system

a　This ophthalmic laser photocoagulation system is newly developed and enables us to perform a retinal laser photocoagulation with less invasive, shorter time and more efficient than the former.

C　Ophthalmic exam system

a　LED visual acuity exam equipment helps a cost reduction and do not need a light bulb replacement. Automated perimeter has a variety of useful analysis software for diagnosis and follow-up in clinical use. Axial length measurement system can measure all parameters for a calculation of intraocular lens based on the principle of a low coherence interferometer. Non-contact intraocular pressure measurement equipment can detect the force of the air jet at the instance of applanation and estimate an intraocular pressure. Non-contact corneal endothelium measurement equipment can evaluate a corneal endothelium by various modes, such as automatic or manual camera control, and automatic or manual analysis