⑤　レベルセンサー（貯血量監視装置）を有し、貯血槽の貯血量が設定以下になった場合に、警報を発し、指定したローラーポンプの回転を制御あるいは停止、または、遠心ポンプ用の送血クランプを作動させることが可能であること。また、貯血量が回復した場合、自動的にローラーポンプの回転を復帰、または、送血クランプを解除させることが可能であること。

⑥　圧力モニターを有し、心肺回路内の圧力が設定値を超えた場合に、指定したポンプの回転を制御あるいは停止させることが可能であること。また、測定圧力は、－200～＋500㎜Hg以上の範囲であること。

⑦　電動方式であること。

⑧　送血ポンプと連動し、ポンプ停止時にオクルーダーが連動して全閉状態となる機能を有すること。

⑨　レベルセンサー動作による送血ポンプ停止時には時間差でオクルーダーが全閉になる機能を有すること。

2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年５月10日17時00分（郵送の場合は必着のこと。）

⑵　提出先　〒892―0853 鹿児島県鹿児島市城山町８番１号　独立行政法人国立病院機構鹿児島医療センター　契約係長　吉田　健司　電話099―223―1151　内線7310

3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成25年３月25日から平成25年５月10日まで。

⑵　交付場所　上記２⑵に同じ。

4　説明会の開催　本公表に基づく導入説明会は開催しない。

5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Artificial heart-lung machines 1 set

⑶　Type of the procurement : Purchase

⑷　Introduction planned time : The 2013 third-quarter

⑸　Basic requirements of the procurement :

①　Capable of master-slave function using 2 or more pumps.

②　At least one pump should be capable of having the pump head and the controller separated for a free layout and to reduce the priming volume of the cardiopulmonary bypass system.

③　The flow meter should be an ultrasonic transit type, does not need the measurement cell and gel, and is capable to place it freely. Also, the flow meter should be capable to measure flow rate with clear liquid such as priming solution.

④　The electrical automatic arterial clamp should be available which is capable of safety function to clamp the arterial line immediately upon level／bubble alarm to prevent sending air, as well as retrograde flow alarm to prevent back flow.

⑤　Capable of a level sensor with alarm and to control and stop the rotation of the assigned roller pump, or to clamp the arterial clamp of the centrifugal pump when the reservoir level is under the preset volume. And when the reservoir level resumes, the pump should resume or the arterial clamp should be opened automatically.

⑥　Capable of a pressure monitor to control and stop of the rotation of the assigned roller pump, when the perfusion pressure of the system is over the preset pressure. The monitoring range should be －200～＋500 ㎜Hg or more.

⑦　The venous clamp should be electrical.

⑧　It should be capable to close completely simultaneously when the roller pump stops.

⑨　It should be capable to close with a delayed time, when the roller pump stops on the alert of level sensor.

⑹　Time-limit for the submission of the requested material : 5 : 00 P.M. May 10, 2013

⑺　Contact point for the notice : Kenji Yoshida, Unit Chief, Accounting Division, National Hospital Organization Kagoshima Medical Center, 8―1 Shiroyama-cho Kagoshima-shi Kagoshima-ken 892―0853, Japan. TEL 099―223―1151 ext. 7310

意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

平成25年３月25日

国土交通省港湾局長　山縣　宣彦

◎調達機関番号　020　◎所在地番号　13

○第56号

1　調達内容

⑴　品目分類番号　72

⑵　特定役務件名及び数量　全国輸出入コンテナ貨物流動調査業務　一式

2　意見の提出方法

⑴　意見の提出期限　平成25年４月15日16時00分（郵便の場合は、必着のこと。）

⑵　提出先　〒100―8918 東京都千代田区霞が関２－１－３　国土交通省港湾局総務課調整係　磯谷　智彦　電話03―5253―8111　内線46175

3　仕様書案の交付

⑴　交付期間　平成25年３月25日から平成25年４月15日まで

⑵　交付場所　上記２⑵に同じ。

4　Summary

⑴　Classification of the services to be procured : 72

⑵　Nature and quantity of the services to be required : Nationwide export-import container cargo flow survey work, 1 Set

⑶　Time-limit for the submission of comments : 16 : 00　15 April 2013

⑷　Contact point for the notice : Tomohiko Isotani, General Affairs Coordination Section, General Affairs Division, Ports and Harbors Bureau, Ministry of Land, Infrastructure, Transport and Tourism 2―1―3 Kasumigaseki, Chiyoda-ku, Tokyo, 100―8918 Japan TEL 03―5253―8111 Ex. 46175

意見招請に関する公示

次のとおり調達特定役務の仕様書案の作成が完了したので、仕様書案に対する意見を招請します。

平成25年３月25日

関東地方整備局長　森北　佳昭

◎調達機関番号　020　◎所在地番号　11

1　調達内容

⑴　品目分類番号　71、27

⑵　調達件名及び数量　特殊車両オンライン申請システム用サーバ賃貸借

2　意見の提出方法

⑴　意見の提出期限　平成25年４月15日17時00分（郵送の場合は必着のこと）

⑵　提出先　〒330―9724 埼玉県さいたま市中央区新都心２－１　国土交通省関東地方整備局総務部契約課購買係　三浦　　孝　電話048―601―3151　内線2537

3　仕様書案の交付

⑴　交付期間　平成25年３月25日から平成25年４月15日まで

⑵　交付場所　上記２⑵に同じ

4　Summary

⑴　Classification of the products to be procured : 71, 27

⑵　Nature and quantity of the product to be leased : Oversize／Overweight vehicles systems server 1 set

⑶　Time-limit for the submission of comments : 17 : 00　15 April, 2013