資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成25年４月１日

国立大学法人北海道大学総長　山口　佳三

◎調達機関番号　415　◎所在地番号　01

○第１号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量　ＣＰＣユニット　一式

⑶　調達方法　購入等

⑷　導入予定時期　平成25年度３月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　「ヒト幹細胞を用いる臨床研究に関する指針」に適合したシステムであること。

Ｂ　医師主導型治験を実施するシステムであること。

Ｃ　ＣＰＷＳ（Cell processing work station）アイソレータシステムと連動するシステムであること。

Ｄ　ＧＭＰの３原則「人為的ミス防止」、「汚染防止」、「品質保証」を実現するシステムであること。

Ｅ　教育への利用を行う教材として適合するシステムであること。

Ｆ　本院既存のＣＰＣと空調・環境モニタリングシステム・各種情報システム等が接続・連動すること。

Ｇ　緊急時の避難および対応が可能なこと。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年５月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒060―8648 札幌市北区北14条西５丁目　北海道大学病院管理課用度第一係　林　　寛　電話011―706―7401

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成25年４月１日から平成25年５月17日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成25年４月10日11時00分

⑵　開催場所　北海道大学病院管理課入札室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : Cell Processing Center Unit 1 Set

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　A CPC unit must comply with “Guidelines on clinical research using human stem cells”.

B　A CPC unit must be for conducting investigator-initiated clinical trials.

C　A CPC unit must work with a CPWS (Cell Processing work station) isolator system.

D　A CPC unit must realize the three principles of GMP, “Prevention of Human Error”, “Prevention of Contamination”, “Quality Assurance”.

E　A CPC unit must suit as teaching materials used for education.

F　Air-conditioning and environmental monitoring system as well as various information systems must connect and work with the CPC unit existing in our facility.

G　Emergency evacuation and response must be possible.

⑸　Time limit for the submission of the requested material : 5 : 00 PM 17, May, 2013

⑹　Contact point for the notice : HAYASHI Yutaka, supplies Ⅰ Section, Financial Division, Hokkaido University Hospital, Kita 14 Nishi 5 Kita-ku Sapporo-shi 060―8648 Japan, TEL 011―706―7401

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

平成25年４月１日

国立大学法人東北大学理事　佃　　良彦

◎調達機関番号　415　◎所在地番号　04

○第１号

１　調達内容

⑴　品目分類番号　22、31

⑵　導入計画物品及び数量

Ｘ線ＣＴ診断システム　二式

⑶　調達方法　購入等

⑷　導入予定時期　平成25年度10月以降

⑸　調達に必要とされる基本的な要求要件

Ａ　高性能Ｘ線ＣＴシステム　１式

ａ　128スライス以上の同時収集ができること。

ｂ　心臓ＣＴ検査および機能診断に対応したＣＴ装置であること。

ｃ　被曝量を減らす再構成処理が可能であり、低線量におけるアーチファクトの軽減対策がなされていること。

ｄ　安定した迅速な保守・支援体制があり、ユーザーの利用時に十分な支援を行えること。

ｅ　本装置は、ＤＩＣＯＭ規格およびＩＨＥの統合プロファイルに準拠したシステムであること。

Ｂ　Ｘ線ＣＴシステム　１式

ａ　64スライス以上の同時収集ができること。

ｂ　全身の形態診断に対応したＣＴ装置であること。

ｃ　被曝量を減らす再構成処理が可能であり、低線量におけるアーチファクトの軽減対策がなされていること。

ｄ　安定した迅速な保守・支援体制があり、ユーザーの利用時に十分な支援を行えること。

ｅ　本装置は、ＤＩＣＯＭ規格およびＩＨＥの統合プロファイルに準拠したシステムであること。

２　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

⑴　資料等の提供期限　平成25年５月17日17時00分（郵送の場合は必着のこと。）

⑵　提供先　〒980―8577 仙台市青葉区片平２－１－１　東北大学財務部調達室調達第一係長　菊地　　崇　電話022―217―4869

３　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。

⑴　交付期間　平成25年４月１日から平成25年５月17日まで。

⑵　交付場所　上記２⑵に同じ。

４　説明会の開催　本公表に基づく導入説明会を開催する。

⑴　開催日時　平成25年４月５日16時00分

⑵　開催場所　東北大学病院仮管理棟４階第１会議室

５　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

６　Summary

⑴　Classification of the products to be procured : 22, 31

⑵　Nature and quantity of the products to be purchased : CT diagnosis system 2 Sets

⑶　Type of the procurement : Purchase

⑷　Basic requirements of the procurement :

A　The high quality CT system 1 Set

a　The CT scanner must have 128 or more multislice capability.

b　The CT scanner can perform ECG-gated cardiac scan and functional scan.

c　The device to reduce X-ray exposure and the techniques to reduce artifacts induced by low-dose X-ray should be provided.

d　Stable and prompt maintenance and support for use are required.

e　This system should be based on DICOM and the Integration Profile which is defined by IHE-J (Integrating the Healthcare Enterprise Japan).

B　The CT system 1 Set

a　The CT scanner must have 64 or more multislice capability.

b　The system must allow acquisition of CT images of the whole body for diagnosis by anatomical means.

c　The device to reduce X-ray exposure and the techniques to reduce artifacts induced by low-dose X-ray should be provided.

d　Stable and prompt maintenance and support for use are required.

e　This system should be based on DICOM and the Integration Profile which is defined by IHE-J (Integrating the Healthcare Enterprise Japan).

⑸　Time limit for the submission of the requested material : 5 : 00 PM 17, May, 2013

⑹　Contact point for the notice : Takashi Kikuchi, Procurement Services Office, Finance Department, Tohoku University, 2―1―1 Katahira Aoba-ku Sendai-shi 980―8577 Japan, TEL 022―217―4869