3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月１日から平成25年５月17日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年４月８日14時00分
⑵　開催場所　鳥取大学医学部附属病院第三会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Digital imaging System 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　X-ray equipment
a　The latest application including the stitching radiography and the tomosynthesis must be equipped with at this system.
b　This system must be able to add wireless type flat panel detector.
c　It is necessary to be this system which can be equivalent to various patient cases.
d　This system must be having the quality assurance for images.
e　This system must be able to keep the function similar to the latest system by upgrade.
f　This system must have a strict quality control system which can be used as needed.
g　Manufacture must insure prompt and efficient service for maintenance and repair.
h　This system should connect to established RIS, HIS, PACS and improve radiography operating smoothly and efficiently.
i　Manufacture must support our medical investigate.
B　X-ray equipment (infant)
a　This system must be able to add wireless type flat panel detector.
b　As for this system, a delay time of the X-ray generation must be as short as possible.
c　Manufacture must insure prompt and efficient service for maintenance and repair.
d　This system should connect to established RIS, HIS, PACS and improve radiography operating smoothly and efficiently.
e　Manufacture must support our medical investigate.
C　Panoramic-Cephalography X-ray system
a　This system must be able to perform the cone beam computed radiography (CBCT).
b　This system must be able to projection the panoramic radiography, cephalography and CBCT in one device.
c　The cephalography must be a one-shot projection.
d　It is necessary to be equipped with the safety device.
e　This system must be having the workstation of the image processing.
f　This system should be able to perform all operation with one console.
g　This system must have a strict quality control system which can be used as needed.
h　Manufacture must insure prompt and efficient service for maintenance and repair.
i　This system should connect to established RIS, HIS, PACS and improve radiography operating smoothly and efficiently.
D　Dental system
a　In this system, all output images must be a digital image.
b　This system must be able to choose several kinds of film size.
c　At the time of the projection, the X-ray tube must be able to point to the many directions.
d　Being the system that the darkroom becomes needless.
e　It is necessary to be able to confirm the image as soon as possible after the projection.
f　Manufacture must insure prompt and efficient service for maintenance and repair.
g　This system should connect to established RIS, HIS, PACS and improve radiography operating smoothly and efficiently.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 17, May, 2013
⑹　Contact point for the notice : Akihiro Urata, Hospital Procurement Section, Accounting and Procurement Divisio, Faculty of Medicine, Tottori University, 36―1 Nishi-cho Yonago-shi 683―8504 Japan, TEL 0859―38―7087

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成25年４月１日
国立大学法人島根大学長　小林　祥泰
◎調達機関番号　415　◎所在地番号　32

○第１号

1　調達内容
⑴　品目分類番号　４、31
⑵　導入計画物品及び数量　デジタル型エックス線撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成25年度12月以降
⑸　調達に必要とされる基本的な要求要件
A　デジタル型エックス線撮影システムは、骨部Ｘ線撮影装置、胸腹部Ｘ線撮影装置、乳房Ｘ線撮影装置、パノラマ・セファロＸ線撮影装置、頭部Ｘ線撮影装置、Ｘ線ＴＶ装置およびワイヤレス撮影システム（ポータブル）、画像処理装置で構成されること。
B　各Ｘ線撮影装置の基本構成は、全てＦＰＤ（Flat Panel Detector）としＨＩＳ―ＲＩＳおよびＰＡＣＳとネットワーク接続できること。
C　骨部Ｘ線撮影装置は、ワイヤレス撮影、脊椎・下肢の長尺撮影が可能であること。
D　胸腹部Ｘ線撮影装置は、本学の職員・学生検診用にも使用できスループットが高いこと。
E　乳房Ｘ線撮影装置は高画質が提供でき５Ｍの画像ビューアーを備えること。
F　パノラマ・セファロＸ線撮影装置は下顎骨のＣＴ画像が作成できること。
G　頭部Ｘ線撮影装置は、再現性の高いポジショニングができること。
H　Ｘ線ＴＶ装置は汎用型で非ＩＶＰ補助のため超音波診断装置と心電モニター（12誘導）を備えること。
I　ワイヤレス撮影システム（ポータブル）は、院内ＬＡＮ、ＰＡＣＳとリアルタイムに無線通信できること。
J　画像処理装置は、他のモダリティーとの連携も可能で診断精度を高めるための機能を有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年５月17日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒693―8501 出雲市塩冶町89―１　国立大学法人島根大学医学部会計課契約第二係　福間　　彰　電話0853―20―2043
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月１日から平成25年５月17日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年４月３日11時00分
⑵　開催場所　島根大学医学部本部棟第二会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。