E　個別対応食の調理業務に対応した設備であること。
F　厨房内温度管理及び食品保管及び食品保存機器の温度管理機能を有する設備であること。
G　食器の洗浄・滅菌ができること。
H　機器は、電気厨房機器を基本とし、省エネルギーに対応していること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年５月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒910―1193 福井県吉田郡永平寺町松岡下合月23―３　福井大学病院部総務管理課調達管理係　山崎　　徹　電話0776―61―3111　内線3147
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月11日から平成25年５月27日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年４月17日10時30分
⑵　開催場所　福井大学医学部管理棟中会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 13
⑵　Nature and quantity of the products to be purchased : Patient meal cooking and a hygiene supervision system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The facility/equipment shall be for producing and serving any diets meals to the patients.
B　The facility shall have New Cook Chill system and Cook-Serve system.
C　The facility and equipment shall be able to provide heat retaining foodservice.
D　The facility shall have central cooking, central plating and delivery system.
E　The facility shall be able to cook and serve individual/personal diet meal.
F　The facility shall have temperature control system for kitchen, food storage and food storage equipment.
G　The facility shall be able to clean and sterilization of tableware.
H　The equipment is basically an electric kitchen equipment shall correspond to energy conservation.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 27, May, 2013
⑹　Contact point for the notice : Tooru Yamazaki, Procurement Section, Accounting Division, University of Fukui, 23―3 Matsuokashimoaizuki Eiheiji-cho Yoshida-gun Fukui 910―1193 Japan, TEL 0776―61―3111 ex. 3147

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 25 年４月 11 日
国立大学法人浜松医科大学
理事　瀧川　雅浩

◎調達機関番号　415　◎所在地番号　22

○第１号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　診断用Ｘ線ＣＴ撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成25年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　64列以上の検出器を有していること。
B　被ばく低減及びアーチファクト低減機構を有していること。
C　デュアルエナジースキャンが可能であること。
D　冠動脈ＣＴ検査に対応可能であること。
E　高速画像処理および解析機能を有していること。
F　既存の放射線情報システム（ＲＩＳ）、および医用画像管理システム（ＰＡＣＳ）、３Ｄ画像処理システムと接続可能であること。
G　システムを設置するための工事を含むこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年５月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒431―3192 浜松市東区半田山１―20―１　浜松医科大学会計課専門職員（契約担当）　高橋　冠次　電話053―435―2148
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月11日から平成25年５月27日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年４月17日13時00分
⑵　開催場所　浜松医科大学管理棟４階第二会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Diagnostic X-ray computed tomography system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The system should be equipped with detector-rows of 64 or more.
B　The system should have the device to reduce X-ray exposure and the technique to reduce artifacts induced by low-dose X-ray.
C　The system should have a capability of dual energy scan.
D　The system should have a capability of coronary artery scan.
E　The system should have the functions for high-speed image processing and analysis.
F　The system should have an interface to communicate with the existing radiology information system (RIS), picture archiving and communication system (PACS), 3D image processing system.
G　The cost of construction and installation should be included.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 27, May, 2013
⑹　Contact point for the notice : Kanji Takahashi, Expert Official, Accounting Section, Hamamatsu University School of Medicine, 1―20―1 Handayama Higashi-ku Hamamatsu-shi　431―3192　Japan, TEL 053―435―2148

○第２号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　診断用磁気共鳴断層撮影システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成25年度11月以降
⑸　調達に必要とされる基本的な要求要件
A　均一性に優れた静磁場強度3.0Ｔ以上の超伝導型マグネットを有していること。
B　高速撮像が可能なグラディエントシステムを有していること。
C　独立した複数のRF送受信システムを有し、不均一を防止する機能を有すること。
D　多チャンネルフェーズドアレイコイルシステムを有していること。
E　全身のＭＲＩ／ＭＲＡ撮像の他、先進的アプリケーションおよび周辺機器を有していること。
F　既存の放射線情報システム（ＲＩＳ）、および医用画像管理システム（ＰＡＣＳ）、３Ｄ画像処理システムと接続可能であること。
G　システムを設置するための工事を含むこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年５月27日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒431―3192 浜松市東区半田山１―20―１　浜松医科大学会計課専門職員（契約担当）　高橋　冠次　電話053―435―2148