⑩　操作性、患者や術者の快適性と安全性および空間効率に配慮した設計であること。
⑪　6,000×5,800㎜の部屋にＸ線装置及び発生器等の器材が収まること。
⑫　その維持について適切な保守体制と迅速なユーザー支援体制を有すること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年６月10日17時00分（郵送の場合は必着のこと。）
⑵　提出先　〒565―8565 大阪府吹田市藤白台５－７－１　独立行政法人国立循環器病研究センター　総務課契約係　右田　涼一　電話06―6833―5012　内線2207
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月25日から平成25年６月10日まで。
⑵　交付場所　上記２⑵に同じ。
4　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Full Digital Angiographic System 1 set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
①　The angiographic system should be applied for catheter examinations and interventional procedures for cerebrovascular disease.
②　The biplane-angiographic system should show us high performance for digital fluoroscopy and angiography.
③　The biplane-angiographic system should enable us to obtain high quality fluoroscopic and angiographic images.
④　The angiographic system should enable us to perform rotational angiography procedure in combination with 3D reconstruction image. In addition, the system should also have much memory for data storage.
⑤　The angiographic system should show us cross-sectional, soft tissue CT image, and also enables us the analysis of the acquired data.
⑥　The angiographic system should have ability for viewing, recording DICOM volume data.
⑦　The angiographic machine should provide us viewing and analysis of 3D volume on intra-network system in National Cerebral Cardiovascular Center.
⑧　After the access to the intra-network in National Cerebral and Cardiovascular Center including the electronic medical rcord, the angiographic system should have a capability to accept the patient information as well as post-processing data.
⑨　Using the angiographic system, radiation exposure of the patient, doctors, and co-medical stuffs during the procedure should be reduced.
⑩　The angiographic system should provide us safety, comfortableness during the interventional procedure. In addition, the system should be suitable for the indicated angiographic room in terms of sizeand layout.
⑪　It is necessary to install X-ray devices and generators on the room of 6,000×5,800 ㎜ for the angiographic system.
⑫　The angiographic system and users should be supported by appropriate and prompt maintenance services.
⑸　Time-limit for the submission of the requested material : 5 : 00 P.M 10, June 2013
⑹　Contact point for the notice : Ryoichi Migita General Affairs Division, National Carebral and Cardiovascular Center, 5―7―1 Fujishirodai Suita-shi, Osaka-fu 565―8565 Japan. TEL 06―6833―5012 ext. 2207

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成 25 年４月 25 日
国立大学法人宮崎大学
契約担当役　理事　松ヶ迫和峰

◎調達機関番号　415　◎所在地番号　45
○第１号
1　調達内容
⑴　品目分類番号　22
⑵　導入計画物品及び数量　ＳＰＥＣＴ／ＣＴシステム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成25年度８月以降
⑸　調達に必要とされる基本的な要求要件
Ａ　２検出器以上を有するＳＰＥＣＴ／ＣＴ装置であること。
Ｂ　ＣＴ装置はマルチスライスＣＴであること。
Ｃ　ＳＰＥＣＴ／ＣＴ装置は、全身のシンチグラム撮影が可能であること。
Ｄ　ＳＰＥＣＴ／ＣＴ装置は、全身のＳＰＥＣＴ／ＣＴ及びＣＴ撮影が可能であること。
Ｅ　ＳＰＥＣＴ／ＣＴ装置は、ＳＰＥＣＴ／ＣＴ画像及びＣＴ画像の自動融合画像作成が可能であること。
Ｆ　各種検査に対応したアプリケーションソフトウェアが利用可能であること。
Ｇ　ＳＰＥＣＴ／ＣＴシステムは、既設の放射線情報システムよりＤＩＣＯＭ3.0通信規格に準拠した患者及び検査情報を取得できること。
Ｈ　ＳＰＥＣＴ／ＣＴシステムで取得した画像は、既存の画像保存システム（ＰＡＣＳ）にＤＩＣＯＭ3.0通信規格に準拠したデータとして保存でき、既存の端末で参照可能なこと。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年６月10日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒889―2192 宮崎市学園木花台西一丁目１番地　国立大学法人宮崎大学財務部経理調達課経理調達係政府調達担当係長　迫口佳代子　電話0985―58―7574
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年４月25日から平成25年６月10日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年５月１日14時00分
⑵　開催場所　宮崎大学木花キャンパス事務局４階会議室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。
6　Summary
⑴　Classification of the products to be procured : 22
⑵　Nature and quantity of the products to be purchased : SPECT／CT system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The SPECT／CT system should be equipped with 2 or more detectors.
B　The CT portion should be equipped with multi-slice CT detector system.
C　The SPECT system should be able to acquire whole-body data.
D　The SPECT／CT system should be able to acquire whole-body SPECT and CT data.
E　The SPECT／CT system should be able to create automatically fused images of SPECT and CT data.
F　Application software appropriate for various types of exams should be available.
G　The SPECT／CT system must be connected to the radiological information system (RIS) installed in the Miyazaki University hospital, and it can communicate patient's and examination's information data with the RIS.
H　The image data obtained with SPECT／CT system must conform exactly to the standard of the DICOM 3.0 and they can be viewed by all the client's viewing system of Miyazaki University hospital.