5　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Integrative managerial system of radiation image information 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The required system should succeed to all functions that had been realized by the present system. Moreover, the system should increase efficiency in radiology department work, and contribute to patient service and medical safety.
B　The required system should succeed to the specifications and functions on cooperation and connection with the hospital information system and present modalities.
C　The required system should offer the environment in which the users can operate present subsystems (cine server, image analysis and processing system, radiation therapy information management system, etc.) seamlessly.
D　The required system including the network system should have enough capabilities for security management and preventive measures for system failure.
E　The required system including the network system should have the performance that it can be operated stably with out a stop.
F　The vender should have a maintenance system and the operation training schedule of the whole system.
G　The vender should disclose the technical information, if necessary, to the third party whom we ask to convert the data from this system to another one.
⑸　Time limit for the submission of the requested material : 5 : 00 PM 8, July, 2013
⑹　Contact point for the notice : Hidenori Mukai, Procurement Section, Management Division, Kanazawa University Hospital part, 13―1 Takara-machi Kanazawa-shi 920―8641 Japan, TEL 076―265―2067

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成25年5月21日
国立大学法人名古屋大学
財務担当理事　藤井　良一
◎調達機関番号　415　◎所在地番号　23

○第５号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　ハイブリッド手術システム　一式
⑶　調達方法　購入等
⑷　導入予定時期　平成25年度１月以降
⑸　調達に必要とされる基本的な要求要件
A　全身麻酔管理下に治療を行うことができること。
B　血管内治療と開頭・開胸・開腹手術が可能なシステムであること。
C　患者及び従事者の被ばく低減機能を有すること。
D　医療機器として安全性が考慮されていること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年７月５日16時00分（郵送の場合は必着のこと。）
⑵　提供先　〒466―8560 名古屋市昭和区鶴舞町65　名古屋大学医学部・医学系研究科経理課専門職員　小林　智亮　電話052―744―2798
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年５月21日から平成25年７月５日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年５月28日15時00分
⑵　開催場所　名古屋大学医学部附属病院外来棟４階会議室４
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be purchased : Hybrid Operating Room system 1 Set
⑶　Type of the procurement : Purchase
⑷　Basic requirements of the procurement :
A　The patient can be treated under standard general anesthesia.
B　The system allows IVR, craniotomy, thoracotomy and laparotomy procedures to be performed.
C　The system should have functions to decrease exposure dose for patients and medical staffs.
D　The system must be considered as safe medical equipment and devices.
⑸　Time limit for the submission of the requested material : 4 : 00 PM 5, July, 2013
⑹　Contact point for the notice : Kobayashi Tomoaki, Accounting Division, Nagoya University School of Medicine, 65 Tsurumai-cho Showa-ku Nagoya-shi 466―8560 Japan, TEL 052―744―2798

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。
平成25年5月21日
国立大学法人神戸大学
契約担当役　理事　竹園　充夫
◎調達機関番号　415　◎所在地番号　28

○第２号

1　調達内容
⑴　品目分類番号　22、31
⑵　導入計画物品及び数量　鏡視下手術システム　一式
⑶　調達方法　借入
⑷　導入予定時期　平成25年度12月以降
⑸　調達に必要とされる基本的な要求要件
A　各診療科で施行する内視鏡下手術に対応できること。
B　内視鏡映像制御装置は高画質の映像を常に最適の状態で供給ができかつ、記録ができること。
C　内視鏡下手術用エネルギー装置は、患者に与える侵襲が最小に抑えられ、手術を行う上で安全で必要十分なエネルギーを提供できること。
2　資料及びコメントの提供方法　上記１⑵の物品に関する一般的な参考資料及び同⑸の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。
⑴　資料等の提供期限　平成25年７月８日17時00分（郵送の場合は必着のこと。）
⑵　提供先　〒650―0017 神戸市中央区楠町７―５―１　神戸大学医学部管理課病院契約係　吉田　　靖　電話078―382―5120
3　説明書の交付　本公表に基づき応募する供給者に対して導入説明書を交付する。
⑴　交付期間　平成25年５月21日から平成25年７月８日まで。
⑵　交付場所　上記２⑵に同じ。
4　説明会の開催　本公表に基づく導入説明会を開催する。
⑴　開催日時　平成25年５月31日14時00分
⑵　開催場所　神戸大学医学部管理棟入札室
5　その他　この導入計画の詳細は導入説明書による。なお、本公表内容は予定であり、変更することがあり得る。

6　Summary
⑴　Classification of the products to be procured : 22, 31
⑵　Nature and quantity of the products to be rent : Endoscopic surgery system 1 Set
⑶　Type of the procurement : Rent
⑷　Basic requirements of the procurement :
A　The system should be suitable to multidisciplinary endoscopic surgeries.
B　Endoscopic image control device should be able to provide optimal video pictures at high resolution and the pictures can be recorded.
C　Cautery device for endoscopic surgery should be minimally invasive to patient and can provide a necessary and sufficient energy in safety during the operations.